

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010322\
 Data File : VX026226.D
 Acq On : 03 Jan 2022 17:33
 Operator : JC/MD
 Sample : M5335-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 31276

Quant Time: Jan 04 03:52:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	181459	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	312944	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	275120	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	123160	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	140170	53.394	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.780%	
35) Dibromofluoromethane	5.391	113	103122	50.685	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.380%	
50) Toluene-d8	8.653	98	369460	48.820	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 97.640%	
62) 4-Bromofluorobenzene	11.079	95	132543	49.199	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.400%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	6615	4.651	ug/l #	85
19) Methyl tert-butyl Ether	3.117	73	4988	0.711	ug/l	95
22) Diisopropyl ether	3.770	45	1695	0.224	ug/l #	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010322\
Data File : VX026226.D
Acq On : 03 Jan 2022 17:33
Operator : JC/MD
Sample : M5335-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
31276

Quant Time: Jan 04 03:52:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

