

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010322\
 Data File : VX026231.D
 Acq On : 03 Jan 2022 19:28
 Operator : JC/MD
 Sample : M5242-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Jan 04 03:54:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	206569	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	360496	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	310155	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	131615	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	156465	52.357	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 104.720%		
35) Dibromofluoromethane	5.391	113	117352	50.071	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.140%		
50) Toluene-d8	8.653	98	425400	48.797	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 97.600%		
62) 4-Bromofluorobenzene	11.079	95	145020	46.730	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 93.460%		
Target Compounds						
16) Acetone	2.386	43	3022	1.866	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010322\
Data File : VX026231.D
Acq On : 03 Jan 2022 19:28
Operator : JC/MD
Sample : M5242-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Jan 04 03:54:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

