

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010322\
 Data File : VX026245.D
 Acq On : 04 Jan 2022 01:38
 Operator : JC/MD
 Sample : PB141783ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141783ZHE#03

Quant Time: Jan 04 07:59:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	208631	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	366015	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	326553	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	136202	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	165367	54.789	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery	= 109.580%		
35) Dibromofluoromethane	5.391	113	119958	50.411	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 100.820%		
50) Toluene-d8	8.653	98	442021	49.939	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery	= 99.880%		
62) 4-Bromofluorobenzene	11.079	95	153107	48.592	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery	= 97.180%		
Target Compounds						Qvalue
16) Acetone	2.380	43	18470	11.294	ug/l	97
20) Methylene Chloride	2.794	84	7688	2.033	ug/l	96
43) Isopropyl Acetate	6.348	43	21616	2.980	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010322\
Data File : VX026245.D
Acq On : 04 Jan 2022 01:38
Operator : JC/MD
Sample : PB141783ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141783ZHE#03

Quant Time: Jan 04 07:59:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

