

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010322\
 Data File : VX026249.D
 Acq On : 04 Jan 2022 03:10
 Operator : JC/MD
 Sample : PB141783ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141783ZHE#07

Quant Time: Jan 04 08:00:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	201677	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	357991	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	319965	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	132107	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	161011	55.185	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.360%
35) Dibromofluoromethane	5.391	113	116308	49.973	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.940%
50) Toluene-d8	8.653	98	431178	49.806	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.620%
62) 4-Bromofluorobenzene	11.079	95	150926	48.973	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.940%
Target Compounds						
						Qvalue
16) Acetone	2.373	43	11144	7.049	ug/l	# 88
20) Methylene Chloride	2.794	84	32988	12.339	ug/l	94
43) Isopropyl Acetate	6.342	43	21311	3.004	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010322\
Data File : VX026249.D
Acq On : 04 Jan 2022 03:10
Operator : JC/MD
Sample : PB141783ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141783ZHE#07

Quant Time: Jan 04 08:00:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

