

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010322\
 Data File : VX026251.D
 Acq On : 04 Jan 2022 03:56
 Operator : JC/MD
 Sample : PB141783ZHE#09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141783ZHE#09

Quant Time: Jan 04 08:01:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	218186	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	386113	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	355504	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	146772	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	171503	54.333	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 108.660%			
35) Dibromofluoromethane	5.391	113	125660	50.059	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.120%			
50) Toluene-d8	8.653	98	468503	50.176	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 100.360%			
62) 4-Bromofluorobenzene	11.079	95	167030	50.251	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 100.500%			
Target Compounds					Qvalue	
16) Acetone	2.380	43	10214	5.972	ug/l	95
20) Methylene Chloride	2.794	84	8391	2.164	ug/l	94
43) Isopropyl Acetate	6.342	43	21822	2.852	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010322\
Data File : VX026251.D
Acq On : 04 Jan 2022 03:56
Operator : JC/MD
Sample : PB141783ZHE#09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141783ZHE#09

Quant Time: Jan 04 08:01:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

