

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010322\
 Data File : VX026252.D
 Acq On : 04 Jan 2022 04:19
 Operator : JC/MD
 Sample : PB141783ZHE#10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141783ZHE#10

Quant Time: Jan 04 08:01:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	209501	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	374192	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	329757	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	135348	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	167903	55.398	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.800%
35) Dibromofluoromethane	5.391	113	121930	50.120	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.240%
50) Toluene-d8	8.653	98	443256	48.985	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.960%
62) 4-Bromofluorobenzene	11.079	95	153004	47.498	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.000%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	10654	6.488	ug/l	95
20) Methylene Chloride	2.794	84	8407	2.300	ug/l	89
43) Isopropyl Acetate	6.342	43	21758	2.934	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010322\
Data File : VX026252.D
Acq On : 04 Jan 2022 04:19
Operator : JC/MD
Sample : PB141783ZHE#10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141783ZHE#10

Quant Time: Jan 04 08:01:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

