

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010322\
 Data File : VX026247.D
 Acq On : 04 Jan 2022 02:24
 Operator : JC/MD
 Sample : PB141783ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141783ZHE#05

Quant Time: Jan 04 08:00:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.555	168	198628	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	350586	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	320694	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	132484	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	155976	54.280	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.560%	
35) Dibromofluoromethane	5.391	113	113716	49.891	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.780%	
50) Toluene-d8	8.652	98	425993	50.247	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.500%	
62) 4-Bromofluorobenzene	11.079	95	150334	49.811	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.620%	
Target Compounds						Qvalue
16) Acetone	2.379	43	10897	6.999	ug/l	92
20) Methylene Chloride	2.794	84	32693	12.422	ug/l	91
43) Isopropyl Acetate	6.342	43	21138	3.043	ug/l	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010322\
Data File : VX026247.D
Acq On : 04 Jan 2022 02:24
Operator : JC/MD
Sample : PB141783ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141783ZHE#05


Quant Time: Jan 04 08:00:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.555 min Scan# 71
Delta R.T. -0.001 min
Lab File: VX026247.D
Acq: 04 Jan 2022 02:24

Instrument :
MSVOA_X
ClientSampleId :
PB141783ZHE#05







Tgt Ion:168 Resp: 198628
Ion Ratio Lower Upper
168 100
99 65.4 45.8 68.8



#16
Acetone
Concen: 6.999 ug/l
RT: 2.379 min Scan# 212
Delta R.T. -0.007 min
Lab File: VX026247.D
Acq: 04 Jan 2022 02:24

Tgt Ion: 43 Resp: 10897
Ion Ratio Lower Upper
43 100
58 26.5 24.8 37.2









#20
Methylene Chloride
Concen: 12.422 ug/l
RT: 2.794 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX026247.D
Acq: 04 Jan 2022 02:24

Instrument :
MSVOA_X
ClientSampleId :
PB141783ZHE#05

Tgt Ion: 84 Resp: 32693




Ion	Ratio	Lower	Upper
84	100		
49	133.8	97.1	145.7
51	43.8	29.3	43.9
86	62.1	50.8	76.2




#33
1,2-Dichloroethane-d4
Concen: 54.280 ug/l
RT: 5.958 min Scan# 799
Delta R.T. -0.006 min
Lab File: VX026247.D
Acq: 04 Jan 2022 02:24

Tgt Ion: 65 Resp: 155976

Ion	Ratio	Lower	Upper
65	100		
67	50.4	0.0	103.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 91
Delta R.T. -0.006 min
Lab File: VX026247.D
Acq: 04 Jan 2022 02:24




Instrument :
MSVOA_X
ClientSampleId :
PB141783ZHE#05


Tgt Ion:114 Resp: 350586
Ion Ratio Lower Upper
114 100
63 22.9 0.0 40.0
88 16.2 0.0 30.6



#35
Dibromofluoromethane
Concen: 49.891 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.000 min
Lab File: VX026247.D
Acq: 04 Jan 2022 02:24

Tgt Ion:113 Resp: 113716
Ion Ratio Lower Upper
113 100
111 103.0 81.8 122.8
192 18.5 15.0 22.4









#43
Isopropyl Acetate
Concen: 3.043 ug/l
RT: 6.342 min Scan# 862
Delta R.T. -0.006 min
Lab File: VX026247.D
Acq: 04 Jan 2022 02:24




Instrument :
MSVOA_X
ClientSampleId :
PB141783ZHE#05


Tgt Ion: 43 Resp: 21138
Ion Ratio Lower Upper
43 100
61 17.5 15.7 23.5
87 12.4 11.2 16.8



#50
Toluene-d8
Concen: 50.247 ug/l
RT: 8.652 min Scan# 1241
Delta R.T. -0.001 min
Lab File: VX026247.D
Acq: 04 Jan 2022 02:24

Tgt Ion: 98 Resp: 425993
Ion Ratio Lower Upper
98 100
100 65.4 52.6 78.8









#62
4-Bromofluorobenzene
Concen: 49.811 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.006 min
Lab File: VX026247.D
Acq: 04 Jan 2022 02:24




Instrument :
MSVOA_X
ClientSampleId :
PB141783ZHE#05


Tgt Ion: 95 Resp: 150334
Ion Ratio Lower Upper
95 100
174 73.4 0.0 157.4
176 70.4 0.0 154.6



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX026247.D
Acq: 04 Jan 2022 02:24

Tgt Ion: 117 Resp: 320694
Ion Ratio Lower Upper
117 100
82 58.5 43.8 65.6
119 32.8 25.2 37.8





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX026247.D


Acq: 04 Jan 2022 02:24

Instrument :

MSVOA_X

ClientSampleId :

PB141783ZHE#05



Tgt Ion:152 Resp: 132484

Ion Ratio Lower Upper

152 100

115 62.9 41.3 123.9

150 158.1 0.0 351.8

