

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010322\
 Data File : VX026248.D
 Acq On : 04 Jan 2022 02:47
 Operator : JC/MD
 Sample : PB141783ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141783ZHE#06

Quant Time: Jan 04 08:00:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.556	168	200794	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	353199	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	316486	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	133060	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	159274	54.829	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.660%	
35) Dibromofluoromethane	5.391	113	116644	50.797	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.600%	
50) Toluene-d8	8.653	98	424074	49.650	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.300%	
62) 4-Bromofluorobenzene	11.079	95	149123	49.044	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.080%	
Target Compounds						Qvalue
16) Acetone	2.380	43	10893	6.921	ug/l	93
20) Methylene Chloride	2.794	84	32325	12.129	ug/l	91
43) Isopropyl Acetate	6.342	43	21092	3.014	ug/l	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010322\
Data File : VX026248.D
Acq On : 04 Jan 2022 02:47
Operator : JC/MD
Sample : PB141783ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141783ZHE#06


Quant Time: Jan 04 08:00:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX026248.D
Acq: 04 Jan 2022 02:47

Instrument :
MSVOA_X
ClientSampleId :
PB141783ZHE#06







Tgt Ion:168 Resp: 200794
Ion Ratio Lower Upper
168 100
99 64.8 45.8 68.8



#16
Acetone
Concen: 6.921 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX026248.D
Acq: 04 Jan 2022 02:47

Tgt Ion: 43 Resp: 10893
Ion Ratio Lower Upper
43 100
58 27.2 24.8 37.2









#20
Methylene Chloride
Concen: 12.129 ug/l
RT: 2.794 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX026248.D
Acq: 04 Jan 2022 02:47

Instrument :
MSVOA_X
ClientSampleId :
PB141783ZHE#06

Tgt Ion: 84 Resp: 32325




Ion	Ratio	Lower	Upper
84	100		
49	136.1	97.1	145.7
51	40.3	29.3	43.9
86	66.1	50.8	76.2




#33
1,2-Dichloroethane-d4
Concen: 54.829 ug/l
RT: 5.958 min Scan# 799
Delta R.T. -0.006 min
Lab File: VX026248.D
Acq: 04 Jan 2022 02:47

Tgt Ion: 65 Resp: 159274

Ion	Ratio	Lower	Upper
65	100		
67	50.4	0.0	103.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 91
Delta R.T. -0.006 min
Lab File: VX026248.D
Acq: 04 Jan 2022 02:47




Instrument :
MSVOA_X
ClientSampleId :
PB141783ZHE#06


Tgt Ion:114 Resp: 353199
Ion Ratio Lower Upper
114 100
63 22.7 0.0 40.0
88 16.3 0.0 30.6



#35
Dibromofluoromethane
Concen: 50.797 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX026248.D
Acq: 04 Jan 2022 02:47

Tgt Ion:113 Resp: 116644
Ion Ratio Lower Upper
113 100
111 102.1 81.8 122.8
192 18.2 15.0 22.4









#43
Isopropyl Acetate
Concen: 3.014 ug/l
RT: 6.342 min Scan# 862
Delta R.T. -0.006 min
Lab File: VX026248.D
Acq: 04 Jan 2022 02:47




Instrument :
MSVOA_X
ClientSampleId :
PB141783ZHE#06


Tgt Ion: 43 Resp: 21092
Ion Ratio Lower Upper
43 100
61 17.1 15.7 23.5
87 11.5 11.2 16.8



#50
Toluene-d8
Concen: 49.650 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX026248.D
Acq: 04 Jan 2022 02:47

Tgt Ion: 98 Resp: 424074
Ion Ratio Lower Upper
98 100
100 64.7 52.6 78.8









#62
4-Bromofluorobenzene
Concen: 49.044 ug/l
RT: 11.079 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX026248.D
Acq: 04 Jan 2022 02:47




Instrument :
MSVOA_X
ClientSampleId :
PB141783ZHE#06


Tgt Ion: 95 Resp: 149123
Ion Ratio Lower Upper
95 100
174 72.4 0.0 157.4
176 70.2 0.0 154.6



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX026248.D
Acq: 04 Jan 2022 02:47


Tgt Ion: 117 Resp: 316486
Ion Ratio Lower Upper
117 100
82 58.1 43.8 65.6
119 31.7 25.2 37.8





#72
 1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 12.024 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VX026248.D
 Acq: 04 Jan 2022 02:47

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141783ZHE#06



Tgt Ion:152 Resp: 133060
 Ion Ratio Lower Upper
 152 100
 115 62.4 41.3 123.9
 150 158.5 0.0 351.8

