

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010324\
 Data File : VX039593.D
 Acq On : 03 Jan 2024 11:49
 Operator : JC/MD
 Sample : P1001-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ORA-1205

Quant Time: Jan 04 03:29:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	84540	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	179532	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	182071	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83117	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	72547	46.843	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.680%
35) Dibromofluoromethane	5.379	113	57676	46.502	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.000%
50) Toluene-d8	8.647	98	224037	48.655	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.300%
62) 4-Bromofluorobenzene	11.079	95	94706	49.849	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.700%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	3.026	59	9044	23.255	ug/l #	88
16) Acetone	2.386	43	59754	47.357	ug/l #	86
18) Methyl Acetate	2.703	43	9983	2.980	ug/l	98
25) 2-Butanone	4.562	43	45452	34.010	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010324\
Data File : VX039593.D
Acq On : 03 Jan 2024 11:49
Operator : JC/MD
Sample : P1001-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ORA-1205

Quant Time: Jan 04 03:29:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

