

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010324\
 Data File : VX039598.D
 Acq On : 03 Jan 2024 13:43
 Operator : JC/MD
 Sample : P1002-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1210

Quant Time: Jan 04 03:31:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	100698	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	210473	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	204811	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86067	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	87623	47.499	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	95.000%		
35) Dibromofluoromethane	5.385	113	67576	46.474	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.940%		
50) Toluene-d8	8.647	98	259223	48.020	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	96.040%		
62) 4-Bromofluorobenzene	11.079	95	106795	47.949	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	95.900%		
Target Compounds						Qvalue
11) Tert butyl alcohol	3.001	59	8344	18.012	ug/l #	92
16) Acetone	2.380	43	48025	23.607	ug/l #	85
18) Methyl Acetate	2.703	43	28523	7.149	ug/l	100
25) 2-Butanone	4.556	43	114842	72.144	ug/l	98
68) m/p-Xylenes	10.305	106	1082	0.363	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	4218	0.727	ug/l	70
95) Naphthalene	13.780	128	2115	0.336	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010324\
Data File : VX039598.D
Acq On : 03 Jan 2024 13:43
Operator : JC/MD
Sample : P1002-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1210

Quant Time: Jan 04 03:31:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

