

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010324\
 Data File : VX039610.D
 Acq On : 03 Jan 2024 18:15
 Operator : JC/MD
 Sample : P1002-11
 Misc : 1.09g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1132

Quant Time: Jan 04 03:37:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	94125	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	200898	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	208544	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	105107	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81162	47.069	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.140%
35) Dibromofluoromethane	5.391	113	62910	45.328	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.660%
50) Toluene-d8	8.647	98	249535	48.429	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.860%
62) 4-Bromofluorobenzene	11.079	95	114076	53.659	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.320%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.788	84	1631	1.135	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010324\
Data File : VX039610.D
Acq On : 03 Jan 2024 18:15
Operator : JC/MD
Sample : P1002-11
Misc : 1.09g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1132

Quant Time: Jan 04 03:37:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

