

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010324\
 Data File : VX039612.D
 Acq On : 03 Jan 2024 19:00
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jan 04 03:38:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	96584	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	206084	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	217684	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	110797	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	83125	46.981	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.960%
35) Dibromofluoromethane	5.385	113	66415	46.649	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.300%
50) Toluene-d8	8.647	98	259093	49.018	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.040%
62) 4-Bromofluorobenzene	11.079	95	119773	54.921	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.840%
Target Compounds						
					Qvalue	
70) Styrene	10.659	104	4663	1.040	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010324\
Data File : VX039612.D
Acq On : 03 Jan 2024 19:00
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jan 04 03:38:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

