

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010324\
 Data File : VX039588.D
 Acq On : 03 Jan 2024 09:42
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 04 03:26:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	84	0.00
2 T	Dichlorodifluoromethane	0.703	0.675	4.0	78	0.00
3 P	Chloromethane	0.779	0.773	0.8	83	0.00
4 C	Vinyl Chloride	0.724	0.763	-5.4#	90	0.00
5 T	Bromomethane	0.417	0.354	15.1	65	0.00
6 T	Chloroethane	0.437	0.368	15.8	75	-0.02
7 T	Trichlorofluoromethane	1.061	1.097	-3.4	87	-0.01
8 T	Diethyl Ether	0.367	0.408	-11.2	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.647	0.682	-5.4	90	-0.01
10 T	Methyl Iodide	0.669	0.690	-3.1	87	0.00
11 T	Tert butyl alcohol	0.230	0.208	9.6	79	0.05
12 CM	1,1-Dichloroethene	0.643	0.639	0.6#	84	0.00
13 T	Acrolein	0.183	0.248	-35.5#	116	0.00
14 T	Allyl chloride	1.303	1.286	1.3	84	0.00
15 T	Acrylonitrile	0.460	0.470	-2.2	86	0.00
16 T	Acetone	0.661	0.618	6.5	101	0.00
17 T	Carbon Disulfide	1.882	1.604	14.8	73	-0.01
18 T	Methyl Acetate	1.981	2.201	-11.1	92	0.00
19 T	Methyl tert-butyl Ether	2.325	2.367	-1.8	87	0.00
20 T	Methylene Chloride	0.764	0.756	1.0	86	0.00
21 T	trans-1,2-Dichloroethene	0.700	0.668	4.6	81	0.00
22 T	Diisopropyl ether	2.371	2.538	-7.0	91	0.00
23 T	Vinyl Acetate	1.765	1.822	-3.2	86	0.00
24 P	1,1-Dichloroethane	1.340	1.363	-1.7	88	0.00
25 T	2-Butanone	0.790	0.728	7.8	87	0.00
26 T	2,2-Dichloropropane	1.170	1.113	4.9	81	-0.01
27 T	cis-1,2-Dichloroethene	0.772	0.790	-2.3	85	0.00
28 T	Bromochloromethane	0.654	0.674	-3.1	82	0.00
29 T	Tetrahydrofuran	0.436	0.447	-2.5	86	0.00
30 C	Chloroform	1.339	1.352	-1.0#	87	0.00
31 T	Cyclohexane	1.183	1.178	0.4	84	-0.02
32 T	1,1,1-Trichloroethane	1.163	1.150	1.1	83	0.00
33 S	1,2-Dichloroethane-d4	0.916	0.841	8.2	83	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	82	0.00
35 S	Dibromofluoromethane	0.345	0.346	-0.3	87	-0.01
36 T	1,1-Dichloropropene	0.545	0.546	-0.2	83	0.00
37 T	Ethyl Acetate	0.673	0.710	-5.5	86	0.00
38 T	Carbon Tetrachloride	0.520	0.509	2.1	80	-0.01
39 T	Methylcyclohexane	0.655	0.682	-4.1	86	0.00
40 TM	Benzene	1.542	1.615	-4.7	86	0.00
41 T	Methacrylonitrile	0.359	0.384	-7.0	86	0.00
42 TM	1,2-Dichloroethane	0.624	0.630	-1.0	83	0.00
43 T	Isopropyl Acetate	1.095	1.138	-3.9	85	0.00
44 TM	Trichloroethene	0.389	0.388	0.3	84	0.00
45 C	1,2-Dichloropropane	0.414	0.437	-5.6#	87	0.00
46 T	Dibromomethane	0.291	0.297	-2.1	85	0.00
47 T	Bromodichloromethane	0.556	0.578	-4.0	84	0.00
48 T	Methyl methacrylate	0.640	0.667	-4.2	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010324\
 Data File : VX039588.D
 Acq On : 03 Jan 2024 09:42
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 04 03:26:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.010	-11.1	88	0.01
50 S	Toluene-d8	1.282	1.273	0.7	85	0.00
51 T	4-Methyl-2-Pentanone	0.698	0.758	-8.6	89	0.00
52 CM	Toluene	0.931	0.979	-5.2#	86	0.00
53 T	t-1,3-Dichloropropene	0.612	0.619	-1.1	81	0.00
54 T	cis-1,3-Dichloropropene	0.659	0.675	-2.4	82	0.00
55 T	1,1,2-Trichloroethane	0.389	0.409	-5.1	88	0.00
56 T	Ethyl methacrylate	0.474	0.526	-11.0	87	0.00
57 T	1,3-Dichloropropane	0.672	0.706	-5.1	87	0.00
58 T	2-Chloroethyl Vinyl ether	0.279	0.381	-36.6#	107	0.00
59 T	2-Hexanone	0.602	0.636	-5.6	92	0.00
60 T	Dibromochloromethane	0.393	0.403	-2.5	84	0.00
61 T	1,2-Dibromoethane	0.415	0.435	-4.8	85	0.00
62 S	4-Bromofluorobenzene	0.529	0.498	5.9	83	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	80	0.00
64 T	Tetrachloroethene	0.327	0.333	-1.8	86	0.00
65 PM	Chlorobenzene	1.091	1.124	-3.0	86	0.00
66 T	1,1,1,2-Tetrachloroethane	0.385	0.410	-6.5	86	0.00
67 C	Ethyl Benzene	2.031	2.136	-5.2#	85	0.00
68 T	m/p-Xylenes	0.728	0.770	-5.8	86	0.00
69 T	o-Xylene	0.704	0.763	-8.4	88	0.00
70 T	Styrene	1.030	1.126	-9.3	88	0.00
71 P	Bromoform	0.287	0.301	-4.9	82	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
73 T	Isopropylbenzene	3.858	4.283	-11.0	88	0.00
74 T	N-amyl acetate	1.952	2.272	-16.4	90	0.00
75 P	1,1,2,2-Tetrachloroethane	1.435	1.650	-15.0	92	0.00
76 T	1,2,3-Trichloropropane	1.294	1.403	-8.4	86	0.00
77 T	Bromobenzene	0.853	0.938	-10.0	88	0.00
78 T	n-propylbenzene	4.753	5.211	-9.6	87	0.00
79 T	2-Chlorotoluene	2.875	3.075	-7.0	86	0.00
80 T	1,3,5-Trimethylbenzene	3.286	3.690	-12.3	88	0.00
81 T	trans-1,4-Dichloro-2-butene	0.459	0.480	-4.6	81	0.00
82 T	4-Chlorotoluene	3.377	3.579	-6.0	85	0.00
83 T	tert-Butylbenzene	3.128	3.515	-12.4	89	0.00
84 T	1,2,4-Trimethylbenzene	3.370	3.797	-12.7	89	0.00
85 T	sec-Butylbenzene	3.980	4.450	-11.8	89	0.00
86 T	p-Isopropyltoluene	3.340	3.714	-11.2	88	0.00
87 T	1,3-Dichlorobenzene	1.676	1.805	-7.7	87	0.00
88 T	1,4-Dichlorobenzene	1.745	1.775	-1.7	86	0.00
89 T	n-Butylbenzene	3.283	3.586	-9.2	85	0.00
90 T	Hexachloroethane	0.603	0.660	-9.5	84	0.00
91 T	1,2-Dichlorobenzene	1.604	1.731	-7.9	87	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.378	0.395	-4.5	77	0.00
93 T	1,2,4-Trichlorobenzene	1.050	1.067	-1.6	80	0.00
94 T	Hexachlorobutadiene	0.396	0.456	-15.2	94	0.00
95 T	Naphthalene	3.657	4.088	-11.8	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010324\
Data File : VX039588.D
Acq On : 03 Jan 2024 09:42
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 04 03:26:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.014	1.068	-5.3	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6