

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX010419\
 Data File : VX006914.D
 Acq On : 04 Jan 2019 15:38
 Operator : JC/MD
 Sample : K1032-04DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S32-0646DL

Quant Time: Jan 04 22:44:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	644722	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	993573	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	902145	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	436451	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	339948	48.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.84%	
35) Dibromofluoromethane	5.51	113	306232	48.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.48%	
50) Toluene-d8	8.71	98	1155080	48.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.52%	
62) 4-Bromofluorobenzene	11.13	95	411023	47.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.76%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.41	62	9555	1.440	ug/l	93
16) Acetone	2.45	43	4547	1.424	ug/l	93
21) trans-1,2-Dichloroethene	3.17	96	53197	8.335	ug/l	97
27) cis-1,2-Dichloroethene	4.60	96	336162	47.737	ug/l	92
44) Trichloroethene	7.21	130	14166	1.871	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX010419\
Data File : VX006914.D
Acq On : 04 Jan 2019 15:38
Operator : JC/MD
Sample : K1032-04DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S32-0646DL

Quant Time: Jan 04 22:44:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

