

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX010419\
 Data File : VX006916.D
 Acq On : 04 Jan 2019 16:31
 Operator : JC/MD
 Sample : PB116150TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB116150TB

Quant Time: Jan 04 22:51:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	646932	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	990459	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	938205	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	445712	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	339412	48.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.36%	
35) Dibromofluoromethane	5.50	113	302066	48.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.46%	
50) Toluene-d8	8.71	98	1190637	50.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.84%	
62) 4-Bromofluorobenzene	11.13	95	420307	48.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.20%	
Target Compounds						
16) Acetone	2.45	43	24318	7.592	ug/l	99
20) Methylene Chloride	2.86	84	11298	1.706	ug/l #	86
43) Isopropyl Acetate	6.46	43	29336	2.258	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX010419\
Data File : VX006916.D
Acq On : 04 Jan 2019 16:31
Operator : JC/MD
Sample : PB116150TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB116150TB

Quant Time: Jan 04 22:51:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

