

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX010419\
 Data File : VX006917.D
 Acq On : 04 Jan 2019 16:58
 Operator : JC/MD
 Sample : K1040-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RBR-200030-RBR-200028-RT-3483

Quant Time: Jan 04 22:54:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	624857	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	965069	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	909221	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	436068	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	328550	48.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.58%	
35) Dibromofluoromethane	5.50	113	296178	48.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.06%	
50) Toluene-d8	8.71	98	1160605	50.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.88%	
62) 4-Bromofluorobenzene	11.13	95	410428	48.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.42%	
Target Compounds						
16) Acetone	2.45	43	33424	10.803	ug/l	99
18) Methyl Acetate	2.78	43	16093	1.769	ug/l	93
20) Methylene Chloride	2.86	84	941010	147.082	ug/l	98
43) Isopropyl Acetate	6.46	43	25262	1.996	ug/l	97
68) m/p-Xylenes	10.35	106	14297	1.153	ug/l	95
84) 1,2,4-Trimethylbenzene	11.80	105	26273	1.017	ug/l	99
95) Naphthalene	13.83	128	153438	4.793	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX010419\
Data File : VX006917.D
Acq On : 04 Jan 2019 16:58
Operator : JC/MD
Sample : K1040-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RBR-200030-RBR-200028-RT-3483

Quant Time: Jan 04 22:54:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

