

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010422\
 Data File : VX026265.D
 Acq On : 04 Jan 2022 12:00
 Operator : JC/MD
 Sample : VX0104MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0104MBL01

Quant Time: Jan 05 06:07:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	257978	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	445621	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	379665	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	159609	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	199977	53.582	ug/l	-0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.160%	
35) Dibromofluoromethane	5.385	113	147002	50.741	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.480%	
50) Toluene-d8	8.647	98	518894	48.152	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 96.300%	
62) 4-Bromofluorobenzene	11.079	95	186458	48.605	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.200%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010422\
Data File : VX026265.D
Acq On : 04 Jan 2022 12:00
Operator : JC/MD
Sample : VX0104MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0104MBL01

Quant Time: Jan 05 06:07:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

