

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010422\
 Data File : VX026266.D
 Acq On : 04 Jan 2022 12:55
 Operator : JC/MD
 Sample : VX0104WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0104WBL01

Quant Time: Jan 05 06:07:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	151914	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	259776	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	235242	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99278	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	114719	52.198	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.400%	
35) Dibromofluoromethane	5.385	113	84060	49.772	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.540%	
50) Toluene-d8	8.647	98	309280	49.233	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.460%	
62) 4-Bromofluorobenzene	11.079	95	117894	52.718	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.440%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010422\
Data File : VX026266.D
Acq On : 04 Jan 2022 12:55
Operator : JC/MD
Sample : VX0104WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0104WBL01

Quant Time: Jan 05 06:07:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

