

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010422\
 Data File : VX026269.D
 Acq On : 04 Jan 2022 14:18
 Operator : JC/MD
 Sample : M5335-01DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20968DL

Quant Time: Jan 05 06:09:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	158963	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	273120	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	240648	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100431	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	120612	52.446	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.900%
35) Dibromofluoromethane	5.391	113	88408	49.789	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.580%
50) Toluene-d8	8.653	98	326635	49.455	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.900%
62) 4-Bromofluorobenzene	11.079	95	119427	50.794	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.580%
Target Compounds						Qvalue
16) Acetone	2.380	43	86046	69.056	ug/l	97
51) 4-Methyl-2-Pentanone	8.574	43	398514	118.325	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010422\
Data File : VX026269.D
Acq On : 04 Jan 2022 14:18
Operator : JC/MD
Sample : M5335-01DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20968DL

Quant Time: Jan 05 06:09:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

