

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010422\
 Data File : VX026275.D
 Acq On : 04 Jan 2022 16:37
 Operator : JC/MD
 Sample : M5333-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Jan 05 06:11:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	143013	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	247701	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	222757	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101732	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	111278	53.784	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 107.560%		
35) Dibromofluoromethane	5.391	113	80936	50.259	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.520%		
50) Toluene-d8	8.653	98	294300	49.132	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 98.260%		
62) 4-Bromofluorobenzene	11.079	95	112409	52.715	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 105.440%		
Target Compounds						
16) Acetone	2.380	43	2426	2.164	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010422\
Data File : VX026275.D
Acq On : 04 Jan 2022 16:37
Operator : JC/MD
Sample : M5333-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Jan 05 06:11:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

