

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010422\
 Data File : VX026278.D
 Acq On : 04 Jan 2022 17:46
 Operator : JC/MD
 Sample : M5333-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-6

Quant Time: Jan 05 06:12:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	156440	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	268237	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	231844	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98559	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	119552	52.824	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.640%	
35) Dibromofluoromethane	5.391	113	87476	50.161	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.320%	
50) Toluene-d8	8.653	98	313134	48.274	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 96.540%	
62) 4-Bromofluorobenzene	11.079	95	114043	49.387	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.780%	
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.953	59	1299	2.874	ug/l	# 73
16) Acetone	2.380	43	2114	1.724	ug/l	97
19) Methyl tert-butyl Ether	3.117	73	70878	11.714	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010422\
Data File : VX026278.D
Acq On : 04 Jan 2022 17:46
Operator : JC/MD
Sample : M5333-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-6

Quant Time: Jan 05 06:12:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

