

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010422\
 Data File : VX026280.D
 Acq On : 04 Jan 2022 18:33
 Operator : JC/MD
 Sample : M5333-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-4

Quant Time: Jan 05 06:13:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	149109	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	256906	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	220760	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94427	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	115925	53.739	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.480%	
35) Dibromofluoromethane	5.391	113	84502	50.593	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.180%	
50) Toluene-d8	8.653	98	296961	47.800	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 95.600%	
62) 4-Bromofluorobenzene	11.079	95	110215	49.835	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.660%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	2514	2.151	ug/l	97
19) Methyl tert-butyl Ether	3.117	73	842884	146.150	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010422\
Data File : VX026280.D
Acq On : 04 Jan 2022 18:33
Operator : JC/MD
Sample : M5333-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4

Quant Time: Jan 05 06:13:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

