

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010422\
 Data File : VX026291.D
 Acq On : 04 Jan 2022 23:56
 Operator : JC/MD
 Sample : VX0104WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0104WBL02

Quant Time: Jan 05 06:19:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	163890	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	275064	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	240908	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100864	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	119853	50.549	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 101.100%	
35) Dibromofluoromethane	5.391	113	89198	49.879	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.760%	
50) Toluene-d8	8.653	98	326258	49.049	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.100%	
62) 4-Bromofluorobenzene	11.079	95	119013	50.260	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.520%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010422\
Data File : VX026291.D
Acq On : 04 Jan 2022 23:56
Operator : JC/MD
Sample : VX0104WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0104WBL02

Quant Time: Jan 05 06:19:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

