

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010422\
 Data File : VX026294.D
 Acq On : 05 Jan 2022 01:05
 Operator : JC/MD
 Sample : PB141821TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141821TB

Quant Time: Jan 05 06:41:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	158237	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	269335	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	232868	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96011	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	117535	51.343	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.680%	
35) Dibromofluoromethane	5.391	113	88111	50.319	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.640%	
50) Toluene-d8	8.653	98	312022	47.906	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 95.820%	
62) 4-Bromofluorobenzene	11.079	95	115194	49.682	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.360%	
Target Compounds						Qvalue
16) Acetone	2.380	43	8376	6.753	ug/l	90
20) Methylene Chloride	2.794	84	26663	12.740	ug/l	97
43) Isopropyl Acetate	6.342	43	16373	3.068	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010422\
Data File : VX026294.D
Acq On : 05 Jan 2022 01:05
Operator : JC/MD
Sample : PB141821TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141821TB

Quant Time: Jan 05 06:41:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

