

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010422\
 Data File : VX026295.D
 Acq On : 05 Jan 2022 01:28
 Operator : JC/MD
 Sample : PB141821ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141821ZHE#01

Quant Time: Jan 05 06:42:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	152599	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	262749	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	226062	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97396	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	116661	52.844	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 105.680%			
35) Dibromofluoromethane	5.391	113	85683	50.159	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.320%			
50) Toluene-d8	8.653	98	305567	48.091	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 96.180%			
62) 4-Bromofluorobenzene	11.079	95	113166	50.031	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 100.060%			
Target Compounds						
16) Acetone	2.380	43	8342	6.974	ug/l	97
20) Methylene Chloride	2.794	84	25390	12.568	ug/l	95
43) Isopropyl Acetate	6.348	43	16263	3.124	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010422\
Data File : VX026295.D
Acq On : 05 Jan 2022 01:28
Operator : JC/MD
Sample : PB141821ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141821ZHE#01

Quant Time: Jan 05 06:42:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

