

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010422\
 Data File : VX026296.D
 Acq On : 05 Jan 2022 01:51
 Operator : JC/MD
 Sample : PB141821ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141821ZHE#02

Quant Time: Jan 05 06:42:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	151385	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	261540	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	226873	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98008	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	116452	53.172	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 106.340%			
35) Dibromofluoromethane	5.391	113	85772	50.444	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.880%			
50) Toluene-d8	8.653	98	305405	48.288	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 96.580%			
62) 4-Bromofluorobenzene	11.079	95	114936	51.048	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 102.100%			
Target Compounds						
16) Acetone	2.379	43	8456	7.126	ug/l	93
20) Methylene Chloride	2.794	84	25578	12.777	ug/l	95
43) Isopropyl Acetate	6.348	43	15672	3.024	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010422\
Data File : VX026296.D
Acq On : 05 Jan 2022 01:51
Operator : JC/MD
Sample : PB141821ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141821ZHE#02

Quant Time: Jan 05 06:42:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

