

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010422\
 Data File : VX026299.D
 Acq On : 05 Jan 2022 03:01
 Operator : JC/MD
 Sample : PB141821ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141821ZHE#05

Quant Time: Jan 05 06:43:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	154996	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	268739	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	232594	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99297	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	118000	52.624	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.240%	
35) Dibromofluoromethane	5.391	113	86615	49.575	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.140%	
50) Toluene-d8	8.653	98	311804	47.979	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 95.960%	
62) 4-Bromofluorobenzene	11.079	95	116883	50.523	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.040%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	13868	11.415	ug/l	90
20) Methylene Chloride	2.794	84	8416	3.452	ug/l	98
43) Isopropyl Acetate	6.342	43	16841	3.163	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010422\
Data File : VX026299.D
Acq On : 05 Jan 2022 03:01
Operator : JC/MD
Sample : PB141821ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141821ZHE#05

Quant Time: Jan 05 06:43:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

