

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010422\
 Data File : VX026301.D
 Acq On : 05 Jan 2022 03:47
 Operator : JC/MD
 Sample : PB141821ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141821ZHE#07

Quant Time: Jan 05 06:43:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	145509	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	251048	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	221196	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99082	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	113139	53.746	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 107.500%			
35) Dibromofluoromethane	5.391	113	82921	50.805	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.600%			
50) Toluene-d8	8.653	98	295992	48.755	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 97.520%			
62) 4-Bromofluorobenzene	11.079	95	114275	52.876	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 105.760%			
Target Compounds					Qvalue	
16) Acetone	2.380	43	13517	11.851	ug/l	98
20) Methylene Chloride	2.794	84	8261	3.654	ug/l	96
43) Isopropyl Acetate	6.342	43	15529	3.122	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010422\
Data File : VX026301.D
Acq On : 05 Jan 2022 03:47
Operator : JC/MD
Sample : PB141821ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141821ZHE#07

Quant Time: Jan 05 06:43:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

