

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010422\
 Data File : VX026303.D
 Acq On : 05 Jan 2022 04:33
 Operator : JC/MD
 Sample : PB141821ZHE#09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141821ZHE#09

Quant Time: Jan 05 06:44:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	147586	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	256087	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	220426	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95387	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	114941	53.833	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.660%	
35) Dibromofluoromethane	5.391	113	84375	50.679	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.360%	
50) Toluene-d8	8.653	98	298165	48.147	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 96.300%	
62) 4-Bromofluorobenzene	11.079	95	110220	49.996	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.000%	
Target Compounds						Qvalue
16) Acetone	2.380	43	13647	11.797	ug/l	92
20) Methylene Chloride	2.794	84	8587	3.768	ug/l	88
43) Isopropyl Acetate	6.342	43	15810	3.116	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010422\
Data File : VX026303.D
Acq On : 05 Jan 2022 04:33
Operator : JC/MD
Sample : PB141821ZHE#09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141821ZHE#09

Quant Time: Jan 05 06:44:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

