

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX010422\
 Data File : VX026288.D
 Acq On : 04 Jan 2022 22:01
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 05 06:18:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	113	0.00
2 T	Dichlorodifluoromethane	50.000	42.504	15.0	92	0.00
3 P	Chloromethane	50.000	36.856	26.3#	83	0.00
4 C	Vinyl Chloride	50.000	37.385	25.2#	84	0.00
5 T	Bromomethane	50.000	24.465	51.1#	54	0.00
6 T	Chloroethane	50.000	33.843	32.3#	79	0.00
7 T	Trichlorofluoromethane	50.000	34.922	30.2#	77	0.00
8 T	Diethyl Ether	50.000	37.175	25.7#	82	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	40.699	18.6	91	0.00
10 T	Methyl Iodide	50.000	26.779	46.4#	58	0.00
11 T	Tert butyl alcohol	250.000	192.968	22.8	94	-0.02
12 CM	1,1-Dichloroethene	50.000	42.051	15.9#	95	0.00
13 T	Acrolein	250.000	204.121	18.4	102	0.00
14 T	Allyl chloride	50.000	39.918	20.2	89	0.00
15 T	Acrylonitrile	250.000	199.344	20.3	89	0.00
16 T	Acetone	250.000	168.673	32.5#	79	0.00
17 T	Carbon Disulfide	50.000	41.628	16.7	95	0.00
18 T	Methyl Acetate	50.000	39.585	20.8	89	0.00
19 T	Methyl tert-butyl Ether	50.000	45.433	9.1	100	0.00
20 T	Methylene Chloride	50.000	41.796	16.4	93	0.00
21 T	trans-1,2-Dichloroethene	50.000	41.749	16.5	96	0.00
22 T	Diisopropyl ether	50.000	42.887	14.2	94	0.00
23 T	Vinyl Acetate	250.000	215.105	14.0	92	0.00
24 P	1,1-Dichloroethane	50.000	42.533	14.9	94	0.00
25 T	2-Butanone	250.000	190.387	23.8	86	0.00
26 T	2,2-Dichloropropane	50.000	43.221	13.6	96	0.00
27 T	cis-1,2-Dichloroethene	50.000	43.035	13.9	96	0.00
28 T	Bromochloromethane	50.000	41.484	17.0	95	0.00
29 T	Tetrahydrofuran	250.000	194.606	22.2	86	0.00
30 C	Chloroform	50.000	41.968	16.1#	94	0.00
31 T	Cyclohexane	50.000	41.928	16.1	92	0.00
32 T	1,1,1-Trichloroethane	50.000	43.581	12.8	97	0.00
33 S	1,2-Dichloroethane-d4	50.000	42.970	14.1	105	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	115	0.00
35 S	Dibromofluoromethane	50.000	44.256	11.5	107	0.00
36 T	1,1-Dichloropropene	50.000	42.532	14.9	94	0.00
37 T	Ethyl Acetate	50.000	40.350	19.3	88	0.00
38 T	Carbon Tetrachloride	50.000	40.047	19.9	89	0.00
39 T	Methylcyclohexane	50.000	42.764	14.5	94	0.00
40 TM	Benzene	50.000	42.850	14.3	96	0.00
41 T	Methacrylonitrile	50.000	41.519	17.0	90	0.00
42 TM	1,2-Dichloroethane	50.000	41.637	16.7	94	0.00
43 T	Isopropyl Acetate	50.000	42.594	14.8	94	0.00
44 TM	Trichloroethene	50.000	45.167	9.7	102	0.00
45 C	1,2-Dichloropropane	50.000	43.075	13.8#	94	0.00
46 T	Dibromomethane	50.000	40.888	18.2	92	0.00
47 T	Bromodichloromethane	50.000	40.185	19.6	90	0.00
48 T	Methyl methacrylate	50.000	42.937	14.1	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010422\
 Data File : VX026288.D
 Acq On : 04 Jan 2022 22:01
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 05 06:18:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	696.968	30.3#	76	-0.03
50 S	Toluene-d8	50.000	43.606	12.8	106	0.00
51 T	4-Methyl-2-Pentanone	250.000	198.603	20.6	86	0.00
52 CM	Toluene	50.000	44.610	10.8#	98	0.00
53 T	t-1,3-Dichloropropene	50.000	39.285	21.4	92	0.00
54 T	cis-1,3-Dichloropropene	50.000	42.559	14.9	92	0.00
55 T	1,1,2-Trichloroethane	50.000	41.686	16.6	93	0.00
56 T	Ethyl methacrylate	50.000	46.474	7.1	99	0.00
57 T	1,3-Dichloropropane	50.000	43.027	13.9	96	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	218.304	12.7	101	0.00
59 T	2-Hexanone	250.000	195.812	21.7	84	0.00
60 T	Dibromochloromethane	50.000	40.162	19.7	89	0.00
61 T	1,2-Dibromoethane	50.000	42.370	15.3	94	0.00
62 S	4-Bromofluorobenzene	50.000	45.891	8.2	111	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	112	0.00
64 T	Tetrachloroethene	50.000	49.542	0.9	112	0.00
65 PM	Chlorobenzene	50.000	43.586	12.8	95	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	43.092	13.8	92	0.00
67 C	Ethyl Benzene	50.000	45.326	9.3#	96	0.00
68 T	m/p-Xylenes	100.000	92.134	7.9	97	0.00
69 T	o-Xylene	50.000	45.513	9.0	95	0.00
70 T	Styrene	50.000	46.030	7.9	94	0.00
71 P	Bromoform	50.000	37.581	24.8	79	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	106	0.00
73 T	Isopropylbenzene	50.000	46.835	6.3	93	0.00
74 T	N-amyl acetate	50.000	48.204	3.6	92	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	40.063	19.9	83	0.00
76 T	1,2,3-Trichloropropane	50.000	42.606	14.8	88	0.00
77 T	Bromobenzene	50.000	46.226	7.5	92	0.00
78 T	n-propylbenzene	50.000	46.927	6.1	91	0.00
79 T	2-Chlorotoluene	50.000	46.585	6.8	94	0.00
80 T	1,3,5-Trimethylbenzene	50.000	47.210	5.6	93	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	31.757	36.5#	64	0.00
82 T	4-Chlorotoluene	50.000	46.979	6.0	92	0.00
83 T	tert-Butylbenzene	50.000	45.330	9.3	89	0.00
84 T	1,2,4-Trimethylbenzene	50.000	48.041	3.9	95	0.00
85 T	sec-Butylbenzene	50.000	45.961	8.1	90	0.00
86 T	p-Isopropyltoluene	50.000	46.876	6.2	92	0.00
87 T	1,3-Dichlorobenzene	50.000	45.995	8.0	91	0.00
88 T	1,4-Dichlorobenzene	50.000	42.944	14.1	88	0.00
89 T	n-Butylbenzene	50.000	45.052	9.9	88	0.00
90 T	Hexachloroethane	50.000	34.477	31.0#	69	0.00
91 T	1,2-Dichlorobenzene	50.000	42.811	14.4	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	38.336	23.3	79	0.00
93 T	1,2,4-Trichlorobenzene	50.000	44.425	11.2	89	0.00
94 T	Hexachlorobutadiene	50.000	43.166	13.7	87	0.00
95 T	Naphthalene	50.000	45.407	9.2	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010422\
Data File : VX026288.D
Acq On : 04 Jan 2022 22:01
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 05 06:18:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	44.125	11.8	87	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6