

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010424\
 Data File : VX039618.D
 Acq On : 04 Jan 2024 10:48
 Operator : JC/MD
 Sample : VX0104WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0104WBL01

Quant Time: Jan 05 00:37:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	89065	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	188171	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	205549	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	99611	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	76865	47.110	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.220%
35) Dibromofluoromethane	5.385	113	62655	48.197	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.400%
50) Toluene-d8	8.647	98	241720	50.085	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.160%
62) 4-Bromofluorobenzene	11.079	95	111621	56.055	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	112.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010424\
Data File : VX039618.D
Acq On : 04 Jan 2024 10:48
Operator : JC/MD
Sample : VX0104WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0104WBL01

Quant Time: Jan 05 00:37:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

