

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX010424\  
 Data File : VX039625.D  
 Acq On : 04 Jan 2024 13:58  
 Operator : JC/MD  
 Sample : IBLK  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 IBLK

Quant Time: Jan 05 00:44:13 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X121323W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Dec 14 01:28:15 2023  
 Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 5.544  | 168  | 81672    | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 6.757  | 114  | 174420   | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 10.055 | 117  | 189353   | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 12.018 | 152  | 93348    | 50.000 | ug/l  | 0.00     |

|                             |        |       |          |          |      |          |
|-----------------------------|--------|-------|----------|----------|------|----------|
| System Monitoring Compounds |        |       |          |          |      |          |
| 33) 1,2-Dichloroethane-d4   | 5.952  | 65    | 70240    | 46.946   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =    | 93.900%  |
| 35) Dibromofluoromethane    | 5.385  | 113   | 57743    | 47.920   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =    | 95.840%  |
| 50) Toluene-d8              | 8.647  | 98    | 221152   | 49.436   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =    | 98.880%  |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 102278   | 55.413   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =    | 110.820% |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX010424\  
Data File : VX039625.D  
Acq On : 04 Jan 2024 13:58  
Operator : JC/MD  
Sample : IBLK  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
IBLK

Quant Time: Jan 05 00:44:13 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X121323W.M  
Quant Title : SW846 8260  
QLast Update : Thu Dec 14 01:28:15 2023  
Response via : Initial Calibration

