

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010424\
 Data File : VX039633.D
 Acq On : 04 Jan 2024 20:05
 Operator : JC/MD
 Sample : P1034-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1

Quant Time: Jan 05 00:47:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	81922	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	176409	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	190738	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98672	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	71106	47.380	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.760%
35) Dibromofluoromethane	5.385	113	57213	46.945	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.900%
50) Toluene-d8	8.647	98	223002	49.287	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.580%
62) 4-Bromofluorobenzene	11.079	95	104037	55.730	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	111.460%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.319	96	443	0.421	ug/l	83
30) Chloroform	5.093	83	2343	1.068	ug/l #	76
44) Trichloroethene	7.135	130	1291	0.940	ug/l	73
64) Tetrachloroethene	9.275	164	6092	4.888	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010424\
Data File : VX039633.D
Acq On : 04 Jan 2024 20:05
Operator : JC/MD
Sample : P1034-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1

Quant Time: Jan 05 00:47:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

