

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX010424\
 Data File : VX039635.D
 Acq On : 04 Jan 2024 20:51
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 05 00:48:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	80	0.00
2 T	Dichlorodifluoromethane	50.000	44.584	10.8	69	0.00
3 P	Chloromethane	50.000	45.430	9.1	72	0.00
4 C	Vinyl Chloride	50.000	47.700	4.6#	77	0.00
5 T	Bromomethane	50.000	50.151	-0.3	73	0.00
6 T	Chloroethane	50.000	46.648	6.7	78	0.00
7 T	Trichlorofluoromethane	50.000	47.171	5.7	76	0.00
8 T	Diethyl Ether	50.000	51.777	-3.6	83	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	47.549	4.9	77	0.00
10 T	Methyl Iodide	50.000	46.059	7.9	74	0.00
11 T	Tert butyl alcohol	250.000	206.439	17.4	69	0.00
12 CM	1,1-Dichloroethene	50.000	46.553	6.9#	75	0.00
13 T	Acrolein	250.000	244.783	2.1	80	0.00
14 T	Allyl chloride	50.000	46.425	7.2	75	0.00
15 T	Acrylonitrile	250.000	251.422	-0.6	80	0.00
16 T	Acetone	250.000	185.834	25.7#	64	0.00
17 T	Carbon Disulfide	50.000	40.467	19.1	66	0.00
18 T	Methyl Acetate	50.000	53.567	-7.1	84	0.00
19 T	Methyl tert-butyl Ether	50.000	48.219	3.6	79	0.00
20 T	Methylene Chloride	50.000	47.075	5.8	78	0.00
21 T	trans-1,2-Dichloroethene	50.000	46.365	7.3	75	0.00
22 T	Diisopropyl ether	50.000	51.604	-3.2	84	0.00
23 T	Vinyl Acetate	250.000	245.149	1.9	77	0.00
24 P	1,1-Dichloroethane	50.000	48.251	3.5	79	0.00
25 T	2-Butanone	250.000	207.852	16.9	75	0.00
26 T	2,2-Dichloropropane	50.000	41.788	16.4	68	0.00
27 T	cis-1,2-Dichloroethene	50.000	49.782	0.4	79	0.00
28 T	Bromochloromethane	50.000	51.213	-2.4	78	0.00
29 T	Tetrahydrofuran	250.000	247.264	1.1	79	0.00
30 C	Chloroform	50.000	48.735	2.5#	80	0.00
31 T	Cyclohexane	50.000	47.046	5.9	76	0.00
32 T	1,1,1-Trichloroethane	50.000	47.119	5.8	75	0.00
33 S	1,2-Dichloroethane-d4	50.000	47.433	5.1	81	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	80	0.00
35 S	Dibromofluoromethane	50.000	50.414	-0.8	86	0.00
36 T	1,1-Dichloropropene	50.000	47.260	5.5	77	0.00
37 T	Ethyl Acetate	50.000	49.143	1.7	79	0.00
38 T	Carbon Tetrachloride	50.000	46.105	7.8	73	0.00
39 T	Methylcyclohexane	50.000	46.333	7.3	75	0.00
40 TM	Benzene	50.000	48.877	2.2	79	0.00
41 T	Methacrylonitrile	50.000	49.961	0.1	78	0.00
42 TM	1,2-Dichloroethane	50.000	46.574	6.9	75	0.00
43 T	Isopropyl Acetate	50.000	48.763	2.5	78	0.00
44 TM	Trichloroethene	50.000	46.853	6.3	77	0.00
45 C	1,2-Dichloropropane	50.000	50.190	-0.4#	81	0.00
46 T	Dibromomethane	50.000	47.946	4.1	78	0.00
47 T	Bromodichloromethane	50.000	48.723	2.6	77	0.00
48 T	Methyl methacrylate	50.000	49.703	0.6	78	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010424\
 Data File : VX039635.D
 Acq On : 04 Jan 2024 20:51
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 05 00:48:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	978.769	2.1	73	0.00
50 S	Toluene-d8	50.000	50.069	-0.1	84	0.00
51 T	4-Methyl-2-Pentanone	250.000	252.047	-0.8	81	0.00
52 CM	Toluene	50.000	48.489	3.0#	78	0.00
53 T	t-1,3-Dichloropropene	50.000	46.395	7.2	72	0.00
54 T	cis-1,3-Dichloropropene	50.000	47.380	5.2	75	0.00
55 T	1,1,2-Trichloroethane	50.000	50.385	-0.8	82	0.00
56 T	Ethyl methacrylate	50.000	52.019	-4.0	80	0.00
57 T	1,3-Dichloropropane	50.000	49.900	0.2	81	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	270.807	-8.3	83	0.00
59 T	2-Hexanone	250.000	231.786	7.3	79	0.00
60 T	Dibromochloromethane	50.000	48.782	2.4	78	0.00
61 T	1,2-Dibromoethane	50.000	49.790	0.4	79	0.00
62 S	4-Bromofluorobenzene	50.000	49.214	1.6	85	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	82	0.00
64 T	Tetrachloroethene	50.000	45.651	8.7	79	0.00
65 PM	Chlorobenzene	50.000	46.604	6.8	80	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	47.414	5.2	79	0.00
67 C	Ethyl Benzene	50.000	47.652	4.7#	79	0.00
68 T	m/p-Xylenes	100.000	94.673	5.3	79	0.00
69 T	o-Xylene	50.000	48.251	3.5	80	0.00
70 T	Styrene	50.000	49.067	1.9	81	0.00
71 P	Bromoform	50.000	47.017	6.0	75	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	82	0.00
73 T	Isopropylbenzene	50.000	48.084	3.8	79	0.00
74 T	N-ethyl acetate	50.000	50.141	-0.3	81	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	49.519	1.0	83	0.00
76 T	1,2,3-Trichloropropane	50.000	46.509	7.0	77	0.00
77 T	Bromobenzene	50.000	47.625	4.8	79	0.00
78 T	n-propylbenzene	50.000	47.621	4.8	79	0.00
79 T	2-Chlorotoluene	50.000	46.501	7.0	79	0.00
80 T	1,3,5-Trimethylbenzene	50.000	48.353	3.3	80	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	42.160	15.7	69	0.00
82 T	4-Chlorotoluene	50.000	46.833	6.3	79	0.00
83 T	tert-Butylbenzene	50.000	48.081	3.8	80	0.00
84 T	1,2,4-Trimethylbenzene	50.000	47.758	4.5	79	0.00
85 T	sec-Butylbenzene	50.000	48.759	2.5	82	0.00
86 T	p-Isopropyltoluene	50.000	48.424	3.2	80	0.00
87 T	1,3-Dichlorobenzene	50.000	47.086	5.8	80	0.00
88 T	1,4-Dichlorobenzene	50.000	45.629	8.7	80	0.00
89 T	n-Butylbenzene	50.000	49.088	1.8	80	0.00
90 T	Hexachloroethane	50.000	47.613	4.8	76	0.00
91 T	1,2-Dichlorobenzene	50.000	47.161	5.7	80	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	45.349	9.3	70	0.00
93 T	1,2,4-Trichlorobenzene	50.000	47.584	4.8	79	0.00
94 T	Hexachlorobutadiene	50.000	49.562	0.9	85	0.00
95 T	Naphthalene	50.000	50.084	-0.2	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010424\
Data File : VX039635.D
Acq On : 04 Jan 2024 20:51
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 05 00:48:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	48.277	3.4	81	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6