

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX010521\
 Data File : VX020506.D
 Acq On : 05 Jan 2021 18:53
 Operator : JC/MD
 Sample : VX0105WBL01
 Misc : 25.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK304

Quant Time: Jan 06 05:09:59 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SFAMXTR010521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jan 06 04:56:09 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.82	114	155699	5.00	ug/L	0.00
28) Chlorobenzene-d5	10.10	117	161029	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.06	152	70877	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	61073	4.87	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.40%
7) Chloroethane-d5	1.69	69	54989	5.02	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.40%
11) 1,1-Dichloroethene-d2	2.34	63	94693	3.90	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.00%
20) 2-Butanone-d5	4.54	46	140502	51.77	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.54%
24) Chloroform-d	5.14	84	119908	4.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.20%
26) 1,2-Dichloroethane-d4	6.03	65	60489	5.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.00%
32) Benzene-d6	6.04	84	249968	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.00%
36) 1,2-Dichloropropane-d6	7.37	67	74724	4.73	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.60%
41) Toluene-d8	8.70	98	209045	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.60%
43) trans-1,3-Dichloropropene-	8.99	79	24798	4.54	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.80%
46) 2-Hexanone-d5	9.43	63	126328	51.07	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.14%
56) 1,1,2,2-Tetrachloroethane-	11.23	84	52902	4.75	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.00%
66) 1,2-Dichlorobenzene-d4	12.36	152	75930	5.49	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX010521\
Data File : VX020506.D
Acq On : 05 Jan 2021 18:53
Operator : JC/MD
Sample : VX0105WBL01
Misc : 25.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK304

Quant Time: Jan 06 05:09:59 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\SFAMXTR010521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jan 06 04:56:09 2021
Response via : Initial Calibration

