

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010522\
 Data File : VX026344.D
 Acq On : 06 Jan 2022 00:29
 Operator : JC/MD
 Sample : VX0105WBSD02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jan 06 03:14:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	105254	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	175288	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	153633	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	72315	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	78866	51.793	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.580%	
35) Dibromofluoromethane	5.391	113	58279	51.140	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.280%	
50) Toluene-d8	8.653	98	211419	49.876	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.760%	
62) 4-Bromofluorobenzene	11.079	95	81337	53.901	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 107.800%	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.172	85	21002	18.999	ug/l	99
3) Chloromethane	1.294	50	19360	15.045	ug/l	97
4) Vinyl Chloride	1.380	62	21911	16.180	ug/l	100
5) Bromomethane	1.618	94	10475	12.534	ug/l	98
6) Chloroethane	1.691	64	13605	14.287	ug/l	97
7) Trichlorofluoromethane	1.892	101	36006	15.435	ug/l	94
8) Diethyl Ether	2.136	74	13101	16.348	ug/l	94
9) 1,1,2-Trichlorotrifluo...	2.337	101	22619	18.393	ug/l	98
10) Methyl Iodide	2.459	142	22833	14.357	ug/l	97
11) Tert butyl alcohol	2.983	59	26637	87.582	ug/l	99
12) 1,1-Dichloroethene	2.325	96	21399	18.360	ug/l	95
13) Acrolein	2.245	56	8775	80.687	ug/l	99
14) Allyl chloride	2.672	41	40291	17.493	ug/l	97
15) Acrylonitrile	3.068	53	66941	88.899	ug/l	98
16) Acetone	2.386	43	65375	79.240	ug/l	95
17) Carbon Disulfide	2.520	76	58927	17.617	ug/l	99
18) Methyl Acetate	2.709	43	48721	18.206	ug/l	97
19) Methyl tert-butyl Ether	3.117	73	81715	20.072	ug/l	99
20) Methylene Chloride	2.794	84	24334	17.838	ug/l	94
21) trans-1,2-Dichloroethene	3.099	96	23341	18.357	ug/l	91
22) Diisopropyl ether	3.763	45	83321	18.953	ug/l	96
23) Vinyl Acetate	3.727	43	322244	92.312	ug/l	98
24) 1,1-Dichloroethane	3.617	63	44967	18.874	ug/l	100
25) 2-Butanone	4.562	43	99690	86.743	ug/l	98
26) 2,2-Dichloropropane	4.483	77	31029	17.286	ug/l	98
27) cis-1,2-Dichloroethene	4.495	96	27201	18.804	ug/l	93
28) Bromochloromethane	4.910	49	19652	18.611	ug/l	91
29) Tetrahydrofuran	5.013	42	63798	88.073	ug/l	95
30) Chloroform	5.099	83	47485	18.844	ug/l	99
31) Cyclohexane	5.477	56	39951	18.423	ug/l	98
32) 1,1,1-Trichloroethane	5.391	97	41646	19.011	ug/l	99
36) 1,1-Dichloropropene	5.702	75	34361	18.694	ug/l	99
37) Ethyl Acetate	4.715	43	38250	17.947	ug/l	99
38) Carbon Tetrachloride	5.684	117	36618	17.910	ug/l	99
39) Methylcyclohexane	7.385	83	38715	18.673	ug/l	95
40) Benzene	6.044	78	96588	18.412	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010522\
 Data File : VX026344.D
 Acq On : 06 Jan 2022 00:29
 Operator : JC/MD
 Sample : VX0105WBSD02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jan 06 03:14:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	21898	18.195	ug/l	95
42) 1,2-Dichloroethane	6.092	62	39591	18.649	ug/l	97
43) Isopropyl Acetate	6.342	43	62645	18.035	ug/l	99
44) Trichloroethene	7.129	130	28251	20.014	ug/l	97
45) 1,2-Dichloropropane	7.434	63	25066	18.681	ug/l	97
46) Dibromomethane	7.586	93	17616	17.700	ug/l	99
47) Bromodichloromethane	7.824	83	34867	17.219	ug/l	100
48) Methyl methacrylate	7.696	41	31432	18.869	ug/l	93
49) 1,4-Dioxane	7.684	88	10560	343.988	ug/l #	89
51) 4-Methyl-2-Pentanone	8.574	43	197143	91.205	ug/l	99
52) Toluene	8.720	92	60520	18.740	ug/l	100
53) t-1,3-Dichloropropene	8.982	75	32138	16.252	ug/l	97
54) cis-1,3-Dichloropropene	8.366	75	36705	17.151	ug/l	98
55) 1,1,2-Trichloroethane	9.153	97	24914	18.771	ug/l	97
56) Ethyl methacrylate	9.116	69	40171	19.624	ug/l	94
57) 1,3-Dichloropropane	9.311	76	43207	19.004	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	102691	99.979	ug/l	97
59) 2-Hexanone	9.433	43	149810	91.496	ug/l	97
60) Dibromochloromethane	9.525	129	24904	16.563	ug/l	100
61) 1,2-Dibromoethane	9.610	107	26031	18.330	ug/l	99
64) Tetrachloroethene	9.275	164	29423	22.694	ug/l	96
65) Chlorobenzene	10.079	112	63262	18.745	ug/l	97
66) 1,1,1,2-Tetrachloroethane	10.165	131	23876	18.566	ug/l	99
67) Ethyl Benzene	10.195	91	118479	19.633	ug/l	96
68) m/p-Xylenes	10.305	106	88535	39.293	ug/l	93
69) o-Xylene	10.640	106	43255	19.620	ug/l	91
70) Styrene	10.659	104	71394	19.754	ug/l	96
71) Bromoform	10.799	173	15654	15.005	ug/l #	99
73) Isopropylbenzene	10.963	105	116504	20.726	ug/l	97
74) N-amyl acetate	10.841	43	50182	19.442	ug/l	97
75) 1,1,2,2-Tetrachloroethane	11.213	83	32661	17.085	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	38528	22.073	ug/l	87
77) Bromobenzene	11.201	156	26977	20.237	ug/l	91
78) n-propylbenzene	11.305	91	130850	20.246	ug/l	97
79) 2-Chlorotoluene	11.366	91	80753	20.298	ug/l	94
80) 1,3,5-Trimethylbenzene	11.451	105	96079	20.597	ug/l	97
81) trans-1,4-Dichloro-2-b...	11.018	75	6969	11.550	ug/l #	71
82) 4-Chlorotoluene	11.457	91	93492	20.809	ug/l	95
83) tert-Butylbenzene	11.719	119	88575	19.563	ug/l	92
84) 1,2,4-Trimethylbenzene	11.756	105	96617	20.729	ug/l	97
85) sec-Butylbenzene	11.890	105	113777	20.137	ug/l	99
86) p-Isopropyltoluene	12.012	119	96030	20.387	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	49890	19.937	ug/l	99
88) 1,4-Dichlorobenzene	12.042	146	50302	18.749	ug/l	98
89) n-Butylbenzene	12.335	91	81126	19.380	ug/l	99
90) Hexachloroethane	12.542	117	13055	13.877	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	48589	18.718	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	12.945	75	7895	16.133	ug/l	90
93) 1,2,4-Trichlorobenzene	13.591	180	29478	19.217	ug/l	99
94) Hexachlorobutadiene	13.725	225	12225	19.194	ug/l	99
95) Naphthalene	13.780	128	108015	19.326	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	29758	19.067	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010522\
Data File : VX026344.D
Acq On : 06 Jan 2022 00:29
Operator : JC/MD
Sample : VX0105WBSD02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :


Quant Time: Jan 06 03:14:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Jan 06 03:14:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:168 Resp: 105254
Ion Ratio Lower Upper
168 100
99 61.8 45.8 68.8



#2
Dichlorodifluoromethane
Concen: 18.999 ug/l
RT: 1.172 min Scan# 14
Delta R.T. -0.001 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29

Tgt Ion: 85 Resp: 21002
Ion Ratio Lower Upper
85 100
87 34.7 17.0 50.9









#3
Chloromethane
Concen: 15.045 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 50 Resp: 19360
Ion Ratio Lower Upper
50 100
52 33.2 25.4 38.0



#4
Vinyl Chloride
Concen: 16.180 ug/l
RT: 1.380 min Scan# 48
Delta R.T. 0.006 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29

Tgt Ion: 62 Resp: 21911
Ion Ratio Lower Upper
62 100
64 31.9 25.4 38.2









#5
Bromomethane
Concen: 12.534 ug/l
RT: 1.618 min Scan# 81
Delta R.T. 0.007 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 94 Resp: 10475
Ion Ratio Lower Upper
94 100
96 91.4 75.0 112.4



#6
Chloroethane
Concen: 14.287 ug/l
RT: 1.691 min Scan# 99
Delta R.T. -0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29


Tgt Ion: 64 Resp: 13605
Ion Ratio Lower Upper
64 100
66 34.0 25.9 38.9








#7
 Trichlorofluoromethane
 Concen: 15.435 ug/l
 RT: 1.892 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX026344.D
 Acq: 06 Jan 2022 00:29

Instrument :
 MSVOA_X
 ClientSampleId :







Tgt Ion:101 Resp: 36006
 Ion Ratio Lower Upper
 101 100
 103 62.6 54.1 81.1



#8
 Diethyl Ether
 Concen: 16.348 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. -0.000 min
 Lab File: VX026344.D
 Acq: 06 Jan 2022 00:29

Tgt Ion: 74 Resp: 13101
 Ion Ratio Lower Upper
 74 100
 45 107.1 50.4 151.2









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.393 ug/l
RT: 2.337 min Scan# 204
Delta R.T. 0.006 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 22619
Ion Ratio Lower Upper
101 100
85 43.6 36.5 54.7
151 74.9 60.6 90.8



#10
Methyl Iodide
Concen: 14.357 ug/l
RT: 2.459 min Scan# 225
Delta R.T. -0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29


Tgt Ion:142 Resp: 22833
Ion Ratio Lower Upper
142 100
127 49.3 37.5 56.3
141 14.7 11.4 17.0








#11
Tert butyl alcohol
Concen: 87.582 ug/l
RT: 2.983 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 59 Resp: 26637
Ion Ratio Lower Upper
59 100
57 10.7 8.2 12.2



#12
1,1-Dichloroethene
Concen: 18.360 ug/l
RT: 2.325 min Scan# 203
Delta R.T. -0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29

Tgt Ion: 96 Resp: 21399
Ion Ratio Lower Upper
96 100
61 165.9 127.5 191.3
98 60.8 52.0 78.0









#13
Acrolein
Concen: 80.687 ug/l
RT: 2.245 min Scan# 19
Delta R.T. 0.006 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 56 Resp: 8775
Ion Ratio Lower Upper
56 100
55 70.9 57.4 86.2



#14
Allyl chloride
Concen: 17.493 ug/l
RT: 2.672 min Scan# 260
Delta R.T. 0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29


Tgt Ion: 41 Resp: 40291
Ion Ratio Lower Upper
41 100
39 70.0 55.4 83.0
76 29.6 27.6 41.4








#15
Acrylonitrile
Concen: 88.899 ug/l
RT: 3.068 min Scan# 31
Delta R.T. -0.001 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 53 Resp: 66941
Ion Ratio Lower Upper
53 100
52 84.2 66.2 99.2
51 37.9 29.4 44.0



#16
Acetone
Concen: 79.240 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29

Tgt Ion: 43 Resp: 65375
Ion Ratio Lower Upper
43 100
58 28.5 24.8 37.2





#17

Carbon Disulfide

Concen: 17.617 ug/l

RT: 2.520 min Scan# 21

Delta R.T. 0.006 min

Lab File: VX026344.D

Acq: 06 Jan 2022 00:29

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 76 Resp: 58927

Ion Ratio Lower Upper

76 100

78 8.7 7.2 10.8



#18

Methyl Acetate

Concen: 18.206 ug/l

RT: 2.709 min Scan# 266

Delta R.T. -0.000 min

Lab File: VX026344.D




Acq: 06 Jan 2022 00:29


Tgt Ion: 43 Resp: 48721

Ion Ratio Lower Upper

43 100

74 22.6 19.1 28.7









#19
Methyl tert-butyl Ether
Concen: 20.072 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 73 Resp: 81715
Ion Ratio Lower Upper
73 100
57 22.0 17.0 25.6



#20
Methylene Chloride
Concen: 17.838 ug/l
RT: 2.794 min Scan# 280
Delta R.T. 0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29


Tgt Ion: 84 Resp: 24334
Ion Ratio Lower Upper
84 100
49 129.4 97.1 145.7
51 39.0 29.3 43.9
86 67.2 50.8 76.2








#21
trans-1,2-Dichloroethene
Concen: 18.357 ug/l
RT: 3.099 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 96 Resp: 23341
Ion Ratio Lower Upper
96 100
61 147.2 105.7 158.5
98 64.5 52.1 78.1




#22
Diisopropyl ether
Concen: 18.953 ug/l
RT: 3.763 min Scan# 439
Delta R.T. -0.007 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29



Tgt Ion: 45 Resp: 83321
Ion Ratio Lower Upper
45 100
43 76.7 63.8 95.6
87 26.0 24.2 36.4
59 11.6 9.4 14.2









#23
 Vinyl Acetate
 Concen: 92.312 ug/l
 RT: 3.727 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VX026344.D
 Acq: 06 Jan 2022 00:29




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 43 Resp: 322244
 Ion Ratio Lower Upper
 43 100
 86 10.4 9.0 13.6



#24
 1,1-Dichloroethane
 Concen: 18.874 ug/l
 RT: 3.617 min Scan# 415
 Delta R.T. 0.000 min
 Lab File: VX026344.D
 Acq: 06 Jan 2022 00:29

Tgt Ion: 63 Resp: 44967
 Ion Ratio Lower Upper
 63 100
 98 6.6 3.3 9.9
 100 4.3 2.2 6.6









#25
2-Butanone
Concen: 86.743 ug/l
RT: 4.562 min Scan# 51
Delta R.T. 0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 99690
Ion Ratio Lower Upper
43 100
72 25.3 21.0 31.4



#26
2,2-Dichloropropane
Concen: 17.286 ug/l
RT: 4.483 min Scan# 557
Delta R.T. -0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29

Tgt Ion: 77 Resp: 31029
Ion Ratio Lower Upper
77 100
97 22.2 11.7 35.0





#27

cis-1,2-Dichloroethene

Concen: 18.804 ug/l

RT: 4.495 min Scan# 51

Delta R.T. 0.000 min


Lab File: VX026344.D

Acq: 06 Jan 2022 00:29

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion: 96 Resp: 27201

Ion Ratio Lower Upper

96 100

61 150.7 0.0 276.8

98 64.6 0.0 130.6



#28

Bromochloromethane

Concen: 18.611 ug/l

RT: 4.910 min Scan# 627

Delta R.T. 0.006 min

Lab File: VX026344.D

Acq: 06 Jan 2022 00:29




Tgt Ion: 49 Resp: 19652


Ion Ratio Lower Upper

49 100

129 1.9 0.0 4.4

130 75.2 67.1 100.7





#29

Tetrahydrofuran

Concen: 88.073 ug/l

RT: 5.013 min Scan# 64

Delta R.T. 0.000 min


Lab File: VX026344.D

Acq: 06 Jan 2022 00:29

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion: 42 Resp: 63798

Ion Ratio Lower Upper

42 100

72 42.9 37.0 55.6

71 39.7 34.1 51.1



#30

Chloroform

Concen: 18.844 ug/l

RT: 5.099 min Scan# 658

Delta R.T. -0.006 min

Lab File: VX026344.D




Acq: 06 Jan 2022 00:29


Tgt Ion: 83 Resp: 47485

Ion Ratio Lower Upper

83 100


85 64.3 50.6 75.8








#31
Cyclohexane
Concen: 18.423 ug/l
RT: 5.477 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 56 Resp: 39951
Ion Ratio Lower Upper
56 100
69 30.6 23.9 35.9
84 89.3 73.2 109.8



#32
1,1,1-Trichloroethane
Concen: 19.011 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.006 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29

Tgt Ion: 97 Resp: 41646
Ion Ratio Lower Upper
97 100
99 64.0 51.5 77.3
61 43.9 35.5 53.3





#33

1,2-Dichloroethane-d4

Concen: 51.793 ug/l

RT: 5.958 min Scan# 799

Delta R.T. -0.006 min

Lab File: VX026344.D

Acq: 06 Jan 2022 00:29

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 65 Resp: 78866

Ion Ratio Lower Upper

65 100

67 49.9 0.0 103.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.006 min

Lab File: VX026344.D

Acq: 06 Jan 2022 00:29




Tgt Ion: 114 Resp: 175288


Ion Ratio Lower Upper

114 100

63 22.1 0.0 40.0

88 17.1 0.0 30.6





#35

Dibromofluoromethane

Concen: 51.140 ug/l

RT: 5.391 min Scan# 706

Delta R.T. 0.000 min


Lab File: VX026344.D

Acq: 06 Jan 2022 00:29

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion:113 Resp: 58279

Ion Ratio Lower Upper

113 100

111 102.9 81.8 122.8

192 19.2 15.0 22.4



#36

1,1-Dichloropropene

Concen: 18.694 ug/l

RT: 5.702 min Scan# 757

Delta R.T. 0.000 min

Lab File: VX026344.D

Acq: 06 Jan 2022 00:29




Tgt Ion: 75 Resp: 34361


Ion Ratio Lower Upper

75 100

110 34.9 18.2 54.6

77 30.9 24.7 37.1





#37

Ethyl Acetate

Concen: 17.947 ug/l

RT: 4.715 min Scan# 595

Delta R.T. -0.006 min


Lab File: VX026344.D

Acq: 06 Jan 2022 00:29

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion: 43 Resp: 38250

Ion Ratio Lower Upper

43 100

61 13.3 10.7 16.1

70 10.9 8.5 12.7



#38

Carbon Tetrachloride

Concen: 17.910 ug/l

RT: 5.684 min Scan# 754

Delta R.T. -0.000 min

Lab File: VX026344.D

Acq: 06 Jan 2022 00:29




Tgt Ion: 117 Resp: 36618


Ion Ratio Lower Upper

117 100

119 95.4 77.2 115.8


121 29.6 23.8 35.8








#39
Methylcyclohexane
Concen: 18.673 ug/l
RT: 7.385 min Scan# 1033
Delta R.T. -0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 83	Resp: 38715
Ion Ratio	Lower Upper
83 100	
55 79.0	57.8 86.8
98 50.0	40.5 60.7



#40
Benzene
Concen: 18.412 ug/l
RT: 6.044 min Scan# 813
Delta R.T. -0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29

Tgt Ion: 78	Resp: 96588
Ion Ratio	Lower Upper
78 100	
77 24.7	18.6 27.8









#41
Methacrylonitrile
Concen: 18.195 ug/l
RT: 4.928 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:	41	Resp:	21898
Ion	Ratio	Lower	Upper
41	100		
39	57.1	44.2	66.2
67	68.0	59.7	89.5
52	29.7	24.2	36.2



#42
1,2-Dichloroethane
Concen: 18.649 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29

Tgt Ion:	62	Resp:	39591
Ion	Ratio	Lower	Upper
62	100		
98	8.5	0.0	18.8





#43

Isopropyl Acetate

Concen: 18.035 ug/l

RT: 6.342 min Scan# 862

Delta R.T. -0.006 min

Lab File: VX026344.D

Acq: 06 Jan 2022 00:29

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 43 Resp: 62645

Ion Ratio Lower Upper

43 100

61 19.3 15.7 23.5

87 13.0 11.2 16.8



#44

Trichloroethene

Concen: 20.014 ug/l

RT: 7.129 min Scan# 991

Delta R.T. -0.000 min

Lab File: VX026344.D




Acq: 06 Jan 2022 00:29

Tgt Ion:130 Resp: 28251

Ion Ratio Lower Upper

130 100

95 96.1 0.0 186.4





#45

1,2-Dichloropropane

Concen: 18.681 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VX026344.D

Acq: 06 Jan 2022 00:29

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 63 Resp: 25066

Ion Ratio Lower Upper

63 100

65 30.7 25.8 38.6



#46

Dibromomethane

Concen: 17.700 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. -0.000 min

Lab File: VX026344.D

Acq: 06 Jan 2022 00:29




Tgt Ion: 93 Resp: 17616


Ion Ratio Lower Upper

93 100

95 81.9 66.6 100.0

174 95.5 77.3 115.9





#47

Bromodichloromethane

Concen: 17.219 ug/l

RT: 7.824 min Scan# 1105

Delta R.T. -0.000 min


Lab File: VX026344.D

Acq: 06 Jan 2022 00:29




Instrument :

MSVOA_X

ClientSampleId :



Tgt Ion: 83	Resp: 34867
Ion Ratio	Lower Upper
83 100	
85 66.8	53.5 80.3
127 8.7	6.6 10.0



#48

Methyl methacrylate

Concen: 18.869 ug/l




RT: 7.696 min Scan# 1084

Delta R.T. -0.000 min

Lab File: VX026344.D

Acq: 06 Jan 2022 00:29

Tgt Ion: 41	Resp: 31432
Ion Ratio	Lower Upper
41 100	
69 75.9	67.3 100.9
39 59.0	50.2 75.2





#49

1,4-Dioxane

Concen: 343.988 ug/l

RT: 7.684 min Scan# 1082

Delta R.T. -0.000 min


Lab File: VX026344.D

Acq: 06 Jan 2022 00:29

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion: 88 Resp: 10560

Ion Ratio Lower Upper

88 100

43 38.6 25.3 37.9#

58 72.3 51.8 77.6



#50

Toluene-d8

Concen: 49.876 ug/l

RT: 8.653 min Scan# 1241

Delta R.T. -0.000 min

Lab File: VX026344.D




Acq: 06 Jan 2022 00:29


Tgt Ion: 98 Resp: 211419

Ion Ratio Lower Upper

98 100

100 65.8 52.6 78.8









#51
4-Methyl-2-Pentanone
Concen: 91.205 ug/l
RT: 8.574 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 197143
Ion Ratio Lower Upper
43 100
58 38.1 30.8 46.2



#52
Toluene
Concen: 18.740 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29

Tgt Ion: 92 Resp: 60520
Ion Ratio Lower Upper
92 100
91 170.5 136.0 204.0





#53

t-1,3-Dichloropropene

Concen: 16.252 ug/l

RT: 8.982 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX026344.D

Acq: 06 Jan 2022 00:29

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 75 Resp: 32138

Ion Ratio Lower Upper

75 100

77 29.8 25.1 37.7



#54

cis-1,3-Dichloropropene

Concen: 17.151 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.006 min

Lab File: VX026344.D

Acq: 06 Jan 2022 00:29




Tgt Ion: 75 Resp: 36705


Ion Ratio Lower Upper

75 100

77 31.0 25.3 37.9

39 54.0 41.4 62.0









#55
1,1,2-Trichloroethane
Concen: 18.771 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 97	Resp: 24914
Ion Ratio	Lower Upper
97 100	
83 83.2	70.5 105.7
85 56.6	43.8 65.6
99 61.5	49.4 74.2



#56
Ethyl methacrylate
Concen: 19.624 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29

Tgt Ion: 69	Resp: 40171
Ion Ratio	Lower Upper
69 100	
41 73.1	52.8 79.2
39 40.9	32.5 48.7





#57

1,3-Dichloropropane

Concen: 19.004 ug/l

RT: 9.311 min Scan# 11349

Delta R.T. 0.000 min

Lab File: VX026344.D

Acq: 06 Jan 2022 00:29

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 76 Resp: 43207

Ion Ratio Lower Upper

76 100

78 31.7 26.0 39.0



#58

2-Chloroethyl Vinyl ether

Concen: 99.979 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. 0.000 min

Lab File: VX026344.D




Acq: 06 Jan 2022 00:29


Tgt Ion: 63 Resp: 102691

Ion Ratio Lower Upper

63 100

106 28.9 24.4 36.6





#59

2-Hexanone

Concen: 91.496 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.000 min

Lab File: VX026344.D

Acq: 06 Jan 2022 00:29

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 149810

Ion Ratio Lower Upper

43 100

58 51.8 26.9 80.6



#60

Dibromochloromethane

Concen: 16.563 ug/l

RT: 9.525 min Scan# 1384

Delta R.T. -0.000 min

Lab File: VX026344.D




Acq: 06 Jan 2022 00:29


Tgt Ion: 129 Resp: 24904

Ion Ratio Lower Upper

129 100

127 77.0 38.6 115.7









#61
1,2-Dibromoethane
Concen: 18.330 ug/l
RT: 9.610 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:107 Resp: 26031
Ion Ratio Lower Upper
107 100
109 94.8 75.1 112.7



#62
4-Bromofluorobenzene
Concen: 53.901 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.006 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29

Tgt Ion: 95 Resp: 81337
Ion Ratio Lower Upper
95 100
174 71.9 0.0 157.4
176 69.3 0.0 154.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:117 Resp: 153633
Ion Ratio Lower Upper
117 100
82 58.6 43.8 65.6
119 33.2 25.2 37.8



#64
Tetrachloroethene
Concen: 22.694 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29

Tgt Ion:164 Resp: 29423
Ion Ratio Lower Upper
164 100
166 129.3 99.4 149.0
129 102.2 78.6 118.0
131 97.6 75.4 113.0









#65
Chlorobenzene
Concen: 18.745 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:112 Resp: 63262
Ion Ratio Lower Upper
112 100
114 31.4 26.6 39.8



#66
1,1,1,2-Tetrachloroethane
Concen: 18.566 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29

Tgt Ion:131 Resp: 23876
Ion Ratio Lower Upper
131 100
133 94.9 47.1 141.3
119 62.0 31.9 95.9





#67

Ethyl Benzene

Concen: 19.633 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. 0.000 min

Lab File: VX026344.D

Acq: 06 Jan 2022 00:29

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 91 Resp: 118479

Ion Ratio Lower Upper

91 100

106 29.6 25.7 38.5



#68

m/p-Xylenes

Concen: 39.293 ug/l

RT: 10.305 min Scan# 1512

Delta R.T. -0.000 min

Lab File: VX026344.D




Acq: 06 Jan 2022 00:29


Tgt Ion:106 Resp: 88535

Ion Ratio Lower Upper

106 100

91 209.0 158.6 238.0









#69
o-Xylene
Concen: 19.620 ug/l
RT: 10.640 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:106 Resp: 43255
Ion Ratio Lower Upper
106 100
91 225.0 105.5 316.4



#70
Styrene
Concen: 19.754 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29

Tgt Ion:104 Resp: 71394
Ion Ratio Lower Upper
104 100
78 53.8 39.8 59.6
103 55.1 43.0 64.6





#71

Bromoform

Concen: 15.005 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min

Lab File: VX026344.D





Acq: 06 Jan 2022 00:29

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion:173	Resp:	15654
Ion Ratio	Lower	Upper
173	100	
175	49.4	24.4 73.4
254	0.0	0.1 0.1#



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 12.024 min Scan# 1794


Delta R.T. 0.000 min

Lab File: VX026344.D

Acq: 06 Jan 2022 00:29

Tgt Ion:152	Resp:	72315
Ion Ratio	Lower	Upper
152	100	
115	72.7	41.3 123.9
150	164.6	0.0 351.8





#73

Isopropylbenzene

Concen: 20.726 ug/l

RT: 10.963 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX026344.D

Acq: 06 Jan 2022 00:29

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:105 Resp: 116504

Ion Ratio Lower Upper

105 100

120 25.6 13.6 40.7



#74

N-ethyl acetate

Concen: 19.442 ug/l

RT: 10.841 min Scan# 1600

Delta R.T. -0.007 min

Lab File: VX026344.D

Acq: 06 Jan 2022 00:29

Tgt Ion: 43 Resp: 50182




Ion Ratio Lower Upper


43 100

70 40.4 34.1 51.1

55 22.7 19.2 28.8

61 23.1 19.0 28.4









#75
 1,1,2,2-Tetrachloroethane
 Concen: 17.085 ug/l
 RT: 11.213 min Scan# 1661
 Delta R.T. 0.000 min
 Lab File: VX026344.D
 Acq: 06 Jan 2022 00:29




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 83 Resp: 32661
 Ion Ratio Lower Upper
 83 100
 131 10.9 5.1 15.3
 85 63.9 31.6 95.0



#76
 1,2,3-Trichloropropane
 Concen: 22.073 ug/l
 RT: 11.238 min Scan# 1665
 Delta R.T. -0.006 min
 Lab File: VX026344.D
 Acq: 06 Jan 2022 00:29

Tgt Ion: 75 Resp: 38528
 Ion Ratio Lower Upper
 75 100
 77 29.9 19.0 57.0









#77
Bromobenzene
Concen: 20.237 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:156 Resp: 26977
Ion Ratio Lower Upper
156 100
77 171.9 78.2 234.6
158 94.3 48.6 145.9



#78
n-propylbenzene
Concen: 20.246 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29

Tgt Ion: 91 Resp: 130850
Ion Ratio Lower Upper
91 100
120 22.6 12.0 36.1









#79
2-Chlorotoluene
Concen: 20.298 ug/l
RT: 11.366 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 80753
Ion Ratio Lower Upper
91 100
126 32.0 17.8 53.3



#80
1,3,5-Trimethylbenzene
Concen: 20.597 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29

Tgt Ion: 105 Resp: 96079
Ion Ratio Lower Upper
105 100
120 47.8 24.9 74.7









#81
trans-1,4-Dichloro-2-butene
Concen: 11.550 ug/l
RT: 11.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 6969
Ion Ratio Lower Upper
75 100
53 141.1 79.5 119.3#
89 45.5 38.3 57.5



#82
4-Chlorotoluene
Concen: 20.809 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29

Tgt Ion: 91 Resp: 93492
Ion Ratio Lower Upper
91 100
126 28.5 15.6 46.7





#83

tert-Butylbenzene

Concen: 19.563 ug/l

RT: 11.719 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX026344.D

Acq: 06 Jan 2022 00:29

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:119 Resp: 88575

Ion Ratio Lower Upper

119 100

91 65.5 28.7 86.1

134 23.8 11.9 35.9



#84

1,2,4-Trimethylbenzene

Concen: 20.729 ug/l

RT: 11.756 min Scan# 1750

Delta R.T. -0.000 min

Lab File: VX026344.D




Acq: 06 Jan 2022 00:29


Tgt Ion:105 Resp: 96617

Ion Ratio Lower Upper

105 100

120 43.7 22.9 68.7









#85
 sec-Butylbenzene
 Concen: 20.137 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX026344.D
 Acq: 06 Jan 2022 00:29




Instrument :
 MSVOA_X
 ClientSampleId :

Tgt Ion:105 Resp: 113777
 Ion Ratio Lower Upper
 105 100
 134 20.3 10.4 31.1



#86
 p-Isopropyltoluene
 Concen: 20.387 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX026344.D
 Acq: 06 Jan 2022 00:29

Tgt Ion:119 Resp: 96030
 Ion Ratio Lower Upper
 119 100
 134 25.9 13.2 39.6
 91 25.4 11.9 35.9





#87

1,3-Dichlorobenzene

Concen: 19.937 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX026344.D

Acq: 06 Jan 2022 00:29

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:146 Resp: 49890

Ion Ratio Lower Upper

146 100

111 41.5 19.9 59.6

148 63.9 31.9 95.5



#88

1,4-Dichlorobenzene

Concen: 18.749 ug/l

RT: 12.042 min Scan# 1797

Delta R.T. -0.001 min

Lab File: VX026344.D

Acq: 06 Jan 2022 00:29




Tgt Ion:146 Resp: 50302

Ion Ratio Lower Upper

146 100

111 42.4 19.9 59.7

148 63.0 31.9 95.7





#89

n-Butylbenzene

Concen: 19.380 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX026344.D

Acq: 06 Jan 2022 00:29

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 91 Resp: 81126

Ion Ratio Lower Upper

91 100

92 54.1 27.3 81.8

134 26.1 13.5 40.4



#90

Hexachloroethane

Concen: 13.877 ug/l

RT: 12.542 min Scan# 1879

Delta R.T. 0.000 min

Lab File: VX026344.D




Acq: 06 Jan 2022 00:29

Tgt Ion: 117 Resp: 13055

Ion Ratio Lower Upper

117 100


201 70.7 35.4 106.3








#91
 1,2-Dichlorobenzene
 Concen: 18.718 ug/l
 RT: 12.335 min Scan# 1845
 Delta R.T. 0.000 min
 Lab File: VX026344.D
 Acq: 06 Jan 2022 00:29


Instrument :
 MSVOA_X
 ClientSampleId :





Tgt Ion:146 Resp: 48589
 Ion Ratio Lower Upper
 146 100
 111 44.2 20.6 61.9
 148 62.6 31.9 95.7




#92
 1,2-Dibromo-3-Chloropropane
 Concen: 16.133 ug/l
 RT: 12.945 min Scan# 1945
 Delta R.T. -0.000 min
 Lab File: VX026344.D
 Acq: 06 Jan 2022 00:29




Tgt Ion: 75 Resp: 7895
 Ion Ratio Lower Upper
 75 100
 155 77.5 44.0 131.8
 157 100.6 55.5 166.5








#93
1,2,4-Trichlorobenzene
Concen: 19.217 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. -0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:180 Resp: 29478
Ion Ratio Lower Upper
180 100
182 95.2 47.2 141.6
145 33.1 16.3 48.9




#94
Hexachlorobutadiene
Concen: 19.194 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX026344.D
Acq: 06 Jan 2022 00:29



Tgt Ion:225 Resp: 12225
Ion Ratio Lower Upper
225 100
223 63.2 31.6 95.0
227 63.4 32.5 97.5





#95

Naphthalene

Concen: 19.326 ug/l

RT: 13.780 min Scan# 2082

Delta R.T. -0.000 min


Lab File: VX026344.D

Acq: 06 Jan 2022 00:29

Instrument :




MSVOA_X

ClientSampleId :



Tgt Ion:128 Resp: 108015

Ion	Ratio	Lower	Upper
128	100		
127	12.9	10.3	15.5
129	10.7	8.8	13.2



#96

1,2,3-Trichlorobenzene

Concen: 19.067 ug/l

RT: 13.963 min Scan# 2112

Delta R.T. -0.000 min

Lab File: VX026344.D

Acq: 06 Jan 2022 00:29

Tgt Ion:180 Resp: 29758

Ion	Ratio	Lower	Upper
180	100		
182	93.9	47.5	142.6
145	35.1	16.7	50.1

