

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010522\
 Data File : VX026337.D
 Acq On : 05 Jan 2022 21:01
 Operator : JC/MD
 Sample : PB141832TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141832TB

Quant Time: Jan 06 03:12:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	154515	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	257893	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	226311	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	103911	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	104641	46.811	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.620%
35) Dibromofluoromethane	5.391	113	81251	48.460	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.920%
50) Toluene-d8	8.653	98	297562	47.713	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.420%
62) 4-Bromofluorobenzene	11.079	95	115035	51.815	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.620%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	13752	11.354	ug/l	97
20) Methylene Chloride	2.794	84	5834	2.107	ug/l	99
25) 2-Butanone	4.568	43	2675	1.586	ug/l	96
43) Isopropyl Acetate	6.342	43	16059	3.142	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010522\
Data File : VX026337.D
Acq On : 05 Jan 2022 21:01
Operator : JC/MD
Sample : PB141832TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141832TB

Quant Time: Jan 06 03:12:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

