

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010522\
 Data File : VX026338.D
 Acq On : 05 Jan 2022 21:24
 Operator : JC/MD
 Sample : N1014-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 72-11944

Quant Time: Jan 06 03:13:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	157286	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	264026	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	224048	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98819	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	109878	48.288	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.580%
35) Dibromofluoromethane	5.391	113	85692	49.922	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.840%
50) Toluene-d8	8.653	98	301275	47.186	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.380%
62) 4-Bromofluorobenzene	11.079	95	111432	49.026	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.060%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	20322	16.483	ug/l	100
20) Methylene Chloride	2.794	84	8867	3.621	ug/l	97
43) Isopropyl Acetate	6.342	43	14778	2.825	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010522\
Data File : VX026338.D
Acq On : 05 Jan 2022 21:24
Operator : JC/MD
Sample : N1014-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
72-11944

Quant Time: Jan 06 03:13:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

