

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010522\
 Data File : VX026342.D
 Acq On : 05 Jan 2022 23:43
 Operator : JC/MD
 Sample : VX0105WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0105WBL02

Quant Time: Jan 06 03:13:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	162488	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	275131	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	236223	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	102301	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	117999	50.197	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.400%
35) Dibromofluoromethane	5.391	113	88918	49.711	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.420%
50) Toluene-d8	8.653	98	317144	47.667	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.340%
62) 4-Bromofluorobenzene	11.079	95	118396	49.988	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010522\
Data File : VX026342.D
Acq On : 05 Jan 2022 23:43
Operator : JC/MD
Sample : VX0105WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0105WBL02

Quant Time: Jan 06 03:13:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

