

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010522\
 Data File : VX026346.D
 Acq On : 06 Jan 2022 01:15
 Operator : JC/MD
 Sample : N1019-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PZ-7

Quant Time: Jan 06 05:04:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	152463	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	255222	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	225287	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	103355	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	107713	48.834	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.660%
35) Dibromofluoromethane	5.391	113	81613	49.186	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.380%
50) Toluene-d8	8.653	98	299007	48.447	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.900%
62) 4-Bromofluorobenzene	11.079	95	115987	52.791	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.580%
Target Compounds						
					Qvalue	
16) Acetone	2.349	43	62864	52.603	ug/l #	52
31) Cyclohexane	5.483	56	292391	93.085	ug/l	97
39) Methylcyclohexane	7.385	83	74640	24.725	ug/l	94
40) Benzene	6.044	78	517194	67.713	ug/l	99
52) Toluene	8.720	92	398133	84.670	ug/l	99
67) Ethyl Benzene	10.195	91	436658	49.344	ug/l	97
68) m/p-Xylenes	10.305	106	2609842	789.880	ug/l	97
69) o-Xylene	10.646	106	31546	9.758	ug/l	92
73) Isopropylbenzene	10.963	105	587195	73.091	ug/l	98
78) n-propylbenzene	11.305	91	1171724	126.852	ug/l	97
80) 1,3,5-Trimethylbenzene	11.451	105	337830	50.672	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	18898	2.837	ug/l	99
85) sec-Butylbenzene	11.890	105	17797	2.204	ug/l	94
89) n-Butylbenzene	12.329	91	27660	4.623	ug/l #	45
95) Naphthalene	13.780	128	874256	109.446	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010522\
Data File : VX026346.D
Acq On : 06 Jan 2022 01:15
Operator : JC/MD
Sample : N1019-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PZ-7

Quant Time: Jan 06 05:04:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

