

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010522\
 Data File : VX026348.D
 Acq On : 06 Jan 2022 02:02
 Operator : JC/MD
 Sample : N1019-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-26I

Quant Time: Jan 06 05:04:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	153380	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	255986	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	221122	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96317	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	110936	49.995	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.980%
35) Dibromofluoromethane	5.391	113	83056	49.906	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.820%
50) Toluene-d8	8.653	98	295794	47.783	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.560%
62) 4-Bromofluorobenzene	11.079	95	109910	49.875	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.760%
Target Compounds						
16) Acetone	2.386	43	8375	6.966	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010522\
Data File : VX026348.D
Acq On : 06 Jan 2022 02:02
Operator : JC/MD
Sample : N1019-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-26I

Quant Time: Jan 06 05:04:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

