

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010522\
 Data File : VX026359.D
 Acq On : 06 Jan 2022 06:16
 Operator : JC/MD
 Sample : N1020-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2

Quant Time: Jan 06 12:23:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	155031	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	258711	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	222570	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96866	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	111365	49.654	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.300%
35) Dibromofluoromethane	5.391	113	82461	49.027	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.060%
50) Toluene-d8	8.653	98	296911	47.458	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.920%
62) 4-Bromofluorobenzene	11.079	95	110301	49.526	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.060%
Target Compounds						
16) Acetone	2.386	43	13759	11.322	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010522\
Data File : VX026359.D
Acq On : 06 Jan 2022 06:16
Operator : JC/MD
Sample : N1020-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2

Quant Time: Jan 06 12:23:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

