

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010522\
 Data File : VX026360.D
 Acq On : 06 Jan 2022 06:39
 Operator : JC/MD
 Sample : N1020-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-27

Quant Time: Jan 06 12:23:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	157404	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	261771	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	223929	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94734	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	113756	49.955	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.900%
35) Dibromofluoromethane	5.391	113	84916	49.896	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.800%
50) Toluene-d8	8.653	98	302881	47.846	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.700%
62) 4-Bromofluorobenzene	11.079	95	111247	49.366	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.740%
Target Compounds						
16) Acetone	2.386	43	9530	7.724	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010522\
Data File : VX026360.D
Acq On : 06 Jan 2022 06:39
Operator : JC/MD
Sample : N1020-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-27

Quant Time: Jan 06 12:23:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

