

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX010522\
 Data File : VX026314.D
 Acq On : 05 Jan 2022 11:44
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 06 03:07:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	50.000	54.871	-9.7	103	0.00
3 P	Chloromethane	50.000	43.760	12.5	85	0.00
4 C	Vinyl Chloride	50.000	45.522	9.0#	89	0.00
5 T	Bromomethane	50.000	57.521	-15.0	109	0.00
6 T	Chloroethane	50.000	40.445	19.1	82	0.00
7 T	Trichlorofluoromethane	50.000	43.956	12.1	84	0.00
8 T	Diethyl Ether	50.000	45.718	8.6	87	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	53.504	-7.0	104	0.00
10 T	Methyl Iodide	50.000	43.939	12.1	82	0.00
11 T	Tert butyl alcohol	250.000	242.983	2.8	102	-0.02
12 CM	1,1-Dichloroethene	50.000	52.156	-4.3#	102	0.00
13 T	Acrolein	250.000	305.897	-22.4	132	0.00
14 T	Allyl chloride	50.000	51.426	-2.9	100	0.00
15 T	Acrylonitrile	250.000	250.235	-0.1	97	0.00
16 T	Acetone	250.000	293.695	-17.5	119	0.00
17 T	Carbon Disulfide	50.000	50.391	-0.8	99	0.00
18 T	Methyl Acetate	50.000	50.484	-1.0	98	0.00
19 T	Methyl tert-butyl Ether	50.000	57.421	-14.8	110	0.00
20 T	Methylene Chloride	50.000	49.948	0.1	96	0.00
21 T	trans-1,2-Dichloroethene	50.000	51.904	-3.8	103	0.00
22 T	Diisopropyl ether	50.000	52.837	-5.7	100	0.00
23 T	Vinyl Acetate	250.000	269.631	-7.9	100	0.00
24 P	1,1-Dichloroethane	50.000	53.295	-6.6	102	0.00
25 T	2-Butanone	250.000	270.514	-8.2	105	-0.01
26 T	2,2-Dichloropropane	50.000	59.378	-18.8	114	0.00
27 T	cis-1,2-Dichloroethene	50.000	53.482	-7.0	103	0.00
28 T	Bromochloromethane	50.000	47.735	4.5	94	0.00
29 T	Tetrahydrofuran	250.000	251.416	-0.6	96	-0.01
30 C	Chloroform	50.000	52.183	-4.4#	101	-0.01
31 T	Cyclohexane	50.000	53.310	-6.6	101	0.00
32 T	1,1,1-Trichloroethane	50.000	55.309	-10.6	106	-0.01
33 S	1,2-Dichloroethane-d4	50.000	55.058	-10.1	117	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	97	0.00
35 S	Dibromofluoromethane	50.000	55.786	-11.6	114	0.00
36 T	1,1-Dichloropropene	50.000	55.128	-10.3	103	0.00
37 T	Ethyl Acetate	50.000	53.574	-7.1	98	-0.01
38 T	Carbon Tetrachloride	50.000	53.127	-6.3	100	0.00
39 T	Methylcyclohexane	50.000	57.278	-14.6	106	0.00
40 TM	Benzene	50.000	53.142	-6.3	100	0.00
41 T	Methacrylonitrile	50.000	53.513	-7.0	98	-0.01
42 TM	1,2-Dichloroethane	50.000	53.947	-7.9	103	0.00
43 T	Isopropyl Acetate	50.000	54.540	-9.1	102	-0.01
44 TM	Trichloroethene	50.000	56.586	-13.2	107	0.00
45 C	1,2-Dichloropropane	50.000	53.908	-7.8#	99	0.00
46 T	Dibromomethane	50.000	51.234	-2.5	96	0.00
47 T	Bromodichloromethane	50.000	51.194	-2.4	96	0.00
48 T	Methyl methacrylate	50.000	55.319	-10.6	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX010522\
 Data File : VX026314.D
 Acq On : 05 Jan 2022 11:44
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 06 03:07:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	895.754	10.4	82	-0.02
50 S	Toluene-d8	50.000	53.344	-6.7	108	0.00
51 T	4-Methyl-2-Pentanone	250.000	258.169	-3.3	94	0.00
52 CM	Toluene	50.000	54.049	-8.1#	100	0.00
53 T	t-1,3-Dichloropropene	50.000	49.757	0.5	99	0.00
54 T	cis-1,3-Dichloropropene	50.000	54.496	-9.0	99	0.00
55 T	1,1,2-Trichloroethane	50.000	52.371	-4.7	98	0.00
56 T	Ethyl methacrylate	50.000	57.616	-15.2	103	0.00
57 T	1,3-Dichloropropane	50.000	53.048	-6.1	100	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	257.642	-3.1	101	0.00
59 T	2-Hexanone	250.000	270.299	-8.1	98	0.00
60 T	Dibromochloromethane	50.000	49.584	0.8	92	0.00
61 T	1,2-Dibromoethane	50.000	52.163	-4.3	97	0.00
62 S	4-Bromofluorobenzene	50.000	58.489	-17.0	119	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	94	0.00
64 T	Tetrachloroethene	50.000	62.259	-24.5	118	0.00
65 PM	Chlorobenzene	50.000	53.953	-7.9	99	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	54.937	-9.9	98	0.00
67 C	Ethyl Benzene	50.000	56.235	-12.5#	100	0.00
68 T	m/p-Xylenes	100.000	112.282	-12.3	99	0.00
69 T	o-Xylene	50.000	56.168	-12.3	99	0.00
70 T	Styrene	50.000	57.945	-15.9	99	0.00
71 P	Bromoform	50.000	48.913	2.2	86	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	92	0.00
73 T	Isopropylbenzene	50.000	57.724	-15.4	101	0.00
74 T	N-amyl acetate	50.000	58.015	-16.0	97	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	48.442	3.1	87	0.00
76 T	1,2,3-Trichloropropane	50.000	54.479	-9.0	98	0.00
77 T	Bromobenzene	50.000	55.987	-12.0	98	0.00
78 T	n-propylbenzene	50.000	58.351	-16.7	98	0.00
79 T	2-Chlorotoluene	50.000	57.403	-14.8	101	0.00
80 T	1,3,5-Trimethylbenzene	50.000	58.372	-16.7	100	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	45.187	9.6	80	0.00
82 T	4-Chlorotoluene	50.000	58.715	-17.4	101	0.00
83 T	tert-Butylbenzene	50.000	56.500	-13.0	97	0.00
84 T	1,2,4-Trimethylbenzene	50.000	59.015	-18.0	102	0.00
85 T	sec-Butylbenzene	50.000	58.749	-17.5	101	0.00
86 T	p-Isopropyltoluene	50.000	59.619	-19.2	102	0.00
87 T	1,3-Dichlorobenzene	50.000	57.246	-14.5	99	0.00
88 T	1,4-Dichlorobenzene	50.000	53.142	-6.3	95	0.00
89 T	n-Butylbenzene	50.000	59.375	-18.8	102	0.00
90 T	Hexachloroethane	50.000	45.322	9.4	79	0.00
91 T	1,2-Dichlorobenzene	50.000	52.871	-5.7	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	50.669	-1.3	92	0.00
93 T	1,2,4-Trichlorobenzene	50.000	57.945	-15.9	101	0.00
94 T	Hexachlorobutadiene	50.000	59.592	-19.2	106	0.00
95 T	Naphthalene	50.000	55.961	-11.9	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010522\
Data File : VX026314.D
Acq On : 05 Jan 2022 11:44
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 06 03:07:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	56.644	-13.3	98	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6