

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX010522\
 Data File : VX026341.D
 Acq On : 05 Jan 2022 22:57
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 06 03:13:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	50.000	49.035	1.9	91	0.00
3 P	Chloromethane	50.000	39.050	21.9	75	0.00
4 C	Vinyl Chloride	50.000	41.669	16.7#	80	0.00
5 T	Bromomethane	50.000	29.038	41.9#	54	0.00
6 T	Chloroethane	50.000	33.249	33.5#	66	0.00
7 T	Trichlorofluoromethane	50.000	38.019	24.0	71	0.00
8 T	Diethyl Ether	50.000	43.199	13.6	81	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	47.157	5.7	90	0.00
10 T	Methyl Iodide	50.000	37.844	24.3	70	0.00
11 T	Tert butyl alcohol	250.000	237.926	4.8	98	0.02
12 CM	1,1-Dichloroethene	50.000	47.304	5.4#	91	0.00
13 T	Acrolein	250.000	224.541	10.2	95	0.00
14 T	Allyl chloride	50.000	45.888	8.2	88	0.00
15 T	Acrylonitrile	250.000	230.892	7.6	88	0.00
16 T	Acetone	250.000	202.840	18.9	81	0.00
17 T	Carbon Disulfide	50.000	44.903	10.2	87	0.00
18 T	Methyl Acetate	50.000	46.810	6.4	90	0.00
19 T	Methyl tert-butyl Ether	50.000	51.789	-3.6	97	0.00
20 T	Methylene Chloride	50.000	46.767	6.5	88	0.00
21 T	trans-1,2-Dichloroethene	50.000	46.984	6.0	92	0.00
22 T	Diisopropyl ether	50.000	48.060	3.9	90	0.00
23 T	Vinyl Acetate	250.000	241.770	3.3	88	0.00
24 P	1,1-Dichloroethane	50.000	48.459	3.1	92	0.00
25 T	2-Butanone	250.000	225.354	9.9	86	0.00
26 T	2,2-Dichloropropane	50.000	48.818	2.4	92	0.00
27 T	cis-1,2-Dichloroethene	50.000	48.493	3.0	92	0.00
28 T	Bromochloromethane	50.000	45.571	8.9	89	0.00
29 T	Tetrahydrofuran	250.000	227.784	8.9	86	0.00
30 C	Chloroform	50.000	47.652	4.7#	91	0.00
31 T	Cyclohexane	50.000	47.991	4.0	90	0.00
32 T	1,1,1-Trichloroethane	50.000	49.734	0.5	94	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.816	2.4	102	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	100	0.00
35 S	Dibromofluoromethane	50.000	47.472	5.1	100	0.00
36 T	1,1-Dichloropropene	50.000	46.569	6.9	90	0.00
37 T	Ethyl Acetate	50.000	46.378	7.2	88	0.00
38 T	Carbon Tetrachloride	50.000	44.905	10.2	87	0.00
39 T	Methylcyclohexane	50.000	47.512	5.0	91	0.00
40 TM	Benzene	50.000	45.816	8.4	89	0.00
41 T	Methacrylonitrile	50.000	48.083	3.8	91	0.00
42 TM	1,2-Dichloroethane	50.000	46.350	7.3	91	0.00
43 T	Isopropyl Acetate	50.000	47.488	5.0	92	0.00
44 TM	Trichloroethene	50.000	50.341	-0.7	99	0.00
45 C	1,2-Dichloropropane	50.000	46.415	7.2#	88	0.00
46 T	Dibromomethane	50.000	44.623	10.8	87	0.00
47 T	Bromodichloromethane	50.000	44.207	11.6	86	0.00
48 T	Methyl methacrylate	50.000	48.429	3.1	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010522\
 Data File : VX026341.D
 Acq On : 05 Jan 2022 22:57
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 06 03:13:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	841.877	15.8	80	0.00
50 S	Toluene-d8	50.000	45.895	8.2	97	0.00
51 T	4-Methyl-2-Pentanone	250.000	227.883	8.8	86	0.00
52 CM	Toluene	50.000	47.340	5.3#	90	0.00
53 T	t-1,3-Dichloropropene	50.000	42.424	15.2	87	0.00
54 T	cis-1,3-Dichloropropene	50.000	45.957	8.1	87	0.00
55 T	1,1,2-Trichloroethane	50.000	46.304	7.4	89	0.00
56 T	Ethyl methacrylate	50.000	50.627	-1.3	94	0.00
57 T	1,3-Dichloropropane	50.000	47.087	5.8	92	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	234.139	6.3	95	0.00
59 T	2-Hexanone	250.000	226.691	9.3	85	0.00
60 T	Dibromochloromethane	50.000	44.262	11.5	85	0.00
61 T	1,2-Dibromoethane	50.000	46.348	7.3	89	0.00
62 S	4-Bromofluorobenzene	50.000	51.134	-2.3	108	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	99	0.00
64 T	Tetrachloroethene	50.000	56.603	-13.2	113	0.00
65 PM	Chlorobenzene	50.000	46.778	6.4	90	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	47.530	4.9	89	0.00
67 C	Ethyl Benzene	50.000	48.683	2.6#	91	0.00
68 T	m/p-Xylenes	100.000	99.229	0.8	92	0.00
69 T	o-Xylene	50.000	49.407	1.2	91	0.00
70 T	Styrene	50.000	50.968	-1.9	91	0.00
71 P	Bromoform	50.000	41.947	16.1	78	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	94	0.00
73 T	Isopropylbenzene	50.000	51.464	-2.9	92	0.00
74 T	N-ethyl acetate	50.000	51.115	-2.2	87	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	41.681	16.6	77	0.00
76 T	1,2,3-Trichloropropane	50.000	47.598	4.8	88	0.00
77 T	Bromobenzene	50.000	50.838	-1.7	91	0.00
78 T	n-propylbenzene	50.000	51.634	-3.3	89	0.00
79 T	2-Chlorotoluene	50.000	50.818	-1.6	92	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.245	-4.5	92	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	35.909	28.2#	65	0.00
82 T	4-Chlorotoluene	50.000	52.001	-4.0	91	0.00
83 T	tert-Butylbenzene	50.000	49.775	0.5	87	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.940	-3.9	91	0.00
85 T	sec-Butylbenzene	50.000	51.322	-2.6	90	0.00
86 T	p-Isopropyltoluene	50.000	52.354	-4.7	92	0.00
87 T	1,3-Dichlorobenzene	50.000	49.901	0.2	89	0.00
88 T	1,4-Dichlorobenzene	50.000	46.897	6.2	86	0.00
89 T	n-Butylbenzene	50.000	50.157	-0.3	88	0.00
90 T	Hexachloroethane	50.000	38.641	22.7	69	0.00
91 T	1,2-Dichlorobenzene	50.000	46.823	6.4	89	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	44.213	11.6	82	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.160	1.7	88	0.00
94 T	Hexachlorobutadiene	50.000	47.944	4.1	87	0.00
95 T	Naphthalene	50.000	50.315	-0.6	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010522\
Data File : VX026341.D
Acq On : 05 Jan 2022 22:57
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 06 03:13:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	48.619	2.8	86	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6