

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010522\
 Data File : VX026364.D
 Acq On : 06 Jan 2022 08:11
 Operator : JC/MD
 Sample : N1020-09MSD
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7AMSD

Quant Time: Jan 06 12:26:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 01/07/2022
 Supervised By :Semsettin Yesilyurt 01/07/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	106058	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	177525	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	153769	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	71852	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	76116	49.608	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 99.220%			
35) Dibromofluoromethane	5.391	113	57952	50.212	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.420%			
50) Toluene-d8	8.653	98	208432	48.552	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 97.100%			
62) 4-Bromofluorobenzene	11.079	95	79832	52.237	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 104.480%			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.172	85	53672	48.185	ug/l	98
3) Chloromethane	1.294	50	49762	38.379	ug/l	96
4) Vinyl Chloride	1.380	62	57042	41.802	ug/l	99
5) Bromomethane	1.617	94	20678	24.554	ug/l	98
6) Chloroethane	1.691	64	31958	33.306	ug/l	98
7) Trichlorofluoromethane	1.892	101	93350	39.713	ug/l	96
8) Diethyl Ether	2.136	74	34473	42.691	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.337	101	57887	46.714	ug/l	99
10) Methyl Iodide	2.459	142	53295	33.257	ug/l	98
11) Tert butyl alcohol	2.983	59	66664m	217.530	ug/l	
12) 1,1-Dichloroethene	2.325	96	55969	47.657	ug/l	94
13) Acrolein	2.239	56	40982	373.977	ug/l #	1
14) Allyl chloride	2.672	41	106263	45.785	ug/l	96
15) Acrylonitrile	3.068	53	171700	226.292	ug/l	100
16) Acetone	2.386	43	160712	193.319	ug/l	96
17) Carbon Disulfide	2.520	76	154253	45.765	ug/l	99
18) Methyl Acetate	2.709	43	111344	41.291	ug/l	97
19) Methyl tert-butyl Ether	3.117	73	214039	52.178	ug/l #	81
20) Methylene Chloride	2.794	84	59747	44.850	ug/l	96
21) trans-1,2-Dichloroethene	3.099	96	60619	47.315	ug/l	91
22) Diisopropyl ether	3.763	45	216767	48.934	ug/l	93
23) Vinyl Acetate	3.727	43	755086	214.667	ug/l	98
24) 1,1-Dichloroethane	3.617	63	118047	49.173	ug/l	99
25) 2-Butanone	4.562	43	244654	211.265	ug/l	96
26) 2,2-Dichloropropane	4.483	77	75607	41.801	ug/l	99
27) cis-1,2-Dichloroethene	4.495	96	71047	48.743	ug/l	94
28) Bromochloromethane	4.903	49	50227	47.205	ug/l	94
29) Tetrahydrofuran	5.013	42	161541	221.316	ug/l	95
30) Chloroform	5.099	83	121863	47.994	ug/l	94
31) Cyclohexane	5.477	56	154591	70.749	ug/l	97
32) 1,1,1-Trichloroethane	5.391	97	109066	49.410	ug/l	99
36) 1,1-Dichloropropene	5.702	75	90319	48.519	ug/l	98
37) Ethyl Acetate	4.721	43	94479	43.772	ug/l	100
38) Carbon Tetrachloride	5.684	117	94381	45.580	ug/l	99
39) Methylcyclohexane	7.385	83	108377	51.613	ug/l	94
40) Benzene	6.044	78	2054221	386.655	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010522\
 Data File : VX026364.D
 Acq On : 06 Jan 2022 08:11
 Operator : JC/MD
 Sample : N1020-09MSD
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7AMSD

Quant Time: Jan 06 12:26:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 01/07/2022
 Supervised By :Semsettin Yesilyurt 01/07/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	57012	46.775	ug/l	94
42) 1,2-Dichloroethane	6.092	62	117375	54.590	ug/l	95
43) Isopropyl Acetate	6.342	43	161261	45.842	ug/l	97
44) Trichloroethene	7.129	130	72008	50.370	ug/l	92
45) 1,2-Dichloropropane	7.434	63	65232	48.003	ug/l	96
46) Dibromomethane	7.586	93	45563	45.203	ug/l	99
47) Bromodichloromethane	7.824	83	91526	44.632	ug/l	98
48) Methyl methacrylate	7.696	41	80926	47.968	ug/l	93
49) 1,4-Dioxane	7.690	88	24944	802.303	ug/l #	90
51) 4-Methyl-2-Pentanone	8.574	43	491113	224.342	ug/l	99
52) Toluene	8.720	92	358730	109.680	ug/l	100
53) t-1,3-Dichloropropene	8.982	75	89377	41.321	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	98362	45.381	ug/l	96
55) 1,1,2-Trichloroethane	9.153	97	62804	46.721	ug/l	98
56) Ethyl methacrylate	9.116	69	104386	50.350	ug/l	95
57) 1,3-Dichloropropane	9.311	76	108679	47.199	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.141	63	691	10.428	ug/l #	44
59) 2-Hexanone	9.433	43	364963	220.092	ug/l	99
60) Dibromochloromethane	9.525	129	66329	43.558	ug/l	100
61) 1,2-Dibromoethane	9.610	107	66495	46.233	ug/l	99
64) Tetrachloroethene	9.275	164	72392	55.788	ug/l	98
65) Chlorobenzene	10.079	112	162597	48.135	ug/l	97
66) 1,1,1,2-Tetrachloroethane	10.165	131	62001	48.168	ug/l	98
67) Ethyl Benzene	10.195	91	505268	83.654	ug/l	97
68) m/p-Xylenes	10.305	106	562087	249.240	ug/l	95
69) o-Xylene	10.646	106	133383	60.448	ug/l	94
70) Styrene	10.659	104	186810	51.643	ug/l	95
71) Bromoform	10.799	173	43452	41.612	ug/l #	100
73) Isopropylbenzene	10.963	105	366623	65.644	ug/l	98
74) N-amyl acetate	10.848	43	117367	45.765	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.213	83	82921	43.655	ug/l	98
76) 1,2,3-Trichloropropane	11.238	75	82160m	47.373	ug/l	
77) Bromobenzene	11.201	156	68864	51.993	ug/l	93
78) n-propylbenzene	11.305	91	463333	72.153	ug/l	98
79) 2-Chlorotoluene	11.366	91	210327	53.209	ug/l	94
80) 1,3,5-Trimethylbenzene	11.451	105	273303	58.967	ug/l	97
81) trans-1,4-Dichloro-2-b...	11.024	75	19732	32.915	ug/l #	83
82) 4-Chlorotoluene	11.457	91	237105	53.113	ug/l	95
83) tert-Butylbenzene	11.713	119	225870	50.208	ug/l	93
84) 1,2,4-Trimethylbenzene	11.756	105	340077	73.432	ug/l	97
85) sec-Butylbenzene	11.890	105	292325	52.070	ug/l	99
86) p-Isopropyltoluene	12.012	119	248094	53.010	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	127089	51.114	ug/l	98
88) 1,4-Dichlorobenzene	12.042	146	128314	48.133	ug/l	99
89) n-Butylbenzene	12.335	91	212112	50.998	ug/l	100
90) Hexachloroethane	12.542	117	35656	38.144	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	121897	47.261	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.945	75	21187	43.572	ug/l	92
93) 1,2,4-Trichlorobenzene	13.591	180	76948	50.485	ug/l	99
94) Hexachlorobutadiene	13.725	225	30913	48.848	ug/l	98
95) Naphthalene	13.780	128	539964	97.235	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	77484	49.967	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010522\
Data File : VX026364.D
Acq On : 06 Jan 2022 08:11
Operator : JC/MD
Sample : N1020-09MSD
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD


Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022

Quant Time: Jan 06 12:26:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 734
Delta R.T. 0.006 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD




Tgt Ion:168 Resp: 106058
Ion Ratio Lower Upper
168 100
99 60.5 45.8 68.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#2
Dichlorodifluoromethane
Concen: 48.185 ug/l
RT: 1.172 min Scan# 14
Delta R.T. -0.001 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Tgt Ion: 85 Resp: 53672
Ion Ratio Lower Upper
85 100
87 32.5 17.0 50.9





#3
Chloromethane
Concen: 38.379 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD




Tgt Ion: 50 Resp: 49762
Ion Ratio Lower Upper
50 100
52 33.8 25.4 38.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#4
Vinyl Chloride
Concen: 41.802 ug/l
RT: 1.380 min Scan# 48
Delta R.T. 0.006 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Tgt Ion: 62 Resp: 57042
Ion Ratio Lower Upper
62 100
64 32.1 25.4 38.2





#5
Bromomethane
Concen: 24.554 ug/l
RT: 1.617 min Scan# 81
Delta R.T. 0.006 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD




Tgt Ion: 94 Resp: 20678
Ion Ratio Lower Upper
94 100
96 91.7 75.0 112.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#6
Chloroethane
Concen: 33.306 ug/l
RT: 1.691 min Scan# 99
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Tgt Ion: 64 Resp: 31958
Ion Ratio Lower Upper
64 100
66 33.5 25.9 38.9





#7
 Trichlorofluoromethane
 Concen: 39.713 ug/l
 RT: 1.892 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX026364.D
 Acq: 06 Jan 2022 08:11




Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7AMSD



Tgt Ion:101 Resp: 93350
 Ion Ratio Lower Upper
 101 100
 103 64.4 54.1 81.1




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
 Supervised By :Semsettin Yesilyurt 01/07/2022



#8
 Diethyl Ether
 Concen: 42.691 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. -0.000 min
 Lab File: VX026364.D
 Acq: 06 Jan 2022 08:11

Tgt Ion: 74 Resp: 34473
 Ion Ratio Lower Upper
 74 100
 45 102.2 50.4 151.2









#9
1,1,2-Trichlorotrifluoroethane
Concen: 46.714 ug/l
RT: 2.337 min Scan# 204
Delta R.T. 0.006 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11

Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD

Tgt Ion:101 Resp: 5788
Ion Ratio Lower Upper
101 100
85 45.0 36.5 54.7
151 77.1 60.6 90.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#10
Methyl Iodide
Concen: 33.257 ug/l
RT: 2.459 min Scan# 225
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11


Tgt Ion:142 Resp: 53295
Ion Ratio Lower Upper
142 100
127 48.6 37.5 56.3
141 14.4 11.4 17.0





#11
Tert butyl alcohol
Concen: 217.530 ug/l m
RT: 2.983 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD




Tgt Ion: 59 Resp: 66664
Ion Ratio Lower Upper
59 100
57 4.2 8.2 12.2

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#12
1,1-Dichloroethene
Concen: 47.657 ug/l
RT: 2.325 min Scan# 203
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Tgt Ion: 96 Resp: 55969
Ion Ratio Lower Upper
96 100
61 170.0 127.5 191.3
98 65.4 52.0 78.0





#13
Acrolein
Concen: 373.977 ug/l
RT: 2.239 min Scan# 189
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD




Tgt Ion: 56 Resp: 40982
Ion Ratio Lower Upper
56 100
55 995.8 57.4 86.2

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#14
Allyl chloride
Concen: 45.785 ug/l
RT: 2.672 min Scan# 260
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Tgt Ion: 41 Resp: 106263
Ion Ratio Lower Upper
41 100
39 66.7 55.4 83.0
76 30.5 27.6 41.4





#15
Acrylonitrile
Concen: 226.292 ug/l
RT: 3.068 min Scan# 31
Delta R.T. -0.001 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD



Tgt Ion: 53 Resp: 171700
Ion Ratio Lower Upper
53 100
52 83.2 66.2 99.2
51 36.9 29.4 44.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#16
Acetone
Concen: 193.319 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11

Tgt Ion: 43 Resp: 160712
Ion Ratio Lower Upper
43 100
58 28.7 24.8 37.2









#17
Carbon Disulfide
Concen: 45.765 ug/l
RT: 2.520 min Scan# 21
Delta R.T. 0.006 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11

Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD

Tgt Ion: 76 Resp: 15425
Ion Ratio Lower Upper
76 100
78 9.2 7.2 10.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#18
Methyl Acetate
Concen: 41.291 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11

Tgt Ion: 43 Resp: 111344
Ion Ratio Lower Upper
43 100
74 22.4 19.1 28.7









#19
Methyl tert-butyl Ether
Concen: 52.178 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11

Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD

Tgt Ion: 73 Resp: 214039
Ion Ratio Lower Upper
73 100
57 30.1 17.0 25.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#20
Methylene Chloride
Concen: 44.850 ug/l
RT: 2.794 min Scan# 280
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11


Tgt Ion: 84 Resp: 59747
Ion Ratio Lower Upper
84 100
49 125.5 97.1 145.7
51 38.6 29.3 43.9
86 66.9 50.8 76.2





#21
trans-1,2-Dichloroethene
Concen: 47.315 ug/l
RT: 3.099 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD




Tgt Ion: 96 Resp: 60619
Ion Ratio Lower Upper
96 100
61 145.4 105.7 158.5
98 62.4 52.1 78.1

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#22
Diisopropyl ether
Concen: 48.934 ug/l
RT: 3.763 min Scan# 439
Delta R.T. -0.007 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Tgt Ion: 45 Resp: 216767
Ion Ratio Lower Upper
45 100
43 72.7 63.8 95.6
87 26.8 24.2 36.4
59 11.2 9.4 14.2





#23
 Vinyl Acetate
 Concen: 214.667 ug/l
 RT: 3.727 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VX026364.D
 Acq: 06 Jan 2022 08:11




Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7AMSD




Tgt Ion: 43 Resp: 755086
 Ion Ratio Lower Upper
 43 100
 86 10.4 9.0 13.6

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
 Supervised By :Semsettin Yesilyurt 01/07/2022




#24
 1,1-Dichloroethane
 Concen: 49.173 ug/l
 RT: 3.617 min Scan# 415
 Delta R.T. 0.000 min
 Lab File: VX026364.D
 Acq: 06 Jan 2022 08:11




Tgt Ion: 63 Resp: 118047
 Ion Ratio Lower Upper
 63 100
 98 7.1 3.3 9.9
 100 4.4 2.2 6.6





#25
2-Butanone
Concen: 211.265 ug/l
RT: 4.562 min Scan# 51
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD



Tgt Ion: 43 Resp: 244654
Ion Ratio Lower Upper
43 100
72 24.2 21.0 31.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#26
2,2-Dichloropropane
Concen: 41.801 ug/l
RT: 4.483 min Scan# 557
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11

Tgt Ion: 77 Resp: 75607
Ion Ratio Lower Upper
77 100
97 23.0 11.7 35.0






#27
cis-1,2-Dichloroethene
Concen: 48.743 ug/l
RT: 4.495 min Scan# 51
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11

Instrument :

MSVOA_X

ClientSampleId :

MW-7AMSD






Tgt Ion: 96 Resp: 7104
Ion Ratio Lower Upper
96 100
61 148.3 0.0 276.8
98 64.3 0.0 130.6


Manual Integrations

APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#28
Bromochloromethane
Concen: 47.205 ug/l
RT: 4.903 min Scan# 626
Delta R.T. -0.001 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Tgt Ion: 49 Resp: 50227
Ion Ratio Lower Upper
49 100
129 1.9 0.0 4.4
130 78.3 67.1 100.7





#29
Tetrahydrofuran
Concen: 221.316 ug/l
RT: 5.013 min Scan# 644
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD



Tgt Ion: 42 Resp: 16154
Ion Ratio Lower Upper
42 100
72 43.0 37.0 55.6
71 39.9 34.1 51.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#30
Chloroform
Concen: 47.994 ug/l
RT: 5.099 min Scan# 658
Delta R.T. -0.006 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11


Tgt Ion: 83 Resp: 121863
Ion Ratio Lower Upper
83 100
85 67.6 50.6 75.8





#31
Cyclohexane
Concen: 70.749 ug/l
RT: 5.477 min Scan# 706
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD




Tgt Ion: 56 Resp: 15459
Ion Ratio Lower Upper
56 100
69 29.5 23.9 35.9
84 87.4 73.2 109.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#32
1,1,1-Trichloroethane
Concen: 49.410 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.006 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Tgt Ion: 97 Resp: 109066
Ion Ratio Lower Upper
97 100
99 64.3 51.5 77.3
61 43.4 35.5 53.3





#33
1,2-Dichloroethane-d4
Concen: 49.608 ug/l
RT: 5.964 min Scan# 800
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD




Tgt Ion: 65 Resp: 76110
Ion Ratio Lower Upper
65 100
67 51.0 0.0 103.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 931
Delta R.T. -0.006 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Tgt Ion:114 Resp: 177525
Ion Ratio Lower Upper
114 100
63 21.8 0.0 40.0
88 16.3 0.0 30.6





#35
Dibromofluoromethane
Concen: 50.212 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD




Tgt Ion:113 Resp: 57952
Ion Ratio Lower Upper
113 100
111 102.2 81.8 122.8
192 19.5 15.0 22.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#36
1,1-Dichloropropene
Concen: 48.519 ug/l
RT: 5.702 min Scan# 757
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Tgt Ion: 75 Resp: 90319
Ion Ratio Lower Upper
75 100
110 34.5 18.2 54.6
77 31.2 24.7 37.1





#37
Ethyl Acetate
Concen: 43.772 ug/l
RT: 4.721 min Scan# 596
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD



Tgt Ion: 43 Resp: 94479
Ion Ratio Lower Upper
43 100
61 13.3 10.7 16.1
70 10.6 8.5 12.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#38
Carbon Tetrachloride
Concen: 45.580 ug/l
RT: 5.684 min Scan# 754
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11

Tgt Ion:117 Resp: 94381
Ion Ratio Lower Upper
117 100
119 96.2 77.2 115.8
121 30.6 23.8 35.8





#39

Methylcyclohexane

Concen: 51.613 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX026364.D

Acq: 06 Jan 2022 08:11

Instrument :

MSVOA_X

ClientSampleId :

MW-7AMSD

Tgt Ion: 83 Resp: 10837

Ion Ratio Lower Upper

83 100

55 79.7 57.8 86.8





98 49.4 40.5 60.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/07/2022

Supervised By :Semsettin Yesilyurt 01/07/2022



#40

Benzene

Concen: 386.655 ug/l

RT: 6.044 min Scan# 813

Delta R.T. -0.000 min

Lab File: VX026364.D




Acq: 06 Jan 2022 08:11


Tgt Ion: 78 Resp: 2054221

Ion Ratio Lower Upper

78 100


77 24.0 18.6 27.8





#41
Methacrylonitrile
Concen: 46.775 ug/l
RT: 4.928 min Scan# 611
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD



Tgt Ion: 41 Resp: 57012
Ion Ratio Lower Upper
41 100
39 56.7 44.2 66.2
67 65.9 59.7 89.5
52 29.0 24.2 36.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#42
1,2-Dichloroethane
Concen: 54.590 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11


Tgt Ion: 62 Resp: 117375
Ion Ratio Lower Upper
62 100
98 7.7 0.0 18.8





#43
Isopropyl Acetate
Concen: 45.842 ug/l
RT: 6.342 min Scan# 862
Delta R.T. -0.006 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD



Tgt Ion: 43 Resp: 16126
Ion Ratio Lower Upper
43 100
61 18.7 15.7 23.5
87 12.4 11.2 16.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#44
Trichloroethene
Concen: 50.370 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11

Tgt Ion:130 Resp: 72008
Ion Ratio Lower Upper
130 100
95 100.8 0.0 186.4





#45

1,2-Dichloropropane

Concen: 48.003 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VX026364.D

Acq: 06 Jan 2022 08:11

Instrument :

MSVOA_X

ClientSampleId :

MW-7AMSD

Tgt Ion: 63 Resp: 6523

Ion Ratio Lower Upper





63 100

65 30.0 25.8 38.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#46

Dibromomethane

Concen: 45.203 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. -0.000 min

Lab File: VX026364.D

Acq: 06 Jan 2022 08:11




Tgt Ion: 93 Resp: 45563


Ion Ratio Lower Upper

93 100

95 83.9 66.6 100.0

174 98.3 77.3 115.9





#47

Bromodichloromethane

Concen: 44.632 ug/l

RT: 7.824 min Scan# 1105

Delta R.T. -0.000 min

Lab File: VX026364.D

Acq: 06 Jan 2022 08:11

Instrument :

MSVOA_X

ClientSampleId :

MW-7AMSD





Tgt Ion: 83 Resp: 91520

Ion	Ratio	Lower	Upper
83	100		
85	64.8	53.5	80.3
127	8.3	6.6	10.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#48

Methyl methacrylate

Concen: 47.968 ug/l

RT: 7.696 min Scan# 1084




Delta R.T. -0.000 min


Lab File: VX026364.D

Acq: 06 Jan 2022 08:11

Tgt Ion: 41 Resp: 80926


Ion	Ratio	Lower	Upper
41	100		
69	76.9	67.3	100.9
39	57.8	50.2	75.2





#49
1,4-Dioxane
Concen: 802.303 ug/l
RT: 7.690 min Scan# 1082
Delta R.T. 0.006 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD



Tgt Ion: 88 Resp: 24944
Ion Ratio Lower Upper
88 100
43 38.4 25.3 37.9
58 71.8 51.8 77.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#50
Toluene-d8
Concen: 48.552 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11

Tgt Ion: 98 Resp: 208432
Ion Ratio Lower Upper
98 100
100 66.6 52.6 78.8





#51

4-Methyl-2-Pentanone

Concen: 224.342 ug/l

RT: 8.574 min Scan# 1228

Delta R.T. -0.006 min

Lab File: VX026364.D

Acq: 06 Jan 2022 08:11

Instrument :

MSVOA_X

ClientSampleId :

MW-7AMSD

Tgt Ion: 43 Resp: 49111

Ion Ratio Lower Upper





43 100

58 37.9 30.8 46.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#52

Toluene

Concen: 109.680 ug/l

RT: 8.720 min Scan# 1252

Delta R.T. -0.000 min

Lab File: VX026364.D




Acq: 06 Jan 2022 08:11


Tgt Ion: 92 Resp: 358730

Ion Ratio Lower Upper

92 100

91 170.3 136.0 204.0





#53

t-1,3-Dichloropropene

Concen: 41.321 ug/l

RT: 8.982 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX026364.D

Acq: 06 Jan 2022 08:11

Instrument :

MSVOA_X

ClientSampleId :

MW-7AMSD

Tgt Ion: 75 Resp: 8937

Ion Ratio Lower Upper





75 100

77 32.0 25.1 37.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#54

cis-1,3-Dichloropropene

Concen: 45.381 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.006 min

Lab File: VX026364.D

Acq: 06 Jan 2022 08:11




Tgt Ion: 75 Resp: 98362


Ion Ratio Lower Upper

75 100

77 31.1 25.3 37.9


39 55.5 41.4 62.0





#55
1,1,2-Trichloroethane
Concen: 46.721 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD




Tgt Ion: 97 Resp: 62804
Ion Ratio Lower Upper
97 100
83 87.2 70.5 105.7
85 56.6 43.8 65.6
99 64.1 49.4 74.2

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#56
Ethyl methacrylate
Concen: 50.350 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11



Tgt Ion: 69 Resp: 104386
Ion Ratio Lower Upper
69 100
41 71.4 52.8 79.2
39 39.8 32.5 48.7









#57
1,3-Dichloropropane
Concen: 47.199 ug/l
RT: 9.311 min Scan# 1157
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11

Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD

Tgt Ion: 76 Resp: 108679
Ion Ratio Lower Upper
76 100
78 32.2 26.0 39.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#58
2-Chloroethyl Vinyl ether
Concen: 10.428 ug/l
RT: 8.141 min Scan# 1157
Delta R.T. -0.103 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11


Tgt Ion: 63 Resp: 691
Ion Ratio Lower Upper
63 100
106 0.0 24.4 36.6#





#59
2-Hexanone
Concen: 220.092 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD




Tgt Ion: 43 Resp: 36496
Ion Ratio Lower Upper
43 100
58 53.0 26.9 80.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#60
Dibromochloromethane
Concen: 43.558 ug/l
RT: 9.525 min Scan# 1384
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Tgt Ion:129 Resp: 66329
Ion Ratio Lower Upper
129 100
127 76.7 38.6 115.7





#61
1,2-Dibromoethane
Concen: 46.233 ug/l
RT: 9.610 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD




Tgt Ion:107 Resp: 6649
Ion Ratio Lower Upper
107 100
109 92.6 75.1 112.7

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#62
4-Bromofluorobenzene
Concen: 52.237 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.006 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11



Tgt Ion: 95 Resp: 79832
Ion Ratio Lower Upper
95 100
174 73.5 0.0 157.4
176 70.3 0.0 154.6






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11

Instrument :

MSVOA_X

ClientSampleId :

MW-7AMSD






Tgt Ion:117 Resp: 153769
Ion Ratio Lower Upper
117 100
82 57.6 43.8 65.6
119 32.2 25.2 37.8


Manual Integrations

APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#64
Tetrachloroethene
Concen: 55.788 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Tgt Ion:164 Resp: 72392
Ion Ratio Lower Upper
164 100
166 127.3 99.4 149.0
129 96.9 78.6 118.0
131 97.2 75.4 113.0





#65
Chlorobenzene
Concen: 48.135 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD




Tgt Ion:112 Resp: 16259
Ion Ratio Lower Upper
112 100
114 31.3 26.6 39.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#66
1,1,1,2-Tetrachloroethane
Concen: 48.168 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Tgt Ion:131 Resp: 62001
Ion Ratio Lower Upper
131 100
133 95.4 47.1 141.3
119 62.3 31.9 95.9





#67
Ethyl Benzene
Concen: 83.654 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD




Tgt Ion: 91 Resp: 505268
Ion Ratio Lower Upper
91 100
106 30.5 25.7 38.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#68
m/p-Xylenes
Concen: 249.240 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11



Tgt Ion:106 Resp: 562087
Ion Ratio Lower Upper
106 100
91 206.4 158.6 238.0









#69
o-Xylene
Concen: 60.448 ug/l
RT: 10.646 min Scan# 1568
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11

Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD

Tgt Ion:106 Resp: 13338
Ion Ratio Lower Upper
106 100
91 219.5 105.5 316.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#70
Styrene
Concen: 51.643 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11


Tgt Ion:104 Resp: 186810
Ion Ratio Lower Upper
104 100
78 55.1 39.8 59.6
103 56.1 43.0 64.6





#71
Bromoform
Concen: 41.612 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD




Tgt Ion:173 Resp: 43452
Ion Ratio Lower Upper
173 100
175 48.8 24.4 73.4
254 0.0 0.1 0.1

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Tgt Ion:152 Resp: 71852
Ion Ratio Lower Upper
152 100
115 88.5 41.3 123.9
150 176.4 0.0 351.8





#73
Isopropylbenzene
Concen: 65.644 ug/l
RT: 10.963 min Scan# 1601
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD




Tgt Ion:105 Resp: 36662
Ion Ratio Lower Upper
105 100
120 25.9 13.6 40.7

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#74
N-ethyl acetate
Concen: 45.765 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11



Tgt Ion: 43 Resp: 117367
Ion Ratio Lower Upper
43 100
70 41.1 34.1 51.1
55 22.7 19.2 28.8
61 23.5 19.0 28.4









#75
1,1,2,2-Tetrachloroethane
Concen: 43.655 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11

Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD

Tgt Ion: 83 Resp: 8292
Ion Ratio Lower Upper
83 100
131 10.6 5.1 15.3
85 64.6 31.6 95.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#76
1,2,3-Trichloropropane
Concen: 47.373 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. -0.006 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11


Tgt Ion: 75 Resp: 82160
Ion Ratio Lower Upper
75 100
77 36.0 19.0 57.0





#77
Bromobenzene
Concen: 51.993 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD




Tgt Ion:156 Resp: 68864
Ion Ratio Lower Upper
156 100
77 170.0 78.2 234.6
158 96.6 48.6 145.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#78
n-propylbenzene
Concen: 72.153 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11



Tgt Ion: 91 Resp: 463333
Ion Ratio Lower Upper
91 100
120 23.0 12.0 36.1









#79
2-Chlorotoluene
Concen: 53.209 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11

Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD

Tgt Ion: 91 Resp: 21032
Ion Ratio Lower Upper
91 100
126 31.8 17.8 53.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#80
1,3,5-Trimethylbenzene
Concen: 58.967 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11


Tgt Ion:105 Resp: 273303
Ion Ratio Lower Upper
105 100
120 47.7 24.9 74.7





#81
trans-1,4-Dichloro-2-butene
Concen: 32.915 ug/l
RT: 11.024 min Scan# 1630
Delta R.T. 0.006 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD




Tgt Ion: 75 Resp: 1973
Ion Ratio Lower Upper
75 100
53 123.2 79.5 119.3
89 46.9 38.3 57.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#82
4-Chlorotoluene
Concen: 53.113 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Tgt Ion: 91 Resp: 237105
Ion Ratio Lower Upper
91 100
126 28.1 15.6 46.7





#83
tert-Butylbenzene
Concen: 50.208 ug/l
RT: 11.713 min Scan# 1174
Delta R.T. -0.006 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD




Tgt Ion:119 Resp: 225870
Ion Ratio Lower Upper
119 100
91 64.2 28.7 86.1
134 24.3 11.9 35.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#84
1,2,4-Trimethylbenzene
Concen: 73.432 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Tgt Ion:105 Resp: 340077
Ion Ratio Lower Upper
105 100
120 44.1 22.9 68.7





#85
 sec-Butylbenzene
 Concen: 52.070 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX026364.D
 Acq: 06 Jan 2022 08:11




Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7AMSD




Tgt Ion:105 Resp: 29232
 Ion Ratio Lower Upper
 105 100
 134 20.2 10.4 31.1

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
 Supervised By :Semsettin Yesilyurt 01/07/2022




#86
 p-Isopropyltoluene
 Concen: 53.010 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. -0.000 min
 Lab File: VX026364.D
 Acq: 06 Jan 2022 08:11



Tgt Ion:119 Resp: 248094
 Ion Ratio Lower Upper
 119 100
 134 25.6 13.2 39.6
 91 25.3 11.9 35.9





#87

1,3-Dichlorobenzene

Concen: 51.114 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX026364.D

Acq: 06 Jan 2022 08:11

Instrument :

MSVOA_X

ClientSampleId :

MW-7AMSD

Tgt Ion:146 Resp: 127089

Ion Ratio Lower Upper

146 100

111 42.1 19.9 59.6





148 63.5 31.9 95.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/07/2022

Supervised By :Semsettin Yesilyurt 01/07/2022



#88

1,4-Dichlorobenzene

Concen: 48.133 ug/l

RT: 12.042 min Scan# 1797

Delta R.T. -0.001 min

Lab File: VX026364.D

Acq: 06 Jan 2022 08:11




Tgt Ion:146 Resp: 128314


Ion Ratio Lower Upper

146 100

111 41.5 19.9 59.7


148 63.8 31.9 95.7





#89
n-Butylbenzene
Concen: 50.998 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD



Tgt Ion: 91 Resp: 21211
Ion Ratio Lower Upper
91 100
92 54.8 27.3 81.8
134 26.6 13.5 40.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#90
Hexachloroethane
Concen: 38.144 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11


Tgt Ion:117 Resp: 35656
Ion Ratio Lower Upper
117 100
201 70.9 35.4 106.3





#91
1,2-Dichlorobenzene
Concen: 47.261 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD




Tgt Ion:146 Resp: 12189
Ion Ratio Lower Upper
146 100
111 44.1 20.6 61.9
148 64.4 31.9 95.7

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#92
1,2-Dibromo-3-Chloropropane
Concen: 43.572 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Tgt Ion: 75 Resp: 21187
Ion Ratio Lower Upper
75 100
155 80.2 44.0 131.8
157 103.1 55.5 166.5





#93
1,2,4-Trichlorobenzene
Concen: 50.485 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD




Tgt Ion:180 Resp: 76948
Ion Ratio Lower Upper
180 100
182 95.2 47.2 141.6
145 33.5 16.3 48.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#94
Hexachlorobutadiene
Concen: 48.848 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Tgt Ion:225 Resp: 30913
Ion Ratio Lower Upper
225 100
223 63.5 31.6 95.0
227 62.8 32.5 97.5





#95
Naphthalene
Concen: 97.235 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11




Instrument :
MSVOA_X
ClientSampleId :
MW-7AMSD




Tgt Ion:128 Resp: 539964
Ion Ratio Lower Upper
128 100
127 12.9 10.3 15.5
129 11.1 8.8 13.2

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#96
1,2,3-Trichlorobenzene
Concen: 49.967 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX026364.D
Acq: 06 Jan 2022 08:11



Tgt Ion:180 Resp: 77484
Ion Ratio Lower Upper
180 100
182 95.3 47.5 142.6
145 35.0 16.7 50.1

