

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010522\
 Data File : VX026365.D
 Acq On : 06 Jan 2022 08:34
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Jan 06 12:26:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 01/07/2022
 Supervised By :Semsettin Yesilyurt 01/07/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	108037	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	180126	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	155460	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	70819	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	75194	48.109	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	96.220%		
35) Dibromofluoromethane	5.391	113	57021	48.692	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.380%		
50) Toluene-d8	8.653	98	203266	46.665	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	93.320%		
62) 4-Bromofluorobenzene	11.079	95	75218	48.508	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	97.020%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.172	85	54350	47.899	ug/l	96
3) Chloromethane	1.294	50	49311	37.335	ug/l	97
4) Vinyl Chloride	1.380	62	58217	41.882	ug/l	98
5) Bromomethane	1.617	94	23523	27.421	ug/l	98
6) Chloroethane	1.691	64	32715	33.470	ug/l	99
7) Trichlorofluoromethane	1.892	101	94636	39.523	ug/l	97
8) Diethyl Ether	2.136	74	35453	43.100	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.337	101	59057	46.785	ug/l	100
10) Methyl Iodide	2.459	142	55431	33.956	ug/l	99
11) Tert butyl alcohol	2.995	59	68445m	219.250	ug/l	
12) 1,1-Dichloroethene	2.325	96	56488	47.218	ug/l	94
13) Acrolein	2.239	56	24557	219.987	ug/l	94
14) Allyl chloride	2.672	41	107562	45.496	ug/l	98
15) Acrylonitrile	3.068	53	172595	223.305	ug/l	99
16) Acetone	2.386	43	154795	182.790	ug/l	94
17) Carbon Disulfide	2.520	76	155555	45.306	ug/l	100
18) Methyl Acetate	2.709	43	125507	45.691	ug/l	97
19) Methyl tert-butyl Ether	3.117	73	216825	51.889	ug/l	99
20) Methylene Chloride	2.794	84	63155	46.576	ug/l	97
21) trans-1,2-Dichloroethene	3.099	96	61733	47.302	ug/l	92
22) Diisopropyl ether	3.763	45	220526	48.871	ug/l	98
23) Vinyl Acetate	3.727	43	858260	239.529	ug/l	98
24) 1,1-Dichloroethane	3.617	63	119437	48.841	ug/l	98
25) 2-Butanone	4.562	43	250452	212.310	ug/l	98
26) 2,2-Dichloropropane	4.483	77	78961	42.856	ug/l	98
27) cis-1,2-Dichloroethene	4.495	96	72002	48.494	ug/l	94
28) Bromochloromethane	4.903	49	49845	45.988	ug/l	90
29) Tetrahydrofuran	5.007	42	163582	220.007	ug/l	95
30) Chloroform	5.099	83	124005	47.943	ug/l	96
31) Cyclohexane	5.477	56	108536	48.762	ug/l	93
32) 1,1,1-Trichloroethane	5.391	97	111611	49.637	ug/l	99
36) 1,1-Dichloropropene	5.696	75	91852	48.630	ug/l	98
37) Ethyl Acetate	4.714	43	100544	45.909	ug/l	100
38) Carbon Tetrachloride	5.684	117	97529	46.420	ug/l	99
39) Methylcyclohexane	7.385	83	103283	48.476	ug/l	94
40) Benzene	6.044	78	258883	48.025	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010522\
 Data File : VX026365.D
 Acq On : 06 Jan 2022 08:34
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Jan 06 12:26:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 01/07/2022
 Supervised By :Semsettin Yesilyurt 01/07/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	59540	48.143	ug/l	93
42) 1,2-Dichloroethane	6.092	62	103290	47.346	ug/l	98
43) Isopropyl Acetate	6.342	43	168293	47.150	ug/l	98
44) Trichloroethene	7.129	130	73798	50.877	ug/l	91
45) 1,2-Dichloropropane	7.434	63	66223	48.029	ug/l	100
46) Dibromomethane	7.586	93	46448	45.416	ug/l	98
47) Bromodichloromethane	7.824	83	94244	45.293	ug/l	98
48) Methyl methacrylate	7.696	41	81860	47.820	ug/l	93
49) 1,4-Dioxane	7.690	88	25837	819.025	ug/l #	89
51) 4-Methyl-2-Pentanone	8.574	43	490446	220.802	ug/l	99
52) Toluene	8.720	92	157140	47.351	ug/l	98
53) t-1,3-Dichloropropene	8.982	75	91276	41.578	ug/l	97
54) cis-1,3-Dichloropropene	8.366	75	101831	46.303	ug/l	95
55) 1,1,2-Trichloroethane	9.153	97	63164	46.310	ug/l	97
56) Ethyl methacrylate	9.116	69	103955	49.418	ug/l	95
57) 1,3-Dichloropropane	9.311	76	110438	47.270	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	251820	224.959	ug/l	97
59) 2-Hexanone	9.433	43	362466	215.430	ug/l	98
60) Dibromochloromethane	9.525	129	67720	43.829	ug/l	99
61) 1,2-Dibromoethane	9.610	107	66513	45.577	ug/l	100
64) Tetrachloroethene	9.275	164	74411	56.720	ug/l	97
65) Chlorobenzene	10.079	112	162991	47.727	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.165	131	62827	48.279	ug/l	99
67) Ethyl Benzene	10.195	91	306852	50.251	ug/l	97
68) m/p-Xylenes	10.305	106	228594	100.260	ug/l	92
69) o-Xylene	10.640	106	111097	49.801	ug/l	94
70) Styrene	10.659	104	184113	50.344	ug/l	97
71) Bromoform	10.799	173	43393	41.104	ug/l #	100
73) Isopropylbenzene	10.963	105	292587	53.152	ug/l	98
74) N-amyl acetate	10.848	43	129264	51.139	ug/l	97
75) 1,1,2,2-Tetrachloroethane	11.213	83	80380	42.935	ug/l	98
76) 1,2,3-Trichloropropane	11.244	75	81481m	47.666	ug/l	
77) Bromobenzene	11.201	156	67076	51.381	ug/l	92
78) n-propylbenzene	11.305	91	334330	52.824	ug/l	97
79) 2-Chlorotoluene	11.366	91	201633	51.754	ug/l	96
80) 1,3,5-Trimethylbenzene	11.451	105	243782	53.365	ug/l	97
81) trans-1,4-Dichloro-2-b...	11.018	75	19572	33.124	ug/l #	83
82) 4-Chlorotoluene	11.457	91	231966	52.720	ug/l	96
83) tert-Butylbenzene	11.713	119	225705	50.903	ug/l	93
84) 1,2,4-Trimethylbenzene	11.750	105	244003	53.456	ug/l	97
85) sec-Butylbenzene	11.890	105	288719	52.178	ug/l	99
86) p-Isopropyltoluene	12.012	119	242566	52.585	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	124664	50.870	ug/l	99
88) 1,4-Dichlorobenzene	12.042	146	125679	47.833	ug/l	99
89) n-Butylbenzene	12.335	91	204931	49.990	ug/l	99
90) Hexachloroethane	12.542	117	35827	38.886	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	120959	47.581	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.945	75	21295	44.433	ug/l	90
93) 1,2,4-Trichlorobenzene	13.591	180	75882	50.512	ug/l	98
94) Hexachlorobutadiene	13.725	225	29797	47.772	ug/l	99
95) Naphthalene	13.780	128	278464	50.876	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	75893	49.654	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010522\
Data File : VX026365.D
Acq On : 06 Jan 2022 08:34
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022

Quant Time: Jan 06 12:26:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:168 Resp: 10803
Ion Ratio Lower Upper
168 100
99 65.7 45.8 68.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#2
Dichlorodifluoromethane
Concen: 47.899 ug/l
RT: 1.172 min Scan# 14
Delta R.T. -0.001 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34


Tgt Ion: 85 Resp: 54350
Ion Ratio Lower Upper
85 100
87 31.7 17.0 50.9





#3
Chloromethane
Concen: 37.335 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 50 Resp: 4931
Ion Ratio Lower Upper
50 100
52 33.6 25.4 38.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#4
Vinyl Chloride
Concen: 41.882 ug/l
RT: 1.380 min Scan# 48
Delta R.T. 0.006 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Tgt Ion: 62 Resp: 58217
Ion Ratio Lower Upper
62 100
64 30.6 25.4 38.2





#5
Bromomethane
Concen: 27.421 ug/l
RT: 1.617 min Scan# 81
Delta R.T. 0.006 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 94 Resp: 2352
Ion Ratio Lower Upper
94 100
96 95.3 75.0 112.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#6
Chloroethane
Concen: 33.470 ug/l
RT: 1.691 min Scan# 99
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34



Tgt Ion: 64 Resp: 32715
Ion Ratio Lower Upper
64 100
66 32.1 25.9 38.9





#7
Trichlorofluoromethane
Concen: 39.523 ug/l
RT: 1.892 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion:101 Resp: 94630
Ion Ratio Lower Upper
101 100
103 64.9 54.1 81.1

Manual Integrations




APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#8
Diethyl Ether
Concen: 43.100 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34

Tgt Ion: 74 Resp: 35453
Ion Ratio Lower Upper
74 100
45 102.9 50.4 151.2









#9
1,1,2-Trichlorotrifluoroethane
Concen: 46.785 ug/l
RT: 2.337 min Scan# 204
Delta R.T. 0.006 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion:101 Resp: 5905
Ion Ratio Lower Upper
101 100
85 45.3 36.5 54.7
151 75.7 60.6 90.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#10
Methyl Iodide
Concen: 33.956 ug/l
RT: 2.459 min Scan# 225
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34


Tgt Ion:142 Resp: 55431
Ion Ratio Lower Upper
142 100
127 47.6 37.5 56.3
141 14.1 11.4 17.0





#11
Tert butyl alcohol
Concen: 219.250 ug/l m
RT: 2.995 min Scan# 311
Delta R.T. 0.012 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 59 Resp: 6844
Ion Ratio Lower Upper
59 100
57 4.9 8.2 12.2

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#12
1,1-Dichloroethene
Concen: 47.218 ug/l
RT: 2.325 min Scan# 203
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Tgt Ion: 96 Resp: 56488
Ion Ratio Lower Upper
96 100
61 168.9 127.5 191.3
98 63.7 52.0 78.0





#13
Acrolein
Concen: 219.987 ug/l
RT: 2.239 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 56 Resp: 2455
Ion Ratio Lower Upper
56 100
55 76.7 57.4 86.2

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#14
Allyl chloride
Concen: 45.496 ug/l
RT: 2.672 min Scan# 260
Delta R.T. 0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Tgt Ion: 41 Resp: 107562
Ion Ratio Lower Upper
41 100
39 69.3 55.4 83.0
76 30.5 27.6 41.4





#15
Acrylonitrile
Concen: 223.305 ug/l
RT: 3.068 min Scan# 31
Delta R.T. -0.001 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 53 Resp: 17259
Ion Ratio Lower Upper
53 100
52 83.5 66.2 99.2
51 36.9 29.4 44.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#16
Acetone
Concen: 182.790 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34



Tgt Ion: 43 Resp: 154795
Ion Ratio Lower Upper
43 100
58 27.9 24.8 37.2









#17
Carbon Disulfide
Concen: 45.306 ug/l
RT: 2.520 min Scan# 21
Delta R.T. 0.006 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 76 Resp: 15555
Ion Ratio Lower Upper
76 100
78 8.9 7.2 10.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#18
Methyl Acetate
Concen: 45.691 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34


Tgt Ion: 43 Resp: 125507
Ion Ratio Lower Upper
43 100
74 22.5 19.1 28.7





#19
Methyl tert-butyl Ether
Concen: 51.889 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 73 Resp: 21682
Ion Ratio Lower Upper
73 100
57 21.9 17.0 25.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#20
Methylene Chloride
Concen: 46.576 ug/l
RT: 2.794 min Scan# 280
Delta R.T. 0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Tgt Ion: 84 Resp: 63155
Ion Ratio Lower Upper
84 100
49 126.2 97.1 145.7
51 38.2 29.3 43.9
86 61.8 50.8 76.2





#21
trans-1,2-Dichloroethene
Concen: 47.302 ug/l
RT: 3.099 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 96 Resp: 61733
Ion Ratio Lower Upper
96 100
61 144.2 105.7 158.5
98 62.7 52.1 78.1

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#22
Diisopropyl ether
Concen: 48.871 ug/l
RT: 3.763 min Scan# 439
Delta R.T. -0.007 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Tgt Ion: 45 Resp: 220526
Ion Ratio Lower Upper
45 100
43 79.0 63.8 95.6
87 27.4 24.2 36.4
59 11.1 9.4 14.2





#23
 Vinyl Acetate
 Concen: 239.529 ug/l
 RT: 3.727 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VX026365.D
 Acq: 06 Jan 2022 08:34



Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC




Tgt Ion: 43 Resp: 858260
 Ion Ratio Lower Upper
 43 100
 86 10.4 9.0 13.6

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
 Supervised By :Semsettin Yesilyurt 01/07/2022




#24
 1,1-Dichloroethane
 Concen: 48.841 ug/l
 RT: 3.617 min Scan# 415
 Delta R.T. 0.000 min
 Lab File: VX026365.D
 Acq: 06 Jan 2022 08:34



Tgt Ion: 63 Resp: 119437
 Ion Ratio Lower Upper
 63 100
 98 7.3 3.3 9.9
 100 4.1 2.2 6.6





#25

2-Butanone

Concen: 212.310 ug/l

RT: 4.562 min Scan# 51

Delta R.T. 0.000 min

Lab File: VX026365.D

Acq: 06 Jan 2022 08:34

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 43 Resp: 250452

Ion Ratio Lower Upper





43 100

72 24.9 21.0 31.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#26

2,2-Dichloropropane

Concen: 42.856 ug/l

RT: 4.483 min Scan# 557

Delta R.T. -0.000 min

Lab File: VX026365.D




Acq: 06 Jan 2022 08:34


Tgt Ion: 77 Resp: 78961

Ion Ratio Lower Upper

77 100


97 22.2 11.7 35.0





#27
cis-1,2-Dichloroethene
Concen: 48.494 ug/l
RT: 4.495 min Scan# 511
Delta R.T. 0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 96 Resp: 72002
Ion Ratio Lower Upper
96 100
61 148.3 0.0 276.8
98 65.4 0.0 130.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#28
Bromochloromethane
Concen: 45.988 ug/l
RT: 4.903 min Scan# 626
Delta R.T. -0.001 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34



Tgt Ion: 49 Resp: 49845
Ion Ratio Lower Upper
49 100
129 2.1 0.0 4.4
130 75.0 67.1 100.7





#29

Tetrahydrofuran

Concen: 220.007 ug/l

RT: 5.007 min Scan# 64

Delta R.T. -0.006 min

Lab File: VX026365.D

Acq: 06 Jan 2022 08:34

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 42 Resp: 16358

Ion Ratio Lower Upper

42 100

72 43.0 37.0 55.6





71 39.6 34.1 51.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/07/2022

Supervised By :Semsettin Yesilyurt 01/07/2022



#30

Chloroform

Concen: 47.943 ug/l

RT: 5.099 min Scan# 658

Delta R.T. -0.006 min

Lab File: VX026365.D




Acq: 06 Jan 2022 08:34


Tgt Ion: 83 Resp: 124005

Ion Ratio Lower Upper

83 100


85 66.5 50.6 75.8





#31
Cyclohexane
Concen: 48.762 ug/l
RT: 5.477 min Scan# 706
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 56 Resp: 108530
Ion Ratio Lower Upper
56 100
69 31.0 23.9 35.9
84 83.5 73.2 109.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#32
1,1,1-Trichloroethane
Concen: 49.637 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.006 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34



Tgt Ion: 97 Resp: 111611
Ion Ratio Lower Upper
97 100
99 64.1 51.5 77.3
61 43.5 35.5 53.3





#33

1,2-Dichloroethane-d4

Concen: 48.109 ug/l

RT: 5.958 min Scan# 79

Delta R.T. -0.006 min

Lab File: VX026365.D

Acq: 06 Jan 2022 08:34

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 65 Resp: 75194

Ion Ratio Lower Upper

65 100





67 50.6 0.0 103.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/07/2022

Supervised By :Semsettin Yesilyurt 01/07/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.006 min

Lab File: VX026365.D

Acq: 06 Jan 2022 08:34




Tgt Ion:114 Resp: 180126


Ion Ratio Lower Upper

114 100

63 22.5 0.0 40.0


88 17.3 0.0 30.6





#35
Dibromofluoromethane
Concen: 48.692 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 113 Resp: 5702
Ion Ratio Lower Upper
113 100
111 103.5 81.8 122.8
192 18.8 15.0 22.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#36
1,1-Dichloropropene
Concen: 48.630 ug/l
RT: 5.696 min Scan# 756
Delta R.T. -0.006 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Tgt Ion: 75 Resp: 91852
Ion Ratio Lower Upper
75 100
110 34.2 18.2 54.6
77 30.6 24.7 37.1





#37
Ethyl Acetate
Concen: 45.909 ug/l
RT: 4.714 min Scan# 596
Delta R.T. -0.007 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 100544
Ion Ratio Lower Upper
43 100
61 13.3 10.7 16.1
70 10.5 8.5 12.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#38
Carbon Tetrachloride
Concen: 46.420 ug/l
RT: 5.684 min Scan# 754
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34

Tgt Ion:117 Resp: 97529
Ion Ratio Lower Upper
117 100
119 95.8 77.2 115.8
121 30.5 23.8 35.8





#39

Methylcyclohexane

Concen: 48.476 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX026365.D

Acq: 06 Jan 2022 08:34

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 83 Resp: 10328

Ion Ratio Lower Upper

83 100

55 80.1 57.8 86.8





98 49.6 40.5 60.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/07/2022

Supervised By :Semsettin Yesilyurt 01/07/2022



#40

Benzene

Concen: 48.025 ug/l

RT: 6.044 min Scan# 813

Delta R.T. -0.000 min

Lab File: VX026365.D




Acq: 06 Jan 2022 08:34


Tgt Ion: 78 Resp: 258883

Ion Ratio Lower Upper

78 100

77 24.0 18.6 27.8






#41
Methacrylonitrile
Concen: 48.143 ug/l
RT: 4.928 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC






Tgt Ion: 41 Resp: 59540
Ion Ratio Lower Upper
41 100
39 54.6 44.2 66.2
67 63.5 59.7 89.5
52 27.5 24.2 36.2


Manual Integrations

APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#42
1,2-Dichloroethane
Concen: 47.346 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Tgt Ion: 62 Resp: 103290
Ion Ratio Lower Upper
62 100
98 8.7 0.0 18.8





#43
Isopropyl Acetate
Concen: 47.150 ug/l
RT: 6.342 min Scan# 862
Delta R.T. -0.006 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 43 Resp: 16829
Ion Ratio Lower Upper
43 100
61 19.1 15.7 23.5
87 12.7 11.2 16.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#44
Trichloroethene
Concen: 50.877 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Tgt Ion:130 Resp: 73798
Ion Ratio Lower Upper
130 100
95 102.0 0.0 186.4





#45
1,2-Dichloropropane
Concen: 48.029 ug/l
RT: 7.434 min Scan# 1041
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 63 Resp: 6622
Ion Ratio Lower Upper
63 100
65 32.5 25.8 38.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#46
Dibromomethane
Concen: 45.416 ug/l
RT: 7.586 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34



Tgt Ion: 93 Resp: 46448
Ion Ratio Lower Upper
93 100
95 82.0 66.6 100.0
174 98.6 77.3 115.9






#47
Bromodichloromethane
Concen: 45.293 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC






Tgt Ion: 83 Resp: 94244
Ion Ratio Lower Upper
83 100
85 65.3 53.5 80.3
127 8.5 6.6 10.0


Manual Integrations

APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#48
Methyl methacrylate
Concen: 47.820 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Tgt Ion: 41 Resp: 81860
Ion Ratio Lower Upper
41 100
69 76.6 67.3 100.9
39 57.8 50.2 75.2





#49
1,4-Dioxane
Concen: 819.025 ug/l
RT: 7.690 min Scan# 1082
Delta R.T. 0.006 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 88 Resp: 2583
Ion Ratio Lower Upper
88 100
43 38.3 25.3 37.9
58 72.4 51.8 77.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#50
Toluene-d8
Concen: 46.665 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34

Tgt Ion: 98 Resp: 203266
Ion Ratio Lower Upper
98 100
100 66.6 52.6 78.8





#51
4-Methyl-2-Pentanone
Concen: 220.802 ug/l
RT: 8.574 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 43 Resp: 490440
Ion Ratio Lower Upper
43 100
58 37.9 30.8 46.2

Manual Integrations




APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#52
Toluene
Concen: 47.351 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34

Tgt Ion: 92 Resp: 157140
Ion Ratio Lower Upper
92 100
91 173.1 136.0 204.0





#53

t-1,3-Dichloropropene

Concen: 41.578 ug/l

RT: 8.982 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VX026365.D

Acq: 06 Jan 2022 08:34

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 75 Resp: 91270

Ion Ratio Lower Upper





75 100

77 32.8 25.1 37.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#54

cis-1,3-Dichloropropene

Concen: 46.303 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.006 min

Lab File: VX026365.D

Acq: 06 Jan 2022 08:34




Tgt Ion: 75 Resp: 101831


Ion Ratio Lower Upper

75 100

77 32.3 25.3 37.9


39 56.5 41.4 62.0





#55
1,1,2-Trichloroethane
Concen: 46.310 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 97 Resp: 63164
Ion Ratio Lower Upper
97 100
83 82.7 70.5 105.7
85 54.3 43.8 65.6
99 61.3 49.4 74.2

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#56
Ethyl methacrylate
Concen: 49.418 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34



Tgt Ion: 69 Resp: 103955
Ion Ratio Lower Upper
69 100
41 72.3 52.8 79.2
39 40.3 32.5 48.7





#57
1,3-Dichloropropane
Concen: 47.270 ug/l
RT: 9.311 min Scan# 110438
Delta R.T. 0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 76 Resp: 110438
Ion Ratio Lower Upper
76 100
78 31.8 26.0 39.0

Manual Integrations




APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#58
2-Chloroethyl Vinyl ether
Concen: 224.959 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. 0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34

Tgt Ion: 63 Resp: 251820
Ion Ratio Lower Upper
63 100
106 28.9 24.4 36.6





#59

2-Hexanone

Concen: 215.430 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.000 min

Lab File: VX026365.D

Acq: 06 Jan 2022 08:34

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 43 Resp: 362460

Ion Ratio Lower Upper





43 100

58 52.6 26.9 80.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#60

Dibromochloromethane

Concen: 43.829 ug/l

RT: 9.525 min Scan# 1384

Delta R.T. -0.000 min

Lab File: VX026365.D




Acq: 06 Jan 2022 08:34


Tgt Ion:129 Resp: 67720

Ion Ratio Lower Upper

129 100

127 76.5 38.6 115.7





#61
1,2-Dibromoethane
Concen: 45.577 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion:107 Resp: 6651
Ion Ratio Lower Upper
107 100
109 93.9 75.1 112.7

Manual Integrations




APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#62
4-Bromofluorobenzene
Concen: 48.508 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.006 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34


Tgt Ion: 95 Resp: 75218
Ion Ratio Lower Upper
95 100
174 72.9 0.0 157.4
176 69.6 0.0 154.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion:117 Resp: 155460
Ion Ratio Lower Upper
117 100
82 56.2 43.8 65.6
119 31.2 25.2 37.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#64
Tetrachloroethene
Concen: 56.720 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Tgt Ion:164 Resp: 74411
Ion Ratio Lower Upper
164 100
166 129.3 99.4 149.0
129 99.7 78.6 118.0
131 96.2 75.4 113.0





#65
Chlorobenzene
Concen: 47.727 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion:112 Resp: 16299
Ion Ratio Lower Upper
112 100
114 32.0 26.6 39.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#66
1,1,1,2-Tetrachloroethane
Concen: 48.279 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Tgt Ion:131 Resp: 62827
Ion Ratio Lower Upper
131 100
133 95.4 47.1 141.3
119 63.0 31.9 95.9





#67
Ethyl Benzene
Concen: 50.251 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 91 Resp: 306852
Ion Ratio Lower Upper
91 100
106 30.3 25.7 38.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#68
m/p-Xylenes
Concen: 100.260 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34



Tgt Ion:106 Resp: 228594
Ion Ratio Lower Upper
106 100
91 209.7 158.6 238.0









#69
o-Xylene
Concen: 49.801 ug/l
RT: 10.640 min Scan# 11109
Delta R.T. -0.006 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion:106 Resp: 11109
Ion Ratio Lower Upper
106 100
91 221.0 105.5 316.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#70
Styrene
Concen: 50.344 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34


Tgt Ion:104 Resp: 184113
Ion Ratio Lower Upper
104 100
78 53.7 39.8 59.6
103 54.6 43.0 64.6





#71
Bromoform
Concen: 41.104 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:173 Resp: 4339
Ion Ratio Lower Upper
173 100
175 49.2 24.4 73.4
254 0.0 0.1 0.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34

Tgt Ion:152 Resp: 70819
Ion Ratio Lower Upper
152 100
115 88.6 41.3 123.9
150 175.9 0.0 351.8





#73
Isopropylbenzene
Concen: 53.152 ug/l
RT: 10.963 min Scan# 1601
Delta R.T. 0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion:105 Resp: 29258
Ion Ratio Lower Upper
105 100
120 26.3 13.6 40.7

Manual Integrations




APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#74
N-ethyl acetate
Concen: 51.139 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34


Tgt Ion: 43 Resp: 129264
Ion Ratio Lower Upper
43 100
70 40.3 34.1 51.1
55 22.8 19.2 28.8
61 23.1 19.0 28.4





#75
1,1,2,2-Tetrachloroethane
Concen: 42.935 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 83 Resp: 80380
Ion Ratio Lower Upper
83 100
131 10.9 5.1 15.3
85 64.7 31.6 95.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#76
1,2,3-Trichloropropane
Concen: 47.666 ug/l m
RT: 11.244 min Scan# 1666
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Tgt Ion: 75 Resp: 81481
Ion Ratio Lower Upper
75 100
77 36.0 19.0 57.0





#77
Bromobenzene
Concen: 51.381 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion:156 Resp: 67070
Ion Ratio Lower Upper
156 100
77 171.9 78.2 234.6
158 95.9 48.6 145.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#78
n-propylbenzene
Concen: 52.824 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34



Tgt Ion: 91 Resp: 334330
Ion Ratio Lower Upper
91 100
120 22.8 12.0 36.1





#79
2-Chlorotoluene
Concen: 51.754 ug/l
RT: 11.366 min Scan# 1100
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 91 Resp: 201633
Ion Ratio Lower Upper
91 100
126 32.9 17.8 53.3

Manual Integrations




APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#80
1,3,5-Trimethylbenzene
Concen: 53.365 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34


Tgt Ion:105 Resp: 243782
Ion Ratio Lower Upper
105 100
120 47.6 24.9 74.7





#81
trans-1,4-Dichloro-2-butene
Concen: 33.124 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 75 Resp: 19572
Ion Ratio Lower Upper
75 100
53 123.3 79.5 119.3
89 47.0 38.3 57.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#82
4-Chlorotoluene
Concen: 52.720 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Tgt Ion: 91 Resp: 231966
Ion Ratio Lower Upper
91 100
126 28.7 15.6 46.7





#83
tert-Butylbenzene
Concen: 50.903 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. -0.006 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion:119 Resp: 225708
Ion Ratio Lower Upper
119 100
91 63.9 28.7 86.1
134 23.1 11.9 35.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#84
1,2,4-Trimethylbenzene
Concen: 53.456 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.006 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Tgt Ion:105 Resp: 244003
Ion Ratio Lower Upper
105 100
120 43.7 22.9 68.7





#85
 sec-Butylbenzene
 Concen: 52.178 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX026365.D
 Acq: 06 Jan 2022 08:34




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC



Tgt Ion:105 Resp: 288719
 Ion Ratio Lower Upper
 105 100
 134 20.0 10.4 31.1




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
 Supervised By :Semsettin Yesilyurt 01/07/2022



#86
 p-Isopropyltoluene
 Concen: 52.585 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. -0.000 min
 Lab File: VX026365.D
 Acq: 06 Jan 2022 08:34

Tgt Ion:119 Resp: 242566
 Ion Ratio Lower Upper
 119 100
 134 25.9 13.2 39.6
 91 25.2 11.9 35.9





#87

1,3-Dichlorobenzene

Concen: 50.870 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX026365.D


Acq: 06 Jan 2022 08:34

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC






Tgt Ion:146 Resp: 124664

Ion	Ratio	Lower	Upper
146	100		
111	42.1	19.9	59.6
148	63.7	31.9	95.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#88

1,4-Dichlorobenzene


Concen: 47.833 ug/l

RT: 12.042 min Scan# 1797

Delta R.T. -0.001 min



Lab File: VX026365.D


Acq: 06 Jan 2022 08:34



Tgt Ion:146 Resp: 125679

Ion	Ratio	Lower	Upper
146	100		
111	41.2	19.9	59.7
148	64.3	31.9	95.7





#89
n-Butylbenzene
Concen: 49.990 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 91 Resp: 20493
Ion Ratio Lower Upper
91 100
92 54.5 27.3 81.8
134 25.8 13.5 40.4

Manual Integrations




APPROVED


Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#90
Hexachloroethane
Concen: 38.886 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34


Tgt Ion:117 Resp: 35827
Ion Ratio Lower Upper
117 100
201 71.3 35.4 106.3





#91
1,2-Dichlorobenzene
Concen: 47.581 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion:146 Resp: 120955
Ion Ratio Lower Upper
146 100
111 43.3 20.6 61.9
148 62.9 31.9 95.7

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#92
1,2-Dibromo-3-Chloropropane
Concen: 44.433 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34



Tgt Ion: 75 Resp: 21295
Ion Ratio Lower Upper
75 100
155 79.1 44.0 131.8
157 99.6 55.5 166.5






#93
1,2,4-Trichlorobenzene
Concen: 50.512 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC






Tgt Ion:180 Resp: 7588
Ion Ratio Lower Upper
180 100
182 92.4 47.2 141.6
145 32.3 16.3 48.9


Manual Integrations

APPROVED



Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022




#94
Hexachlorobutadiene
Concen: 47.772 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Tgt Ion:225 Resp: 29797
Ion Ratio Lower Upper
225 100
223 64.0 31.6 95.0
227 65.2 32.5 97.5





#95
Naphthalene
Concen: 50.876 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion:128 Resp: 278464
Ion Ratio Lower Upper
128 100
127 13.1 10.3 15.5
129 10.9 8.8 13.2

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 01/07/2022
Supervised By :Semsettin Yesilyurt 01/07/2022



#96
1,2,3-Trichlorobenzene
Concen: 49.654 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX026365.D
Acq: 06 Jan 2022 08:34



Tgt Ion:180 Resp: 75893
Ion Ratio Lower Upper
180 100
182 93.6 47.5 142.6
145 34.5 16.7 50.1

