

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010524\
 Data File : VX039648.D
 Acq On : 05 Jan 2024 14:49
 Operator : JC/MD
 Sample : P1034-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1D3

Quant Time: Jan 05 22:23:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	76651	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	162177	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	175964	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90020	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67231	47.879	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	95.760%		
35) Dibromofluoromethane	5.385	113	53361	47.627	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.260%		
50) Toluene-d8	8.647	98	206343	49.608	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.220%		
62) 4-Bromofluorobenzene	11.079	95	95598	55.703	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	111.400%		
Target Compounds			Qvalue			
12) 1,1-Dichloroethene	2.325	96	390	0.396	ug/l #	65
21) trans-1,2-Dichloroethene	3.099	96	450	0.420	ug/l #	69
27) cis-1,2-Dichloroethene	4.501	96	1683	1.422	ug/l	87
30) Chloroform	5.092	83	872	0.425	ug/l	96
44) Trichloroethene	7.129	130	4083	3.235	ug/l	95
64) Tetrachloroethene	9.275	164	3278	2.851	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010524\
Data File : VX039648.D
Acq On : 05 Jan 2024 14:49
Operator : JC/MD
Sample : P1034-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1D3

Quant Time: Jan 05 22:23:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

