

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX010620\
 Data File : VX014419.D
 Acq On : 06 Jan 2020 11:26
 Operator : JC/SP
 Sample : VX0106MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0106MBL01

Quant Time: Jan 06 16:08:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	584556	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	888702	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	799701	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	360082	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	319485	48.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.44%	
35) Dibromofluoromethane	5.49	113	265936	49.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.46%	
50) Toluene-d8	8.71	98	1038860	49.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.78%	
62) 4-Bromofluorobenzene	11.13	95	360854	46.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.88%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX010620\
Data File : VX014419.D
Acq On : 06 Jan 2020 11:26
Operator : JC/SP
Sample : VX0106MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0106MBL01

Quant Time: Jan 06 16:08:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

