

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010622\
 Data File : VX026370.D
 Acq On : 06 Jan 2022 13:58
 Operator : JC/MD
 Sample : VX0106WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jan 07 04:12:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	103844	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	171623	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	147501	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68143	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	80239	53.410	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.820%
35) Dibromofluoromethane	5.385	113	59629	53.442	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.880%
50) Toluene-d8	8.647	98	210975	50.834	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.660%
62) 4-Bromofluorobenzene	11.079	95	80101	54.216	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.440%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.173	85	21581	19.788	ug/l	97
3) Chloromethane	1.294	50	19775	15.577	ug/l	96
4) Vinyl Chloride	1.380	62	22456	16.807	ug/l	99
5) Bromomethane	1.618	94	10955	13.286	ug/l	99
6) Chloroethane	1.691	64	13936	14.833	ug/l	99
7) Trichlorofluoromethane	1.892	101	37540	16.311	ug/l	94
8) Diethyl Ether	2.136	74	14055	17.777	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.337	101	23301	19.204	ug/l	99
10) Methyl Iodide	2.459	142	20797	13.254	ug/l	98
11) Tert butyl alcohol	2.983	59	21807	72.675	ug/l	98
12) 1,1-Dichloroethene	2.325	96	22024	19.153	ug/l	93
13) Acrolein	2.239	56	10179	94.868	ug/l	98
14) Allyl chloride	2.666	41	41169	18.117	ug/l	97
15) Acrylonitrile	3.069	53	65593	88.291	ug/l	99
16) Acetone	2.386	43	73620	90.445	ug/l	95
17) Carbon Disulfide	2.514	76	58775	17.810	ug/l	100
18) Methyl Acetate	2.703	43	47260	17.900	ug/l	97
19) Methyl tert-butyl Ether	3.117	73	84859	21.128	ug/l	99
20) Methylene Chloride	2.794	84	24596	18.299	ug/l	93
21) trans-1,2-Dichloroethene	3.093	96	23620	18.829	ug/l	90
22) Diisopropyl ether	3.757	45	86506	19.945	ug/l	95
23) Vinyl Acetate	3.721	43	332150	96.442	ug/l	98
24) 1,1-Dichloroethane	3.611	63	46360	19.723	ug/l	99
25) 2-Butanone	4.556	43	99619	87.858	ug/l	97
26) 2,2-Dichloropropane	4.477	77	33459	18.893	ug/l	97
27) cis-1,2-Dichloroethene	4.489	96	28161	19.732	ug/l	93
28) Bromochloromethane	4.898	49	20745	19.912	ug/l	91
29) Tetrahydrofuran	5.007	42	61567	86.147	ug/l	95
30) Chloroform	5.093	83	48427	19.479	ug/l	98
31) Cyclohexane	5.471	56	41439	19.369	ug/l	94
32) 1,1,1-Trichloroethane	5.385	97	42878	19.839	ug/l	99
36) 1,1-Dichloropropene	5.696	75	35291	19.610	ug/l	99
37) Ethyl Acetate	4.715	43	37791	18.111	ug/l	99
38) Carbon Tetrachloride	5.678	117	36713	18.340	ug/l	97
39) Methylcyclohexane	7.385	83	40667	20.033	ug/l	94
40) Benzene	6.038	78	98864	19.249	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010622\
 Data File : VX026370.D
 Acq On : 06 Jan 2022 13:58
 Operator : JC/MD
 Sample : VX0106WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jan 07 04:12:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	22250	18.882	ug/l	94
42) 1,2-Dichloroethane	6.086	62	40982	19.716	ug/l	98
43) Isopropyl Acetate	6.336	43	63477	18.665	ug/l	98
44) Trichloroethene	7.129	130	28551	20.659	ug/l	95
45) 1,2-Dichloropropane	7.434	63	25508	19.416	ug/l	100
46) Dibromomethane	7.580	93	17883	18.352	ug/l	99
47) Bromodichloromethane	7.818	83	36494	18.408	ug/l	95
48) Methyl methacrylate	7.690	41	31108	19.073	ug/l	92
49) 1,4-Dioxane	7.684	88	8796	292.645	ug/l #	89
51) 4-Methyl-2-Pentanone	8.574	43	192601	91.006	ug/l	99
52) Toluene	8.720	92	60992	19.289	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	33143	17.017	ug/l	97
54) cis-1,3-Dichloropropene	8.366	75	38302	18.279	ug/l	98
55) 1,1,2-Trichloroethane	9.153	97	24601	18.931	ug/l	98
56) Ethyl methacrylate	9.116	69	40604	20.259	ug/l	96
57) 1,3-Dichloropropane	9.311	76	43443	19.516	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.238	63	94106	94.207	ug/l	97
59) 2-Hexanone	9.427	43	146043	91.100	ug/l	98
60) Dibromochloromethane	9.519	129	25052	17.017	ug/l	98
61) 1,2-Dibromoethane	9.610	107	26217	18.855	ug/l	99
64) Tetrachloroethene	9.275	164	28263	22.706	ug/l	97
65) Chlorobenzene	10.079	112	64187	19.809	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	24018	19.452	ug/l	99
67) Ethyl Benzene	10.195	91	117403	20.264	ug/l	99
68) m/p-Xylenes	10.305	106	88730	41.017	ug/l	93
69) o-Xylene	10.640	106	43653	20.624	ug/l	94
70) Styrene	10.659	104	71340	20.560	ug/l	97
71) Bromoform	10.799	173	15848	15.822	ug/l #	99
73) Isopropylbenzene	10.963	105	115476	21.801	ug/l	97
74) N-amyl acetate	10.842	43	48984	20.140	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.213	83	32105	17.822	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	37001	22.496	ug/l	90
77) Bromobenzene	11.201	156	26165	20.830	ug/l	93
78) n-propylbenzene	11.305	91	131086	21.525	ug/l	98
79) 2-Chlorotoluene	11.366	91	80291	21.418	ug/l	94
80) 1,3,5-Trimethylbenzene	11.451	105	95955	21.830	ug/l	97
81) trans-1,4-Dichloro-2-b...	11.018	75	7022	12.351	ug/l #	69
82) 4-Chlorotoluene	11.457	91	92661	21.886	ug/l	94
83) tert-Butylbenzene	11.713	119	90596	21.234	ug/l	94
84) 1,2,4-Trimethylbenzene	11.750	105	95742	21.799	ug/l	97
85) sec-Butylbenzene	11.890	105	115427	21.679	ug/l	98
86) p-Isopropyltoluene	12.012	119	96933	21.839	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	49020	20.788	ug/l	97
88) 1,4-Dichlorobenzene	12.043	146	48940	19.358	ug/l	98
89) n-Butylbenzene	12.335	91	82709	20.968	ug/l	99
90) Hexachloroethane	12.542	117	13375	15.087	ug/l	98
91) 1,2-Dichlorobenzene	12.335	146	47293	19.334	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	12.945	75	7537	16.344	ug/l	89
93) 1,2,4-Trichlorobenzene	13.591	180	29196	20.198	ug/l	98
94) Hexachlorobutadiene	13.725	225	12493	20.816	ug/l	98
95) Naphthalene	13.780	128	105418	20.016	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	29564	20.102	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010622\
Data File : VX026370.D
Acq On : 06 Jan 2022 13:58
Operator : JC/MD
Sample : VX0106WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Jan 07 04:12:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010622\
 Data File : VX026370.D
 Acq On : 06 Jan 2022 13:58
 Operator : JC/MD
 Sample : VX0106WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :


Quant Time: Jan 07 04:12:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:168 Resp: 103844
Ion Ratio Lower Upper
168 100
99 64.8 45.8 68.8



#2
Dichlorodifluoromethane
Concen: 19.788 ug/l
RT: 1.173 min Scan# 14
Delta R.T. -0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58

Tgt Ion: 85 Resp: 21581
Ion Ratio Lower Upper
85 100
87 32.0 17.0 50.9









#3
Chloromethane
Concen: 15.577 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 50 Resp: 19775
Ion Ratio Lower Upper
50 100
52 34.0 25.4 38.0



#4
Vinyl Chloride
Concen: 16.807 ug/l
RT: 1.380 min Scan# 48
Delta R.T. 0.006 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58

Tgt Ion: 62 Resp: 22456
Ion Ratio Lower Upper
62 100
64 31.4 25.4 38.2









#5
Bromomethane
Concen: 13.286 ug/l
RT: 1.618 min Scan# 81
Delta R.T. 0.007 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 94 Resp: 10955
Ion Ratio Lower Upper
94 100
96 93.1 75.0 112.4



#6
Chloroethane
Concen: 14.833 ug/l
RT: 1.691 min Scan# 99
Delta R.T. -0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58


Tgt Ion: 64 Resp: 13936
Ion Ratio Lower Upper
64 100
66 32.9 25.9 38.9








#7
 Trichlorofluoromethane
 Concen: 16.311 ug/l
 RT: 1.892 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX026370.D
 Acq: 06 Jan 2022 13:58

Instrument :
 MSVOA_X
 ClientSampleId :







Tgt Ion:101 Resp: 37540
 Ion Ratio Lower Upper
 101 100
 103 63.0 54.1 81.1



#8
 Diethyl Ether
 Concen: 17.777 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. -0.000 min
 Lab File: VX026370.D
 Acq: 06 Jan 2022 13:58

Tgt Ion: 74 Resp: 14055
 Ion Ratio Lower Upper
 74 100
 45 104.7 50.4 151.2









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.204 ug/l
RT: 2.337 min Scan# 204
Delta R.T. 0.006 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 23301
Ion Ratio Lower Upper
101 100
85 45.7 36.5 54.7
151 77.2 60.6 90.8



#10
Methyl Iodide
Concen: 13.254 ug/l
RT: 2.459 min Scan# 225
Delta R.T. -0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58

Tgt Ion:142 Resp: 20797
Ion Ratio Lower Upper
142 100
127 48.5 37.5 56.3
141 14.4 11.4 17.0





#11

Tert butyl alcohol

Concen: 72.675 ug/l

RT: 2.983 min Scan# 311

Delta R.T. 0.000 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 59 Resp: 21807

Ion Ratio Lower Upper

59 100

57 10.9 8.2 12.2



#12

1,1-Dichloroethene

Concen: 19.153 ug/l

RT: 2.325 min Scan# 203

Delta R.T. -0.000 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58




Tgt Ion: 96 Resp: 22024


Ion Ratio Lower Upper

96 100

61 172.4 127.5 191.3


98 64.4 52.0 78.0








#13
Acrolein
Concen: 94.868 ug/l
RT: 2.239 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 56 Resp: 10179
Ion Ratio Lower Upper
56 100
55 73.1 57.4 86.2



#14
Allyl chloride
Concen: 18.117 ug/l
RT: 2.666 min Scan# 259
Delta R.T. -0.006 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58


Tgt Ion: 41 Resp: 41169
Ion Ratio Lower Upper
41 100
39 69.8 55.4 83.0
76 30.7 27.6 41.4








#15
Acrylonitrile
Concen: 88.291 ug/l
RT: 3.069 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 53 Resp: 65593
Ion Ratio Lower Upper
53 100
52 84.0 66.2 99.2
51 37.2 29.4 44.0



#16
Acetone
Concen: 90.445 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58


Tgt Ion: 43 Resp: 73620
Ion Ratio Lower Upper
43 100
58 28.2 24.8 37.2








#17
Carbon Disulfide
Concen: 17.810 ug/l
RT: 2.514 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 76 Resp: 58775
Ion Ratio Lower Upper
76 100
78 9.1 7.2 10.8



#18
Methyl Acetate
Concen: 17.900 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.006 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58

Tgt Ion: 43 Resp: 47260
Ion Ratio Lower Upper
43 100
74 22.6 19.1 28.7









#19
Methyl tert-butyl Ether
Concen: 21.128 ug/l
RT: 3.117 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 73 Resp: 84859
Ion Ratio Lower Upper
73 100
57 20.8 17.0 25.6



#20
Methylene Chloride
Concen: 18.299 ug/l
RT: 2.794 min Scan# 280
Delta R.T. 0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58


Tgt Ion: 84 Resp: 24596
Ion Ratio Lower Upper
84 100
49 130.0 97.1 145.7
51 43.5 29.3 43.9
86 65.8 50.8 76.2








#21
trans-1,2-Dichloroethene
Concen: 18.829 ug/l
RT: 3.093 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 96 Resp: 23620
Ion Ratio Lower Upper
96 100
61 148.2 105.7 158.5
98 63.7 52.1 78.1




#22
Diisopropyl ether
Concen: 19.945 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.013 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58



Tgt Ion: 45 Resp: 86506
Ion Ratio Lower Upper
45 100
43 84.5 63.8 95.6
87 27.4 24.2 36.4
59 10.9 9.4 14.2









#23
 Vinyl Acetate
 Concen: 96.442 ug/l
 RT: 3.721 min Scan# 41
 Delta R.T. -0.006 min
 Lab File: VX026370.D
 Acq: 06 Jan 2022 13:58




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 43 Resp: 332150
 Ion Ratio Lower Upper
 43 100
 86 10.4 9.0 13.6



#24
 1,1-Dichloroethane
 Concen: 19.723 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. -0.006 min
 Lab File: VX026370.D
 Acq: 06 Jan 2022 13:58

Tgt Ion: 63 Resp: 46360
 Ion Ratio Lower Upper
 63 100
 98 6.9 3.3 9.9
 100 4.4 2.2 6.6





#25

2-Butanone

Concen: 87.858 ug/l

RT: 4.556 min Scan# 5

Delta R.T. -0.006 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 99619

Ion Ratio Lower Upper

43 100

72 24.7 21.0 31.4



#26

2,2-Dichloropropane

Concen: 18.893 ug/l

RT: 4.477 min Scan# 556

Delta R.T. -0.006 min

Lab File: VX026370.D




Acq: 06 Jan 2022 13:58


Tgt Ion: 77 Resp: 33459

Ion Ratio Lower Upper

77 100

97 22.0 11.7 35.0





#27

cis-1,2-Dichloroethene

Concen: 19.732 ug/l

RT: 4.489 min Scan# 51

Delta R.T. -0.006 min


Lab File: VX026370.D

Acq: 06 Jan 2022 13:58

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion: 96 Resp: 28161

Ion Ratio Lower Upper

96 100

61 150.4 0.0 276.8

98 64.8 0.0 130.6



#28

Bromochloromethane

Concen: 19.912 ug/l

RT: 4.898 min Scan# 625

Delta R.T. -0.006 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58




Tgt Ion: 49 Resp: 20745


Ion Ratio Lower Upper

49 100

129 1.7 0.0 4.4

130 75.3 67.1 100.7





#29

Tetrahydrofuran

Concen: 86.147 ug/l

RT: 5.007 min Scan# 64

Delta R.T. -0.006 min


Lab File: VX026370.D

Acq: 06 Jan 2022 13:58

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion: 42 Resp: 61567

Ion Ratio Lower Upper

42 100

72 42.8 37.0 55.6

71 39.9 34.1 51.1



#30

Chloroform

Concen: 19.479 ug/l

RT: 5.093 min Scan# 657

Delta R.T. -0.012 min

Lab File: VX026370.D




Acq: 06 Jan 2022 13:58


Tgt Ion: 83 Resp: 48427

Ion Ratio Lower Upper

83 100


85 61.5 50.6 75.8








#31
Cyclohexane
Concen: 19.369 ug/l
RT: 5.471 min Scan# 719
Delta R.T. -0.006 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58

Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion: 56 Resp: 41439




Ion	Ratio	Lower	Upper
56	100		
69	31.5	23.9	35.9
84	85.0	73.2	109.8




#32
1,1,1-Trichloroethane
Concen: 19.839 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.012 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58

Tgt Ion: 97 Resp: 42878

Ion	Ratio	Lower	Upper
97	100		
99	63.2	51.5	77.3
61	43.5	35.5	53.3





#33

1,2-Dichloroethane-d4

Concen: 53.410 ug/l

RT: 5.958 min Scan# 799

Delta R.T. -0.006 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 65 Resp: 80239

Ion Ratio Lower Upper

65 100

67 50.9 0.0 103.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.006 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58




Tgt Ion: 114 Resp: 171623


Ion Ratio Lower Upper

114 100

63 20.9 0.0 40.0

88 17.0 0.0 30.6





#35

Dibromofluoromethane

Concen: 53.442 ug/l

RT: 5.385 min Scan# 706

Delta R.T. -0.006 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:113 Resp: 59629

Ion Ratio Lower Upper

113 100

111 103.9 81.8 122.8

192 18.9 15.0 22.4



#36

1,1-Dichloropropene

Concen: 19.610 ug/l

RT: 5.696 min Scan# 756

Delta R.T. -0.006 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58




Tgt Ion: 75 Resp: 35291


Ion Ratio Lower Upper

75 100

110 35.1 18.2 54.6

77 31.2 24.7 37.1





#37

Ethyl Acetate

Concen: 18.111 ug/l

RT: 4.715 min Scan# 595

Delta R.T. -0.006 min


Lab File: VX026370.D

Acq: 06 Jan 2022 13:58

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion: 43 Resp: 37791

Ion Ratio Lower Upper

43 100

61 13.0 10.7 16.1

70 11.0 8.5 12.7



#38

Carbon Tetrachloride

Concen: 18.340 ug/l

RT: 5.678 min Scan# 753

Delta R.T. -0.006 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58




Tgt Ion: 117 Resp: 36713


Ion Ratio Lower Upper

117 100

119 99.6 77.2 115.8

121 30.1 23.8 35.8





#39

Methylcyclohexane

Concen: 20.033 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 83 Resp: 40667

Ion Ratio Lower Upper

83 100

55 80.9 57.8 86.8

98 50.9 40.5 60.7



#40

Benzene

Concen: 19.249 ug/l

RT: 6.038 min Scan# 812

Delta R.T. -0.006 min

Lab File: VX026370.D




Acq: 06 Jan 2022 13:58


Tgt Ion: 78 Resp: 98864

Ion Ratio Lower Upper

78 100

77 22.6 18.6 27.8









#41
Methacrylonitrile
Concen: 18.882 ug/l
RT: 4.922 min Scan# 61
Delta R.T. -0.006 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:	41	Resp:	22250
Ion	Ratio	Lower	Upper
41	100		
39	56.5	44.2	66.2
67	66.0	59.7	89.5
52	27.9	24.2	36.2



#42
1,2-Dichloroethane
Concen: 19.716 ug/l
RT: 6.086 min Scan# 820
Delta R.T. -0.006 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58

Tgt Ion:	62	Resp:	40982
Ion	Ratio	Lower	Upper
62	100		
98	8.6	0.0	18.8





#43

Isopropyl Acetate

Concen: 18.665 ug/l

RT: 6.336 min Scan# 861

Delta R.T. -0.012 min


Lab File: VX026370.D

Acq: 06 Jan 2022 13:58




Instrument :

MSVOA_X

ClientSampleId :



Tgt Ion:	43	Resp:	63477
Ion	Ratio	Lower	Upper
43	100		
61	19.3	15.7	23.5
87	12.1	11.2	16.8



#44

Trichloroethene

Concen: 20.659 ug/l




RT: 7.129 min Scan# 991


Delta R.T. -0.000 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58

Tgt Ion:	130	Resp:	28551
Ion	Ratio	Lower	Upper
130	100		
95	98.0	0.0	186.4





#45

1,2-Dichloropropane

Concen: 19.416 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 63 Resp: 25508

Ion Ratio Lower Upper

63 100

65 32.1 25.8 38.6



#46

Dibromomethane

Concen: 18.352 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.006 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58




Tgt Ion: 93 Resp: 17883


Ion Ratio Lower Upper

93 100

95 82.8 66.6 100.0

174 97.3 77.3 115.9









#47
Bromodichloromethane
Concen: 18.408 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. -0.006 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 36494
Ion Ratio Lower Upper
83 100
85 62.5 53.5 80.3
127 9.2 6.6 10.0



#48
Methyl methacrylate
Concen: 19.073 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. -0.006 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58

Tgt Ion: 41 Resp: 31108
Ion Ratio Lower Upper
41 100
69 75.6 67.3 100.9
39 58.7 50.2 75.2





#49

1,4-Dioxane

Concen: 292.645 ug/l

RT: 7.684 min Scan# 1082

Delta R.T. -0.000 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 88 Resp: 8796

Ion Ratio Lower Upper

88 100

43 39.3 25.3 37.9#

58 72.8 51.8 77.6



#50

Toluene-d8

Concen: 50.834 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.006 min

Lab File: VX026370.D




Acq: 06 Jan 2022 13:58


Tgt Ion: 98 Resp: 210975

Ion Ratio Lower Upper

98 100

100 66.5 52.6 78.8









#51
4-Methyl-2-Pentanone
Concen: 91.006 ug/l
RT: 8.574 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 192601
Ion Ratio Lower Upper
43 100
58 37.8 30.8 46.2



#52
Toluene
Concen: 19.289 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58

Tgt Ion: 92 Resp: 60992
Ion Ratio Lower Upper
92 100
91 171.2 136.0 204.0





#53

t-1,3-Dichloropropene

Concen: 17.017 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.006 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 75 Resp: 33143

Ion Ratio Lower Upper

75 100

77 29.7 25.1 37.7



#54

cis-1,3-Dichloropropene

Concen: 18.279 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.006 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58




Tgt Ion: 75 Resp: 38302


Ion Ratio Lower Upper

75 100

77 31.3 25.3 37.9

39 53.3 41.4 62.0





#55

1,1,2-Trichloroethane

Concen: 18.931 ug/l

RT: 9.153 min Scan# 1323

Delta R.T. -0.000 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion: 97 Resp: 24601





Ion Ratio Lower Upper

97 100

83 86.4 70.5 105.7

85 55.7 43.8 65.6

99 59.4 49.4 74.2



#56

Ethyl methacrylate

Concen: 20.259 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. 0.000 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58




Tgt Ion: 69 Resp: 40604


Ion Ratio Lower Upper

69 100

41 71.1 52.8 79.2

39 41.0 32.5 48.7





#57

1,3-Dichloropropane

Concen: 19.516 ug/l

RT: 9.311 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 76 Resp: 43443

Ion Ratio Lower Upper

76 100

78 32.3 26.0 39.0



#58

2-Chloroethyl Vinyl ether

Concen: 94.207 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.006 min

Lab File: VX026370.D




Acq: 06 Jan 2022 13:58


Tgt Ion: 63 Resp: 94106

Ion Ratio Lower Upper

63 100

106 29.1 24.4 36.6





#59

2-Hexanone

Concen: 91.100 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.006 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 146043

Ion Ratio Lower Upper

43 100

58 52.4 26.9 80.6



#60

Dibromochloromethane

Concen: 17.017 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. -0.006 min

Lab File: VX026370.D




Acq: 06 Jan 2022 13:58


Tgt Ion: 129 Resp: 25052

Ion Ratio Lower Upper

129 100

127 79.2 38.6 115.7









#61
1,2-Dibromoethane
Concen: 18.855 ug/l
RT: 9.610 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:107 Resp: 26217
Ion Ratio Lower Upper
107 100
109 92.8 75.1 112.7



#62
4-Bromofluorobenzene
Concen: 54.216 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.006 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58

Tgt Ion: 95 Resp: 80101
Ion Ratio Lower Upper
95 100
174 72.6 0.0 157.4
176 70.0 0.0 154.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:117 Resp: 147501
Ion Ratio Lower Upper
117 100
82 56.8 43.8 65.6
119 32.3 25.2 37.8



#64
Tetrachloroethene
Concen: 22.706 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58

Tgt Ion:164 Resp: 28263
Ion Ratio Lower Upper
164 100
166 132.5 99.4 149.0
129 98.6 78.6 118.0
131 95.2 75.4 113.0









#65
Chlorobenzene
Concen: 19.809 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:112 Resp: 64187
Ion Ratio Lower Upper
112 100
114 32.4 26.6 39.8



#66
1,1,1,2-Tetrachloroethane
Concen: 19.452 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58

Tgt Ion:131 Resp: 24018
Ion Ratio Lower Upper
131 100
133 94.9 47.1 141.3
119 63.7 31.9 95.9





#67

Ethyl Benzene

Concen: 20.264 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. 0.000 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 91 Resp: 117403

Ion Ratio Lower Upper

91 100

106 31.7 25.7 38.5



#68

m/p-Xylenes

Concen: 41.017 ug/l

RT: 10.305 min Scan# 1512

Delta R.T. 0.000 min

Lab File: VX026370.D




Acq: 06 Jan 2022 13:58


Tgt Ion:106 Resp: 88730

Ion Ratio Lower Upper

106 100

91 209.5 158.6 238.0









#69
o-Xylene
Concen: 20.624 ug/l
RT: 10.640 min Scan# 1568
Delta R.T. -0.006 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:106 Resp: 43653
Ion Ratio Lower Upper
106 100
91 219.8 105.5 316.4



#70
Styrene
Concen: 20.560 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58

Tgt Ion:104 Resp: 71340
Ion Ratio Lower Upper
104 100
78 53.3 39.8 59.6
103 54.6 43.0 64.6





#71

Bromoform

Concen: 15.822 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:173 Resp: 15848

Ion Ratio Lower Upper

173 100

175 48.1 24.4 73.4

254 0.0 0.1 0.1#



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. 0.000 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58




Tgt Ion:152 Resp: 68143


Ion Ratio Lower Upper

152 100

115 74.3 41.3 123.9


150 164.8 0.0 351.8








#73
Isopropylbenzene
Concen: 21.801 ug/l
RT: 10.963 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:105 Resp: 115476
Ion Ratio Lower Upper
105 100
120 25.7 13.6 40.7



#74
N-ethyl acetate
Concen: 20.140 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. -0.006 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58

Tgt Ion: 43 Resp: 48984
Ion Ratio Lower Upper
43 100
70 40.3 34.1 51.1
55 23.2 19.2 28.8
61 23.0 19.0 28.4









#75
 1,1,2,2-Tetrachloroethane
 Concen: 17.822 ug/l
 RT: 11.213 min Scan# 1661
 Delta R.T. 0.000 min
 Lab File: VX026370.D
 Acq: 06 Jan 2022 13:58




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 83 Resp: 32105
 Ion Ratio Lower Upper
 83 100
 131 11.2 5.1 15.3
 85 62.8 31.6 95.0



#76
 1,2,3-Trichloropropane
 Concen: 22.496 ug/l
 RT: 11.238 min Scan# 1665
 Delta R.T. -0.006 min
 Lab File: VX026370.D
 Acq: 06 Jan 2022 13:58

Tgt Ion: 75 Resp: 37001
 Ion Ratio Lower Upper
 75 100
 77 32.2 19.0 57.0









#77
Bromobenzene
Concen: 20.830 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:156 Resp: 26165
Ion Ratio Lower Upper
156 100
77 170.9 78.2 234.6
158 97.9 48.6 145.9



#78
n-propylbenzene
Concen: 21.525 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58

Tgt Ion: 91 Resp: 131086
Ion Ratio Lower Upper
91 100
120 23.1 12.0 36.1









#79
2-Chlorotoluene
Concen: 21.418 ug/l
RT: 11.366 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 80291
Ion Ratio Lower Upper
91 100
126 31.9 17.8 53.3



#80
1,3,5-Trimethylbenzene
Concen: 21.830 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58

Tgt Ion: 105 Resp: 95955
Ion Ratio Lower Upper
105 100
120 47.8 24.9 74.7









#81
trans-1,4-Dichloro-2-butene
Concen: 12.351 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 7022
Ion Ratio Lower Upper
75 100
53 144.4 79.5 119.3#
89 48.3 38.3 57.5



#82
4-Chlorotoluene
Concen: 21.886 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58

Tgt Ion: 91 Resp: 92661
Ion Ratio Lower Upper
91 100
126 27.8 15.6 46.7





#83

tert-Butylbenzene

Concen: 21.234 ug/l

RT: 11.713 min Scan# 1

Delta R.T. -0.006 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:119 Resp: 90596

Ion Ratio Lower Upper

119 100

91 63.1 28.7 86.1

134 23.7 11.9 35.9



#84

1,2,4-Trimethylbenzene

Concen: 21.799 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.006 min

Lab File: VX026370.D




Acq: 06 Jan 2022 13:58


Tgt Ion:105 Resp: 95742

Ion Ratio Lower Upper

105 100

120 43.7 22.9 68.7









#85
 sec-Butylbenzene
 Concen: 21.679 ug/l
 RT: 11.890 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VX026370.D
 Acq: 06 Jan 2022 13:58




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion:105 Resp: 115427
 Ion Ratio Lower Upper
 105 100
 134 19.7 10.4 31.1



#86
 p-Isopropyltoluene
 Concen: 21.839 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX026370.D
 Acq: 06 Jan 2022 13:58

Tgt Ion:119 Resp: 96933
 Ion Ratio Lower Upper
 119 100
 134 26.1 13.2 39.6
 91 25.8 11.9 35.9





#87

1,3-Dichlorobenzene

Concen: 20.788 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:146 Resp: 49020

Ion Ratio Lower Upper

146 100

111 42.8 19.9 59.6

148 64.5 31.9 95.5



#88

1,4-Dichlorobenzene

Concen: 19.358 ug/l

RT: 12.043 min Scan# 1797

Delta R.T. -0.000 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58




Tgt Ion:146 Resp: 48940


Ion Ratio Lower Upper

146 100

111 43.2 19.9 59.7

148 64.1 31.9 95.7





#89

n-Butylbenzene

Concen: 20.968 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 91 Resp: 82709

Ion Ratio Lower Upper

91 100

92 53.8 27.3 81.8

134 25.6 13.5 40.4



#90

Hexachloroethane

Concen: 15.087 ug/l

RT: 12.542 min Scan# 1879

Delta R.T. 0.000 min

Lab File: VX026370.D




Acq: 06 Jan 2022 13:58


Tgt Ion: 117 Resp: 13375

Ion Ratio Lower Upper

117 100


201 72.4 35.4 106.3








#91
1,2-Dichlorobenzene
Concen: 19.334 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:146 Resp: 47293
Ion Ratio Lower Upper
146 100
111 44.2 20.6 61.9
148 62.9 31.9 95.7



#92
1,2-Dibromo-3-Chloropropane
Concen: 16.344 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58


Tgt Ion: 75 Resp: 7537
Ion Ratio Lower Upper
75 100
155 76.1 44.0 131.8
157 100.5 55.5 166.5








#93
1,2,4-Trichlorobenzene
Concen: 20.198 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:180 Resp: 29196
Ion Ratio Lower Upper
180 100
182 96.3 47.2 141.6
145 34.5 16.3 48.9




#94
Hexachlorobutadiene
Concen: 20.816 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX026370.D
Acq: 06 Jan 2022 13:58



Tgt Ion:225 Resp: 12493
Ion Ratio Lower Upper
225 100
223 64.5 31.6 95.0
227 63.6 32.5 97.5





#95

Naphthalene

Concen: 20.016 ug/l

RT: 13.780 min Scan# 2082

Delta R.T. 0.000 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:128 Resp: 105418

Ion Ratio Lower Upper

128 100

127 12.8 10.3 15.5

129 10.8 8.8 13.2



#96

1,2,3-Trichlorobenzene

Concen: 20.102 ug/l

RT: 13.963 min Scan# 2112

Delta R.T. -0.000 min

Lab File: VX026370.D

Acq: 06 Jan 2022 13:58

Tgt Ion:180 Resp: 29564

Ion Ratio Lower Upper

180 100

182 95.4 47.5 142.6

145 35.5 16.7 50.1

